



Ohio Valley Hospital
One Ross Park
Steubenville, Ohio 43952
614-283-7000

November 15, 1990

U. S. Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Elyn, IL 60137

Re: Decommission Plan and Docket Number 03000429
for License Number 34-13317-01

Gentlemen:

On November 5, 1990, we received a notice stating that we were not in compliance with 10 CFR 30.35. We were relieved of this responsibility in Condition 26 of our License, provided we restricted our possession of licensed material below the minimum limits specified in 10 CFR 30.35(d). Our only source is a 6,000 Ci Co-60 sealed source for our teletherapy unit which is below the 10,000 Ci limit required for financial assurance. In addition, we have recently acquired a Linear Accelerator and are planning to decommission the unit in the not so distant future.

For further information, we would like to reference a letter sent to us from Patricia J. Pelke of the USNRC, Region III Materials Licensing Section, dated November 15, 1989, describing our responsibilities in this matter. We feel that we have complied with the terms and conditions of this letter and our License.

If you have any questions, please do not hesitate to contact our Therapy Physicist, Marcel Szal. His telephone number is 614-283-7683.

Sincerely,

OHIO VALLEY HOSPITAL

Fred B. Brower
President

9102150062 910130
REQ3 LIC30
34-13317-01 PDR

FBB/MP/jaf

Enclosures

(NRC letter of November 15, 1989
and License Amendment 12)

RECEIVED
NOV 19 1990
REGION III

NOV 14 1990



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
799 ROOSEVELT ROAD
GLEN ELLYN, ILLINOIS 60137

November 2, 1990

Docket No. 03000429
License No. 34-13317-01

Ohio Valley Hospital
One Ross Park
Steubenville, OH 43952

Gentlemen:

As of July 27, 1990, you were required to comply with 10 CFR 30.35, of the NRC regulations, which requires licensees with quantities of licensed material that you are authorized to possess to submit a decommissioning funding plan or a certificate that financial assurance for decommissioning has been provided. Our records indicate that NRC has not yet received your response to this requirement. Therefore, you appear to be in violation of 10 CFR 30.35.

The NRC considers violation of this requirement to be a significant regulatory concern. Therefore, in order for the NRC to determine whether your license should be modified, suspended or revoked, or other enforcement action taken to ensure compliance with NRC regulatory requirements, you are required to respond as described in the enclosed Demand for Information within 30 days.

If you have any questions concerning this request, please contact our Materials Licensing Section at (708) 790-5625.

Sincerely,

A. Bert Davis
Regional Administrator

Enclosure: Demand for Information

cc w/enclosure:
DCD/DCB(RIDS)

CERTIFIED MAIL
RETURN RECEIPT REQUESTED

901203046



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
789 ROOSEVELT ROAD
GLEN ELLYN, ILLINOIS 60137

Copy to Janet Gratzmiller 11-22-89

NOV 15 1989

Ohio Valley Hospital
ATTN: Fred Brower, President
One Ross Park
Steubenville, OH 43952

Gentlemen:

Enclosed is your Amendment No. 12 to your NRC License No. 34-13317-01 in accordance with your request.

On June 27, 1988, the Commission established regulations for decommissioning licensed nuclear facilities. The regulations address decommissioning planning needs, timing, funding methods, etc. The intent of the regulations is to ensure that the decommissioning of all licensed facilities will be accomplished in a safe and timely manner, and that licensees will provide adequate funds to cover all costs associated with decommissioning. According to 10 CFR 30.35 (enclosed), financial assurance requirements for a Part 30 licensee are based upon the type, form, and quantity of material authorized on the license. A license which authorizes an unlimited quantity of material (i.e., does not specify a possession limit) requires that a licensee develop a complete decommissioning funding plan. In order to relieve you of this responsibility, we have added License Condition 26. to your license which states that you must restrict your possession of licensed material below the minimum limits specified in 10 CFR 30.35(d) for establishing decommissioning financial assurance. This condition will be deleted should you ever choose to amend your license to specify possession limits.

However, you should note that all licensees, regardless of possession limits, are required to keep records relative to information which the Commission considers important to decommissioning. Therefore, we have added License Condition 27. requiring that you maintain such records as set forth in Section 30.35(g).

Please review your license carefully to assure that you understand and are in compliance with the terms and conditions contained therein. If you have questions, please contact our office at (708) 790-5625.

Sincerely,

Patricia J. Pelke
Materials Licensing Section

Enclosures:

1. Amendment No. 12 to License No. 34-13317-01
2. 10 CFR Part 30

3980 CI 11/90

~~9443244349~~

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number	34-13317-01
Docket or Reference number	030-00429
Amendment No. 12	

Ohio Valley Hospital
One Ross Park
Steubenville, OH 43952

In accordance with letter dated August 8, 1989, License Number 34-13317-01 is amended as follows:

Item Two (Address) is changed from: 380 Summit Avenue
Steubenville, OH 43952

to: One Ross Park
Steubenville, OH 43952

Item 9. is amended to read:

9. Authorized Use:

- A. One source to be used in a Ricker Model 6296 (C9M/80) teletherapy unit for treatment of humans and for instrument calibrations and intercomparisons. One source in its shipping container to be in possession of the licensee as necessary to the replacement of the source in the teletherapy unit only.

Conditions 11. and 12. are amended to read:

11. A. Radiation Safety Officer: Louis James Jacques, M.D.

B. Teletherapy Physicist: Marcel Szal, M.S.

12. Authorized Users:

A. Joseph Concannon, M.D.

E. Julian Bractor, M.D.

B. James Hughes, M.D.

F. Mark Trombetta, M.D.

C. Gerald Medwick, D.O.

G. Marcel Szal, M.S., for instrument calibrations and intercomparisons.

D. Tarit Dutta, M.D.

Conditions 26. and 27. are added:

- 26. The licensee is exempted from decommissioning financial assurance requirements for possession of licensed material in sealed sources in quantities greater than the limits of 10 CFR 30.35(d) for the purpose of source changes only. This exemption is granted for no more than 30 days for any one source change.

9003200309

5

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number 34-13317-01

Docket or Reference number 030-00429

Amendment NO. 12

27. The licensee shall maintain records of information important to safe and effective decommissioning at the address specified in Condition 10, per the provisions of 10 CFR 30.35(g) until this license is terminated by the Commission.



For the U.S. Nuclear Regulatory Commission

Date: November 15, 1989

Original Signed
By Patricia J. Pelke
Materials Licensing Section, Region III
COPY



Ohio Valley Hospital
One Ross Park
Steubenville, Ohio 43952
614-283-7000

November 15, 1990

U. S. Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Elyn, IL 60137

Re: Decommission Plan and Docket Number 03000429
for License Number 34-13317-01

Gentlemen:

On November 5, 1990, we received a notice stating that we were not in compliance with 10 CFR 30.35. We were relieved of this responsibility in Condition 26 of our License, provided we restricted our possession of licensed material below the minimum limits specified in 10 CFR 30.35(d). Our only source is a 6,000 Ci Co-60 sealed source for our teletherapy unit which is below the 10,000 Ci limit required for financial assurance. In addition, we have recently acquired a Linear Accelerator and are planning to decommission the unit in the not so distant future.

For further information, we would like to reference a letter sent to us from Patricia J. Pelke of the USNRC, Region III Materials Licensing Section, dated November 15, 1989, describing our responsibilities in this matter. We feel that we have complied with the terms and conditions of this letter and our License.

If you have any questions, please do not hesitate to contact our Therapy Physicist, Marcel Szal. His telephone number is 614-283-7683.

Sincerely,

OHIO VALLEY HOSPITAL

Fred B. Brower
President

9102150062
FBB/MP/jaf

Enclosures
(NRC letter of November 15, 1989
and License Amendment 12)



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
799 ROOSEVELT ROAD
GLEN ELLYN, ILLINOIS 60137

November 2, 1990

Docket No. 03000429
License No. 34-13317-01

Ohio Valley Hospital
One Ross Park
Steubenville, OH 43952

Gentlemen:

As of July 27, 1990, you were required to comply with 10 CFR 30.35, of the NRC regulations, which requires licensees with quantities of licensed material that you are authorized to possess to submit a decommissioning funding plan or a certificate that financial assurance for decommissioning has been provided. Our records indicate that NRC has not yet received your response to this requirement. Therefore, you appear to be in violation of 10 CFR 30.35.

The NRC considers violation of this requirement to be a significant regulatory concern. Therefore, in order for the NRC to determine whether your license should be modified, suspended or revoked, or other enforcement action taken to ensure compliance with NRC regulatory requirements, you are required to respond as described in the enclosed Demand for Information within 30 days.

If you have any questions concerning this request, please contact our Materials Licensing Section at (708) 790-5625.

Sincerely,

A. Bert Davis
Regional Administrator

Enclosure: Demand for Information

cc w/enclosure:
DCD/DCB(RIDS)

CERTIFIED MAIL
RETURN RECEIPT REQUESTED

9012030440



UNITED STATES
NUCLEAR REGULATORY COMMISSION

Copy to Janet Gratzmiller 11-22-89

REGION III
799 ROOSEVELT ROAD
GLEN ELLYN, ILLINOIS 60137

NOV 15 1989

Ohio Valley Hospital
ATTN: Fred Brower, President
One Ross Park
Steubenville, OH 43952

Gentlemen:

Enclosed is your Amendment No. 12 to your NRC License No. 34-13317-01 in accordance with your request.

On June 27, 1988, the Commission established regulations for decommissioning licensed nuclear facilities. The regulations address decommissioning planning needs, timing, funding methods, etc. The intent of the regulations is to ensure that the decommissioning of all licensed facilities will be accomplished in a safe and timely manner, and that licensees will provide adequate funds to cover all costs associated with decommissioning. According to 10 CFR 30.35 (enclosed), financial assurance requirements for a Part 30 licensee are based upon the type, form, and quantity of material authorized on the license. A license which authorizes an unlimited quantity of material (i.e., does not specify a possession limit) requires that a licensee develop a complete decommissioning funding plan. In order to relieve you of this responsibility, we have added License Condition 26. to your license which states that you must restrict your possession of licensed material below the minimum limits specified in 10 CFR 30.35(d) for establishing decommissioning financial assurance. This condition will be deleted should you ever choose to amend your license to specify possession limits.

However, you should note that all licensees, regardless of possession limits, are required to keep records relative to information which the Commission considers important to decommissioning. Therefore, we have added License Condition 27. requiring that you maintain such records as set forth in Section 30.35(g).

Please review your license carefully to assure that you understand and are in compliance with the terms and conditions contained therein. If you have questions, please contact our office at (708) 790-5625.

Sincerely,

Patricia J. Pelke
Materials Licensing Section

Enclosures:

1. Amendment No. 12 to
License No. 34-13317-01
2. 10 CFR Part 30

3980 CI 11/90

9003200309

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number 34-13317-01

Docket or Reference number 030-00429

Amendment No. 12

Ohio Valley Hospital
One Ross Park
Steubenville, OH 43952

In accordance with letter dated August 8, 1989, License Number 34-13317-01 is amended as follows:

Item Two (Address) is changed from: 380 Summit Avenue
Steubenville, OH 43952

to: One Ross Park
Steubenville, OH 43952

Item 9. is amended to read:

9. Authorized Use:

A. One source to be used in a Picker Model 6295 (C9M-80) teletherapy unit for treatment of humans and for instrument calibrations and intercomparisons. One source in its shipping container to be in possession of the licensee as necessary to the replacement of the source in the teletherapy unit only.

Conditions 11. and 12. are amended to read:

11. A. Radiation Safety Officer: Louis James Jacques, M.D.

B. Teletherapy Physicist: Marcel Szal, M.D.

12. Authorized Users:

A. Joseph Concannon, M.D.

E. Julian Bractor, M.D.

B. James Hughes, M.D.

F. Mark Trombetta, M.D.

C. Gerald Medwick, D.O.

G. Marcel Szal, M.S., for instrument calibrations and intercomparisons.

D. Tarit Dutta, M.D.

Conditions 26. and 27. are added:

26. The licensee is exempted from decommissioning financial assurance requirements for possession of licensed material in sealed sources in quantities greater than the limits of 10 CFR 30.35(d) for the purpose of source changes only. This exemption is granted for no more than 30 days for any one source change.

900320309

COPY

5

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number 34-13317-01

Docket or Reference number 030-00429

Amendment NO. 12

27. The licensee shall maintain records of information important to safe and effective decommissioning at the address specified in Condition 10, per the provisions of 10 CFR 30.35(g) until this license is terminated by the Commission.



For the U.S. Nuclear Regulatory Commission

Date: November 15, 1989

Original Signed
By Patricia J. Pelke
Materials Licensing Section, Region III
COPY