

ATP

233

DCS  
misadministration  
(CF Available)

TO: Referring Physician Dr. [ ]  
NRC Regional Office  
Office of Inspection and Enforcement

Also send copy to:  
U.S. Nuclear Regulatory Commission  
c/o Document Management Branch  
Washington, D.C. 20555



REPORT OF DIAGNOSTIC MISADMINISTRATION

Date of report 7-1-81 Date of Incident 6-30-81

Licensee St. Michael Hospital Number 48-02435-01

Address 2400 W. Villard Ave. 414- 3432

City, State MILW. WI. 53209 263-8880

Referring Physician Dr. [ ] 263-8337

Signature [ ]

Description of event ///

Misadministration of linear material  
instead of bone material.

3mCi TC 99 HAA 20mCi TC HDP

Effect on patient NONE

Preventive action Public check ref.

This report is to be sent within ten (10) days after the end of calendar quarter in which the incident occurred.

The Medical Isotope Committee and the licensee (administration) has been informed of this incident.

Submitted by:

**XA Copy Has Been Sent to PDR**

8108210447 B10701  
NMS LIC30  
48-02435-01

Radiation Safety Officer

Daniel Price M.D.

FE30  
1/0

JUL 10 1981 (13)