

DOCTORS HOSPITAL RADIOLOGY SERVICES, INC.

2730 EAST JEFFERSON AVENUE - DETROIT, MICH. 48207 - TELEPHONE (313) 259-3050

December 3, 1981

United States
Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Gentlemen:

This is a response to your letter dated 12 November 1981. In regards to the areas of non-compliance as stated, I am pleased to say that we are in 100% compliance as of this date.

a. The minutes of the Isotope Committee have been brought up to date as explained to Mr. R. E. Burgin during his visit.

b. Linearity checks on the Dose Calibrator are being conducted daily and documented when used.

c. Molybdenum⁹⁹ checks are being conducted daily and documented when used.

I assure the Commission that there will be closer supervision both by myself and our Health Physicist in these areas, and total compliance will be maintained.

Sincerely yours,



Steven A. Lenzion, R. T., C.N.M.T.
Director, Radiology Services

SAL/jmc

8112240155 811218
NMS LIC30
21-16650-01 PDR

DEC 7 1981