**U.S. NUCLEAR REGULATORY COMMISSION** NRC FLAM 366 (7.77) LICENSEE EVENT REPORT (PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION) CONTROL BLOCK: 0 0 0 10 CON'T REPORT 0 L(6)|00 1 0 0 EVENT DATE EVENT DESCRIPTION AND PROBABLE CONSEQUENCES (10) Routine surveillance during plant operation revealed that Unit No. 1 Post-Accident 0 Monitors, 1-CAC-AQH-1260-1 and 2, 1261-1, 2, and 3, and 1262-1 and 2 were showing 0 3 unexpected upscale indications. In addition, Unit No. 2 Post-Accident Monitors, 0 4 2-CAC-AQH-1260-1 and 3, and 1262-1 and 3, were also showing unexpected upscale indications. This event did not affect the health and safety of the public. This 0 6 resulted in a loss of DW particulate monitoring for Unit No. 1. Technical Specifications 3.3.5.3, 3.4.3.1a, 6.9.1.9b COMP SYSTEM CAUSE CAUSE VALVE COMPONENT CODE SURCODE Z 1 A (13) SI UI(14 Y (15) 1D BI N R (16) T REVISION SEQUENTIAL OCCURRENCE REPORT REPORT NO TYPE 0 11 01 3 0 COMPONENT PRIME COMP. ATTACHMENT SUBMITTED NPRD-4 METHOD FUTURE EFFECT HOURS (22) FORM SUB. SUPPLIER 2 10 0 3 0 5 Х 0 N N CAUSE DESCRIPTION AND CORRECTIVE ACTIONS (27 These indication failures resulted from moisture accumulation in the sampling chambers of each affected monitor, originating from the humid drywell sample flow. The moisture was removed and each monitor was returned to service. An engineering work request has been submitted to plant engineering to evaluate the moisture problem and what modification can be made to eliminate it. A 80 METHOD OF DISCOVERY FACILITY (30) DISCOVERY DESCRIPTION (32) OTHER STATUS % POWER 91 0 9 Operator Surveillance A (31) 1.1 80 CONTENT ACTIVITY AMOUNT OF ACTIVITY (35) LOCATION OF RELEASE (36) OF RELEASE RELEASED ZI 2 (34) NA NA (33) 80 44 PERSONNEL EXPOSURES DESCRIPTION (39) 010 NA 10 Z (38) 80 PERSONNEL INJURIES DESCRIPTION (41) 0 01 0 NA (40)80 LOSS OF OR DAMAGE TO FACILITY (43) DESCRIPTION TYPE Z (42) NA 80 PUBLICITY NRC USE ONLY DESCRIPTION (45 N (44) NA RR 69 80 8203080037 820222 ADOCK 05000325 PDR J. Pastva, (919)457-952 Jr. PHONE: PDR S