

Palos Community Hospital

80th Avenue at McCarthy Road

Palos Heights, Illinois 60463

Phone (312) 361-4500

March 3, 1983

Darryl Wiedeman, Chief
Materials Radiation Protection
U.S. Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, IL 60137

Re: Inspection of license #12-15467-01

Dear Mr. Wiedeman:

This is in response to your February 18, 1983 letter concerning the February 10, 1983 NRC inspection of our facilities.

The item of non-compliance noted in your report was the possession of two (2) rather than one (1) Am-241 Searle Analytic Anatomical Marker Sources as stated in our current license. The corrective action we will take is to obtain a license amendment for possession of two sources. This request was included in our January 1983 license renewal application and therefore full compliance will be achieved as soon as that pending application is processed.

Your inspector was very professional and courteous. We look forward to receiving confirmation from your office that we are again in compliance with our license conditions.

Thank you.

Sincerely,



Fred Fleming, Vice President
Ancillary Services

/lew

8303170349 830310
NMS LIC30
12-15467-01 PDR

MAR 07 1983