

February 28, 1983

U.S. Nuclear Regulatory Commission
Radioisotopes Licensing Section
Region III
799 Roosevelt Road
Glen Ellyn IL 60137

Re: Routine Safety Inspection of License No. 12-13645-01

Gentlemen

This is in follow up to the February 14, 1983 letter from your office regarding the routine radiation safety inspection of our nuclear medicine facilities of January 25, 1983.

We confirm Appendix I of the NRC Medical Licensing Guide, referenced in our current license application, will be followed. Specifically, daily radiation surveys of the nuclear technologist's hands, clothing, and dose preparation area, generator elution and patient injection areas will be conducted. Compliance with this item is immediate.

We trust this response is sufficient to correct this item of non-compliance.

Sincerely



Eugene Pritchard
President

/gb

MAR 2 1983