

DMB

MAR 14 1983

Docket No. 50/409

Dairyland Power Cooperative
LACBWR
ATTN: Bruce Zibung
Health and Safety Superisior
P.O. Box 275
Genoa, WI 54632

Dear Mr. Zibung:

As referred in I&E Inspection Report No. 50-409/82-17, a Sping 4 particulate sample was analyzed by our reference laboratory (Radiological Environmental Services Laboratory - RESL) in an attempt to resolve disagreements identified by our onsite comparisons with the Region III Mobile Laboratory. RESL results were comparable to those of the Mobile Laboratory; ratios of results (LACBWR/RESL) were 2.6 for ^{60}Co and 6.0 for ^{137}Cs . As discussed with Larry Nelson on February 23, 1983, a spiked particulate filter (#A-64) is being forwarded to him under separate cover for analysis with the new calibration data (entered November 1982). Please return the filter and data as soon as the analysis is complete.

Please refer all questions to Nora Nicholson at 312/932-2527.

Sincerely,

C. J. Paperiello
C. J. Paperiello, Chief
Emergency Preparedness and
Radiological Safety Branch

cc: J. Parkyn, Plant Superintendent
DMB/Document Control Desk (RIDS)
Resident Inspector, RIII
John J. Duffy, Chief
Boiler Section
Stanley York, Chairman
Public Service Commission

RIII
hu
Nicholson/jp
3/3/83

RIII
R
Schumacher

RIII
CJP
Paperiello
3/11/83

8303170520 830314
FDR ADOCK 05000409
Q FDR

IE:06