NRC FORM 591 Part 1 (7-91)

10 000 0 001

SAFETY INSPECTION

				Page 1 of /
1. LICENSEE EG&G Florida, Inc. P.O. Box 21296 Kennedy Space Center, Florida 32815		U S N 101 M	REGIONAL OFFICE REGION II U S NUCLEAR REGULATORY COMMISSION 101 MARIETTA STREET NW SUITE 2900 ATLANTA GA 30323	
3. DOCKET NUMBER(S)	4. LICENSE NUMBER(S	S)	5. DATE OF INSPECTION	
030-20073	09-21233-01		5-3 4 4-94	
LICENSEE: The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:				
Within the scope of this inspection, no violations were observed.				
The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on those actions at this time.				
3. During this inspection certain of your activities, as described below or attached, were in violation of NRC requirements. This form is a NOTICE OF VIOLATION, which is required to be posted in accordance with 10 CFR 19.11.				
A			was not properly i	posted to
indicate the presence of a	Indicate the presence of a 10 CFR 20.203(b),(c),(dj.(e) or 34			
B	B of sealed sources were not			
performed at the proper frequencies. 10 CFR or License Condition Number				
C. Records of were not properly maintained.				
10 CFR or License Condition Number				
D. Documents were not properly posted or otherwise made available. 10 CFR 19.11.				
E. Reports or notification of			were not made in	accordance with
10 CFR or License Condition Number				
F				
9405020117 940504 PDR ADDCK 03020073 C PDR				
I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified in the items chacked above. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.				
SIGNATURE - LICENSEE 010020	DATE	SIGNATURE - NRC IN	SPECTOR Sulle	5/1/94