PUBLIC PUBLIC

Sanofi Diagnostics Pasteur, Inc.
ATTN: Terry Bicker
President
1000 Lake Hazeltine Drive
Chasta, MN 55318

License No. 22-13838-01G Docket No. 030-076329 License No. 22-13838-01 Docket No. 030-05055

Dear Mr. Bicker:

This refers to the routine safety inspection conducted by Mark Mitchell of this office on May 11, 1994, of activities authorized by NRC Byproduct Material License Nos. 22-13838-01 and 22-13838-01G, and to the discussion of our findings with Mr. Bill Cotter and other members of your staff at the conclusion of the inspection.

The inspection was an examination of activities conducted under your license as they relate to radiation safety and to compliance with the Commission's rules and regulations and with the conditions of your license. The inspection consisted of a selective examination of procedures and representative records, observations, independent measurements, and interviews with personnel.

No violations of NRC requirements were identified during the course of this inspection.

However, regarding License No. 22-13828-01, our inspector expressed concern during the inspection that the radiation safety program lacks clear procedures for waste segregation that assures proper disposal of radioactive materials. Specifically, radioactive material users lack detailed guidance directing them to segregate wastes of various isotopes. Therefore, please describe those actions taken or planned to ensure that radioactive materials are properly segregated prior to storage for decay and disposal.

Response directed by this letter is not subject to clearance procedures of the Office of Management and Budget as required by the Paperwork Reduction Act of 1980, PL 96-511.

In accordance with 10 CFR 2.790 of the Commission's regulations, a copy of this letter will be placed in the NRC Public Document Room.

We will gladly discuss any questions you have concerning this inspection.

Sincerely, Original Signed by John D. Jones

John D. Jones, Acting Chief Nuclear Materials Inspection Section 2

bcc: PUBLIC

Mitchell/sd 05/24/94 Jones OS 126194

010094

9406020119 940527 PDR ADOCK 03005055 C PDR