

MAY 11 1994

Docket No. 030-33027
License No. 47-25233-01MD

Tri-States Area Nuclear Pharmacy
ATTN: Craig Lewandowski, Pharm. D.
Manager
611 7th Avenue
Huntington, West Virginia 25701

Gentlemen:

SUBJECT: NRC INSPECTION REPORT NO. 47-25233-01MD/94-01

Thank you for your response of April 22, 1994, received on April 28, 1994, to our Notice of Violation, issued on March 28, 1994, concerning activities conducted under NRC License No. 47-25233-01MD. We have evaluated your response and found that it meets the requirements of 10 CFR 2.201. Concerning your Ludlum Model 2200 counting equipment, we agree that as long as you follow the guidance in NRC Information Notice 93-30, including documentation of your efforts, you should be able to demonstrate that your equipment is sensitive enough to meet NRC regulatory requirements. Our records have been modified to reflect that the Ludlum Model 2200 counting equipment example is not included in the violation. We will examine the implementation of your corrective actions during future inspections.

We appreciate your cooperation in this matter.

Sincerely,

Douglas M. Collins, Chief
Nuclear Materials Safety and
Safeguards Branch
Division of Radiation Safety
and Safeguards

cc: State of West Virginia

bcc: Document Control Desk
RII Docket File, DRSS

SEND	OFC	RII:DRSS	RII:DRSS			
TO	NAME	MFuller	CHosey			
PDR?	DATE	05/11/94	05/11/94	/ / 94	/ / 94	/ / 94
Yes	No	COPY?	Yes	No	Yes	No

OFFICIAL RECORD COPY

DOCUMENT NAME:G:\DRSS\NMIS\TRISTATE.ACK

010050

9406020168 940511
PDR ADOCK 03033027
C PDR

10

DF07