

Certified By mar**BALTIMORE GAS AND ELECTRIC COMPANY**

P.O. BOX 1475

BALTIMORE, MARYLAND 21203

NEAR POWER DEPARTMENT  
 WEST CLIFFS NUCLEAR POWER PLANT  
 P.O. BOX 1475  
 BALTIMORE, MARYLAND 21203

September 13, 1982

Mr. Ronald C. Haynes, Director  
 U. S. Nuclear Regulatory Commission  
 Region 1  
 631 Park Avenue  
 King of Prussia, PA 19406

Docket No. 50-317/50-318  
 License No. DPR 53/DPR 69

SUBJECT: LER 82-53 (U-1); LER 82-43 (U-2)

Dear Mr. Haynes:

This letter is to confirm our verbal notification of your Mr. Dave Trimble of the discovery that our only operable hydrogen analyzer was technically inoperable for a brief period of time.

At 1045 on September 10, 1982 with Units No. 1 and 2 in Mode 1, it was discovered that the hydrogen analyzer (T.S. 3.6.5.1) was unintentionally adjusted out of calibration rendering the unit inoperable. The instrument was recalibrated and returned to service at 1125 terminating the event.

A complete description of the events leading to this discovery and any further corrective actions will be addressed in a follow-up report.

Very truly yours,

*L. B. Russell*

L. B. Russell  
 Plant Superintendent

LBR:bsg

cc: J. A. Tiernan  
 Director, Office of Management  
 Information & Program Control  
 J. M. Bills

R. M. Douglass  
 R. E. Architzel  
 L. B. Russell  
 W. S. Gibson