

Appendix

NOTICE OF VIOLATION

St. Vincent Hospital and Medical Center

License No. 34-01216-03

As a result of the inspection conducted on February 25, 1983, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violation was identified:

License Condition No. 17 requires that the licensee shall possess and use licensed material in accordance with statements, representations, and procedures contained in application dated October 27, 1978, and various letters.

Item 14 of application dated October 27, 1978, requires that packages containing radioactive material be opened in accordance with the procedures outlined in Appendix F of attachment I-1 and I-2.

Steps 2, 3, and 6 of Appendix F requires that, when opening packages in which licensed material has been received, the exposure rate 3 feet from the surface of the package and at the surface be measured and recorded, and that a wipe test be performed of the final source container and recorded.

Contrary to the above, as of February 25, 1983, packages containing licensed material received by the licensee from Pharmatopes, Toledo, Ohio, have not been surveyed or wipe tested.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

3/16/83
Dated

for S.R. Lawk
D. J. Sreniawski, Chief
Materials Radiation Protection
Section 2