



Polyclinic Medical Center

Harrisburg, Pennsylvania 17105

DEPARTMENT OF RADIOLOGY
(717) 782-4250

RICHARD M. FENCEL, M.D., Director

Associates:

G. Thomas Balsbaugh, M.D.
Richard R. Brock, M.D.
Milton A. Friedlander, M.D.
Leonard C. Griff, M.D.
John C. Morton, M.D.
Frederick C. Newton, M.D.
Richard J. Pawelski, M.D.
Theodore A. Tristan, M.D.

Divisions:

Angiography
Computerized Tomography
Diagnostic Roentgenology
Nuclear Medicine
Radiation Therapy
Ultrasound

Robert M. Olley, R.T.,
Administrator

28 February 1983

Mr. Thomas T. Martin
Director, Division of Engineering
and Technical Programs
USNRC, Region I
631 Park Avenue
King of Prussia, PA 19406

RE: Polyclinic Medical Center
License Number 37-00358-04
" " 37-00358-05
Inspection Number 030-02947/83-01
" " 030-00452/83-01

Dear Mr. Martin:

Referring to the above inspection, this is the 30 day report required by your office concerning the violations found by your inspectors.

- A) Failure of personnel to adequately monitor their hands after handling isotopes.
- 1) Corrective Steps Taken
 - a) Following your inspection on 1/18/83, the chief technologist conducted a review of hand surveying procedures for the technologists.
 - 2) Corrective Steps Which Will Be Taken
 - a) The hospital's medical radiation physicist will instruct all Nuclear Medicine personnel on the proper procedure for monitoring their hands, and for decontamination when needed. In addition, periodic unannounced surveys will be made of personnel to assure that surveys are being performed.

8303220225 830317
NMS LIC30
37-00358-04 PDR

28 February 1983

3) Date when full compliance will be achieved:

a) 3/1/83

B) Failure to evaluate personnel exposure when monitoring devices are damaged.

1) Corrective Steps Taken

a) An estimate of the missing exposure data was done by averaging the technologists' prior 12 month exposure data. This information has been added to the technologist exposure history.


2) Corrective Steps Which Will Be Taken

a) A standard form has been developed for updating of personnel exposure histories, and will be used when personnel monitoring devices are lost or damaged.

3) Date when full compliance will be achieved:

a) 3/1/83

Sincerely yours,



John C. McGreevy
Administrator
Polyclinic Medical Center

JCM/JRF/crf

cc: Milton A. Friedlander, M.D.
Robert M. Olley, R.T.