

St. Therese Medical Park
St. Therese Hospital

2615 Washington Street
Waukegan, Illinois 60085
(312) 688-5800



*Conducted by Missionary Sisters
Servants of the Holy Spirit*

St. Therese
Area Trauma Satellite

37809 North Route 59
Lake Villa, Illinois 60046
(312) 356-6600

March 9, 1983

D. G. Wiedeman, Chief
Materials Radiation Protection
U.S. Nuclear Reg. Comm.
Region III
799 Roosevelt Road
Glen Ellyn, IL 60137
RE: Inspection of License #12-10655-01

Dear Mr. Wiedeman:

This is in response to your March 4, 1983 report concerning the February 18, 1983 NRC inspection of our facilities. We will respond to each item of non-compliance using the same numerical order as your letter.

1. Quarterly linearity checks and daily Cs-137 constancy checks on our dose calibrator are now being performed at the required frequencies. Compliance with this item is immediate.
2. Wipe tests on the external surface of final source containers are now being performed according to Appendix F of the NRC licensing guide. Compliance with this item is immediate.
3. Radiation surveys and wipe tests are now being performed at the frequencies specified in Appendix I of the NRC licensing guides. Compliance with this item is immediate.

Since each of the above three (3) items involves "gaps" in the required records, we have prepared a "records audit" checklist which the radiation safety officer will periodically use to spot check the nuclear medicine records. We plan to initially use this checklist on approximately a months frequency until we are certain no further non-compliance will occur.

8303220246 830317
NMS LIC30
12-10655-01 PDR

St. Therese Medical Park
St. Therese Hospital

2615 Washington Street
Waukegan, Illinois 60085
(312) 688-5800



*Conducted by Missionary Sisters
Servants of the Holy Spirit*

St. Therese
Area Trauma Satellite

37809 North Route 59
Lake Villa, Illinois 60046
(312) 356-6600

Thank you for your attention to this matter, and we look forward to receiving acknowledgement from your office that these corrective actions are satisfactory until re-inspection is made.

Sincerely,

Piyush I. Vyas, M.D.
Radiation Safety Officer

PIV:kj

cc: Sister Veronica Marie
Dr. P. Vyas
Stan Huber Consultants
Mr. Howard Kanous