

Docket No. 30-14644

License No. 18-01760-04

SEP 23 1982

Pineland Center  
ATTN: George A. Zitnay  
Superintendent  
Box C  
Pownal, Maine 04069

Gentlemen:

Subject: Inspection No. 30-14644/82-01

This refers to your letter dated August 5, 1982, in response to our letter dated July 23, 1982.

Thank you for informing us of the corrective and preventive actions documented in your letter. These actions will be examined during a future inspection of your licensed program.

If you wish to use iodine-125 under the General License in 10 CFR 31.11, and if you have not already done so, please complete the enclosed NRC form 483 and submit it in accordance with the instructions on the form. When you receive the numbered copy of Form 483, you may write to Dr. John Glenn, Chief, Materials Program Section No. 2, at this office to request that the authorization for iodine-125 be deleted from License No. 18-01760-04. Please note that the General License limits possession of iodine-125 to 200 microcuries at any one time, and exempt you from the requirements of 10 CFR Part 20, including the waste disposal requirements in section 20.301.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By:  
John D. Kinneman

*for* Thomas T. Martin, Director  
Division of Engineering and  
Technical Programs

Enclosure: NRC Form 483  
10 CFR Part 31

cc w/encl:  
Robert B. Tanguay, Ph.D., Director of Laboratories  
Public Document Room (PDR)  
Nuclear Safety Information Center (NSIC)  
State of Maine

OFFICIAL RECORD COPY

8209280206 820923  
NMS LIC30  
18-01760-04 PDR

1E:07

SEP 23 1982

bcc w/encl:  
Region I Docket Room (w/ concurrences)

*[Signature]*  
RI:DETP  
Nicolosi:as  
9/16/82

RI:DETP  
Glenn  
9/23/82

*[Signature]*  
RI:DETP  
Kinneman  
9/23/82