

GRACE HOSPITAL

2307 WEST 14TH STREET
CLEVELAND, OHIO 44113

March 15th, 1983

License No. 34-16583-01


NUCLEAR REGULATORY COMMISSION
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137
Attention: D.J. Sreniawski, Chief
Materials Radiation Protection, Section 2

Dear Mr. Sreniawski,

Your letter of March 7th has been received. The items referenced in your letter were discussed with Mr. Pillai, our consultant physicist who has performed the leak-testing. The intervals between testing had exceeded 6 months on several occasions due to scheduling problems. We have requested Mr. Pillai to schedule examinations at a more frequent interval so that in the event that a scheduled examination has been missed, he can be contacted to perform the required studies within the maximum of a 6 month interval.

We have requested Mr. Pillai to write us to indicate the steps he feels may be appropriate to ensure that the leak-testing intervals do not exceed 6 months. If your department desires, we shall also inform you as to the dates of leak-testing with a brief report to be filed on a yearly basis. From a standpoint of date of full compliance, we considered that today is the date that compliance has been achieved.

Sincerely,



Alvin Segel, M.D.
Director Department of Radiology

AS/bc

8303280149 830323
NMS LIC30
34-16583-01 PDR

MAR 21 1983