

Appendix

NOTICE OF VIOLATION

Bio-Metric Systems, Incorporated

License No. 22-20307-01

As a result of the inspection conducted on March 7, 1983, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violation was identified:

License Condition No. 14 requires that licensed material be possessed and used in accordance with statements, representations, and procedures contained in application dated August 2, 1982 and letter dated September 3, 1982.

Item No. 7 of the above letter states that incoming packages containing radioactive material will be monitored with a GM survey instrument and final source containers will be smear tested for contamination.

Contrary to these requirements, packages received at your facility containing radioactive material are not surveyed or smear tested.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

March 18, 1983
Dated

D. G. Wiedeman
D. G. Wiedeman, Chief
Materials Radiation Protection
Section 1

8303280245 830322
NMS LIC30
22-20307-01 FDR