

Appendix

NOTICE OF VIOLATION

Ashtabula General Hospital

License No. 34-01386-01

As a result of the inspection conducted on September 28, 1982, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violation was identified:

License Condition No. 16 requires that licensed material be possessed and used in accordance with statements, representations, and procedures contained in various dated applications and letters.

Letter dated March 16, 1978, Item No. 2, states that weekly radiation level surveys are supplemented with a series of wipe tests to measure contamination levels, as outlined in Appendix I of the Draft Medical Licensing Guide.

Contrary to this requirement, area wipe tests have not been performed at your facility since the inception of this requirement.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated

10/12/82

D. J. Sreniawski
D. J. Sreniawski, Chief
Materials Radiation Protection
Section 2