

VOID SHEET

TO: License Fee Management Branch

FROM:

*Peter J Lee*

SUBJECT:

VOIDED APPLICATION

Control Number:

96623

Applicant:

Wayne State Univ.

Date Voided:

4-19-94

Reason for Void:

Combined with control no. 96633  
96635

*Peter J Lee* 4-28-94  
Signature Date

Attachment:  
Official Record Copy of  
Voided Action

FOR LFMB USE ONLY

Final Review of VOID Completed:

- Refund Authorized and processed
- No Refund Due
- Fee Exempt or Fee Not Required

Comments:

9406130179 940419  
PDR ADOCK 03001995  
C PDR

Log completed

Processed by:

LAC  
5/10/94

190124

ML  
30  
DH

(FOR LFMS USE)  
INFORMATION FROM LTS

BETWEEN:

LICENSE FEE MANAGEMENT BRANCH, ARM  
AND  
REGIONAL LICENSING SECTIONS

PROGRAM CODE: 03800  
STATUS CODE: 2  
FEE CATEGORY: EX 3P  
EXP. DATE: 19940228  
FEE COMMENTS: 170.11(A)(4)  
DECOM FIN ASSUR REQD: N  
.....

LICENSE FEE TRANSMITTAL

A. REGION

1. APPLICATION ATTACHED  
APPLICANT/LICENSEE: WAYNE STATE UNIVERSITY  
RECEIVED DATE: 940228  
DOCKET NO: 3006964  
CONTROL NO.: 396623  
LICENSE NO.: 21-00741-09  
ACTION TYPE: AMENDMENT

2. FEE ATTACHED  
AMOUNT: \_\_\_\_\_  
CHECK NO.: 9

3. COMMENTS

SIGNED P. Stclaff  
DATE 3-27-94

B. LICENSE FEE MANAGEMENT BRANCH (CHECK WHEN MILESTONE 03 IS ENTERED / )

1. FEE CATEGORY AND AMOUNT: EX 3P FEE EXEMPT

2. CORRECT FEE PAID. APPLICATION MAY BE PROCESSED FOR:  
AMENDMENT  \_\_\_\_\_  
RENEWAL \_\_\_\_\_  
LICENSE \_\_\_\_\_

3. OTHER \_\_\_\_\_

SIGNED SL  
DATE 3/22/94

RECEIVED  
APR 04 1994  
REGION III

APR 13 1994



Wayne State University

Health Physics-Radiation Control  
645 Mullett  
Detroit, Michigan 48226

February 21, 1994

Mr. Peter Lee  
U.S. Nuclear Regulatory Commission  
Licensing Section  
801 Warrenville Rd.  
Lisle, Illinois 60532-4351

RECEIVED BY LFDCB	
Date	Mar. 21, 1994
Log	Mar 17 III
By	SC
Date Completed	3/22/94

Dear Mr. Lee;

This letter serves to confirm our conversation on February 17, 1994 regarding an unused, in storage only, cobalt-60 irradiator at Wayne State University, Detroit, MI which we wish to relocate on campus.

Specifically, the building in which this irradiator is currently housed is scheduled for demolition as soon as arrangements can be made for irradiator removal.

The unit is in a Bar-Ray Model R-1001-K licensed under N.R.C. License #21-00741-09 containing approximately 11.07 Ci of Cobalt-60.

Since the State of Michigan generators cannot at present dispose of any solid waste, we will be forced to store this irradiator at our central waste storage facility located at 6000 Cass and approximately 8 blocks from the current irradiator facility.

Therefore, it is our intent to hire a professional commercial rigging company to remove this unit (non-source removal - that is, the source will remain inside its shielding housing) and transport it to 6000 Cass.

Riggers working on this unit will be monitored for any potential radiation exposure. During the entire moving procedure, Mr. Richard D. Cummings, R.S.O., and Director of Health Physics with 24 years experience in Health Physics and Mr. James L. Barrows, Health Physicists II with 32 years experience in Health Physics will be present with appropriate monitoring equipment to monitor this transfer. Again, the source will never leave our possession, nor the campus.

Based on the aforementioned, my question is, may we hire a commercial rigging company to perform these activities? If so, does this company have to be licensed by the N.R.C.?

RECEIVED

FEB 28 1994

REGION III

FEB 21 1994

FEE EXEMPT


170.11(a)(4)

CONTROL NO. 396623

In addition, what are your required procedures that we must comply with?

I thank you in advance for assisting us with these activities.

Sincerely,



James L. Barrows  
Health Physicists II  
Wayne State University  
Health Physics Department  
645 Mullett  
Detroit, MI 48226  
(313) 577-1255

JLB/de

CONTROL NO. 396623



Wayne State University  
Health Physics - Radiation Control  
45 Mullett  
Detroit, Michigan 48226

PRESORTED  
FIRST-CLASS



Mr. Peter Lee  
Nuclear Regulatory Commission  
Licensing Section  
801 Warrenville Rd.  
Lisle, IL 60532-4351



CONVERSATION RECORD

TIME 10:00A DATE 3/14/94

VISIT  CONFERENCE  TELEPHONE

INCOMING  
 OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT: Jim Barrows  
ORGANIZATION (OFFICE, DEPT.ETC.): Wayne University  
TELEPHONE NO.:

SUBJECT  
License expiration

SUMMARY  
I spoke to Jim Barrows about the fact the it appeared that his license expired. Mr. Barrows said that the amendment request that he sent in (received by us on 2/28/94) was meant to be a renewal request. They were suprised that the license had expired, since they had sent in the amendment request.

I told him we would process his letter as an amendment request to extend the expiration date for one year. He said that will be acceptable to him because they want to add this irradiator to their broadscope license.

ACTION REQUIRED  
Process in letter as an amendment request.

NAME OF PERSON DOCUMENTING CONVERSATION: Pat Vacherlon  
SIGNATURE: [Signature]  
DATE: 3/14/94

ACTION TAKEN

SIGNATURE TITLE DATE

CONTROL NO. 396623