

whitworth , foust + stearnes ^{inc.}

architects · engineers · planners

lennie o. whitworth jr., p.e. byron foust, a.i.a. robert stearnes, a.i.a.

September 24, 1982

Mr. D.G. Wiedeman, Chief
Materials Radiation Protection
Section 1
United States Nuclear Regulatory
Commission
Region III
799 Roosevelt Road
Glen Ellyn, IL 60137


RE: License No. 24-20197-01

Dear Mr. Wiedeman:

The following explanations are in reply to the noncompliance item mentioned in the appendix to your letter of August 31, 1982.

- 1) A leak test was performed as mentioned on August 11th, 1982 and the results which are enclosed were received recently.
- 2) The next leak test is scheduled to be performed in February of 1983. A new leak test kit has been ordered so that the test can be performed on time.
- 3) Since the above two items have been initiated, I feel that we are presently in compliance.

Yours truly,



Carl J. Muench, Jr.,
Radiation Protection Officer

CJM/sb

8210250123 821013
NMS LIC30
24-20197-01 PDR

Carl J. Muench Jr., P.E. Associate

412 tanner, box 548, sikeston, missouri 63801 pho. 314 471 4660
626 broadway, box 678, cape girardeau, missouri 63701 pho. 314 334 1329

SEP 27 1982



CPN CORP
130 SO BUCHANAN CIRCLE
PACHECO, CA 94553
415-687-6472

LEAK TEST CERTIFICATE

(Customer mailing address)

Blue = Keep with Gauge
White = RSO file copy
Pink = CPN customer file copy

Attn: Mr. John E. Hill, P.E.
Whitworth, Foust & Stearnes, Inc.
P.O. Box 548
Sikeston, MO 63801

RADIATION SAFETY OFFICER:

One copy of the leak test certificate should be maintained with the device at all times. One copy should be maintained in the Radiation License File at all times for inspection.

The device listed hereon has been leak tested on the date indicated and has less than 0.005 microcuries contamination.

DEVICE Portaprobe SERIAL M21053814

ISOTOPE: CS137 10 mCi AM241 50 mCi RA226 mCi Other mCi

DATE OF TEST 8/11/82

SIGNED

Jim M. Shull
CPN Production Control

Check the applicable license for the appropriate leak test period. This will be six months, one year, or three years depending upon type of device and jurisdictional agency. All PORTAPROBE devices are approved for a one year leak test by the State of California Department of Public Health and this approval is recognized by most other licensing agencies.

Order your next leak test kit from CPN, Part # A-401197.

You will be notified automatically, one month prior to the next required leak test, by CPN if you use the CPN leak test service.