

Appendix

NOTICE OF VIOLATION

Pittsburgh Testing Laboratory

License No. 37-00276-25

As a result of the inspection conducted on August 10, 1982, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violations were identified:

1. License Condition No. 10 states the authorized place of use includes temporary job sites of the licensee throughout the United States and at off-shore drilling platforms located in the Gulf of Mexico. The licensee may also store byproduct material at the following location:

Pittsburgh Testing Laboratory
1715 Brookpark Road
Cleveland, Ohio 44109

Contrary to the above, nuclear moisture and density gauges are being stored at 4734 Spring Road, Brooklyn Heights, Ohio, a location not authorized by the license.

This is a Severity Level IV violation (Supplement VI).

2. License Condition No. 20 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced applications and letters.

The referenced letter dated April 29, 1982, states in the enclosed attachment titled "Radiation Safety Procedures for Use of Nuclear Surface Moisture Density Gauges", dated January 1, 1980, that a survey meter will be available at the laboratory for making radiation surveys of each nuclear unit.

Contrary to the above, on the day of the inspection a survey meter was not available at the Brooklyn Heights laboratory where several gauges were being stored.

This is a Severity Level IV violation (Supplement VI).

8210250172 821013
NMS LIC30
37-00276-25 PDR

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

OCT 13 1982

Dated

ORIGINAL SIGNED BY
JAMES R. MILLER

J. R. Miller, Chief
Technical Inspection Branch