



Franklin
Medical Center

A Member of Baystate Health Systems

June 9, 1994

U.S. Nuclear Regulatory Commission
ATTN: Document Control Desk
Washington, D.C. 20555

RE: License No. 20-02399-01

REPLY TO A NOTICE OF VIOLATION


ITEM A.

In July, 1991 a nuclear pharmacy computer system was installed in our facility. This system very effectively tracks health physics requirements for our license. Unfortunately the system was programmed to flag sealed source inventory every 6 months.

This program error was corrected on May 4, 1994 during our inspection with Penny Nessen, and a sealed source inventory was performed on that day. Heretofore the computer system will trigger every 90 days for a source inventory, and will require this function as a mandatory procedure.

ITEM B.

A license amendment has been applied for under separate cover (copy enclosed).



Harlan J. Smith, President

BM/r=20H

Enclosure: License Amendment Request
cc: Regional Administrator

9406200136 940609
PDR ADDCK 03001846
C PDR

104 STREET, GREENFIELD, MASSACHUSETTS 01301

TELEPHONE (413) 772-0211

TEO 1
11