

October 18, 1982

In Reply Refer To: 566/115

Thomas T. Martin, Director
Division of Engineering
and Technical Programs
United States Nuclear Regulatory
Commission
631 Park Avenue
King of Prussia, PA 19406

Subject: Inspection No. 82-01, Docket No. 30-01788, License No. 19-00975-01.

As a result of the inspection conducted on August 23, 1982, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), certain violations were identified. In compliance with established regulations and notice of violation, appendix A to your letter of October 5, 1982, the following actions were taken:

1. Item 3.B of an attachment to this application entitled, "Dose Calibrator Procedures," requires that dose calibrator linearity be determined at installation, and quarterly thereafter.

a. Corrective steps taken - dose calibrator linearity check has been done and will be done quarterly hereafter. This is done along with the daily calibration using 57 Cobalt, and the yearly calibration using 137 Cesium and 133 Barium.

b. Avoiding further violations - follow up supervision.

c. Achievement date - August 24, 1982.

2. An attachment to this application entitled, "Laboratory Rules for the Use of Radioactive Material," requires that nuclear medicine personnel wear laboratory coats at all times when working in areas where radioactive materials are used.

a. Corrective steps taken - personnel are wearing laboratory coats or other protective clothing when working in areas where radioactive materials are being used.

b. Avoiding further violations - follow up supervision.

c. Achievement date - August 24, 1982.

2.


3. An attachment to this application entitled, "Survey Procedures," requires that a permanent record be kept of all surveys, including negative results.

a. Records of the package surveys and wipe tests that are done immediately upon the receipt of our pre-calibrated doses are being kept. Personnel monitoring areas have been done and presently are being done weekly. These results are kept in the files.

b. Avoiding further violations - follow up supervision.

c. Achievement date - August 24, 1982.

4. Based on attentive evaluation and implementation, we feel we are fully in compliance with all NRC requirements as set forth in Notice of Violation.


FOR ROBERT HAITH, JR.
Director