



UNITED STATES
NUCLEAR REGULATORY COMMISSION

REGION IV

611 RYAN PLAZA DRIVE, SUITE 400
ARLINGTON, TEXAS 76011-8064

JUN 17 1994

Docket: 030-33170
License: 35-27468-01

Regional Medical Laboratory, Inc.
ATTN: C. Terrence Dolan, M.D.
Laboratory Director
1923 South Utica Avenue
Tulsa, Oklahoma 74104

SUBJECT: NRC INSPECTION REPORT 030-33170/94-01

This refers to the routine, announced inspection conducted by Ms. M. Linda McLean of this office on May 26, 1994. The inspection included a review of activities authorized by Byproduct Materials License 35-27468-01. At the conclusion of the inspection, the findings were discussed with members of your staff.

The inspection was an examination of activities conducted under the license as they relate to radiation safety and to compliance with the Commission's rules and regulations and the conditions of the license. The inspection consisted of selective examinations of procedures and representative records, interviews of personnel, independent measurements, and observation of activities in progress.

During this initial inspection, the inspector discussed with your staff the requirements with which you must comply, including NRC and Department of Transportation regulations, and the commitments in your license application. The NRC enforcement policy was also discussed at this time.

No violations or deviations were identified; therefore, no response to this letter is required.

The inspector noted that you possessed one blood irradiator used for irradiation of human blood and blood products. The irradiator, located in the blood bank laboratory within St. Johns Medical Center, contained approximately 1400 curies of cesium-137.

In accordance with 10 CFR 2.790 of the NRC's "Rules of Practice," a copy of this letter will be placed in the NRC Public Document Room.

2100 27

9406220097 940617
PDR ADOCK 03033170
C PDR

JE07/0

Regional Medical Laboratory, Inc. -2-

Should you have any questions concerning this letter, please contact the inspector identified above at (817) 860-8100.

Sincerely,

Charles L. Cain

Charles L. Cain, Acting Chief
Nuclear Materials Inspection Branch

cc:
Oklahoma Radiation Control Program Director

bcc:
DMB - Original (IE-07)
LJCallan
SJCcollins
RAScarano
DWeiss, OC/LFDCB (T-9E10)
*WLFisher
*CLCain
*FAWenslawski
*MLMcLean
*NMIB
*MIS System
*RIV Files (2)

*W/IFS Form

RIV:NMIB	AC:NMIB <i>CLC</i>			
MLMcLean <i>MLM</i>	CLCain			
06/15/94	06/16/94			

bcc:

DMB - Original (IF-07)

LJCallan

SJCollins

RAScarano

DWeiss, OC/LFDCB (T-9E10)

*WLFisher

*CLCain

*FAWenslawski

*MLMcLean

*NMIB

*S System

*RIV Files (2)

*W/IFS Form

RIV:NMIB	AC:NMIBCAC			
MLMcLean <i>MLM</i>	CLCain			
06/15/94	06/16/94			