

TE07
PUBLIC

JUN 2 1994

St. Vincent Hospital & Health
Care Center
ATTN: Shih-Shen Ling, Ph.D.
Radiation Safety Officer
2001 West 86th Street
Indianapolis, IN 46240-0970

License No. 13-00133-02
Docket No. 030-01579

Dear Dr. Ling:

As a result of the inspection conducted on April 18, 1994, the enclosed NRC Form 591, SAFETY INSPECTION, is issued for License No. 13-00133-02. The form sets forth the violations noted. Please acknowledge receipt of this form by signing and dating in the appropriate space on all copies. You are requested to retain the original and return four signed and dated copies to this office within ten days of the receipt of this letter.

I wish to express my appreciation for the cooperation extended to me during the inspection.

Sincerely,

Original Signed by

T. L. Simmons
Radiation Specialist

Enclosure: NRC Form 591

bcc w/enclosure:
PUBLIC

RIII
Simmons/jaw
6/31/94

RIII
Holt
6/1/94

9406220171 940602
PDR ADOCK 03001579
C PDR

SAFETY INSPECTION

<p>1. LICENSEE</p> <p>St. Vincent Hospital & Health Care Ctr. 2001 West 86th Street Indianapolis, IN 46240-0970</p>	<p>2. REGIONAL OFFICE</p> <p>REGION III U S NUCLEAR REGULATORY COMMISSION 801 WARRENVILLE ROAD LISLE IL 60532-4351</p>	
<p>3. DOCKET NUMBER(S)</p> <p>030-01579</p>	<p>4. LICENSE NUMBER(S)</p> <p>13-00133-02</p>	<p>5. DATE OF INSPECTION</p> <p>April 18, 1994</p>

LICENSEE:
The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:

- 1. Within the scope of this inspection, no violations were observed.
- 2. The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on those actions at this time.
- 3. During this inspection certain of your activities, as described below or attached, were in violation of NRC requirements. This form is a **NOTICE OF VIOLATION**, which is required to be posted in accordance with 10 CFR 19.11.
 - A. _____ was not properly posted to indicate the presence of a _____, 10 CFR 20.203(b),(c),(d),(e) or 34.42.
 - B. _____ of sealed sources were not performed at the proper frequencies. 10 CFR _____ or License Condition Number _____.
 - C. Records of _____ were not properly maintained. 10 CFR _____ or License Condition Number _____.
 - D. Documents were not properly posted or otherwise made available. 10 CFR 19.11.
 - E. Reports or notification of _____ were not made in accordance with 10 CFR _____ or License Condition Number _____.
 - F. The Victoreen 450 survey instrument has not been checked with a dedicated check source each day of use. 10 CFR 35.51 (c)

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified in the items checked above. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.

SIGNATURE - LICENSEE	DATE	SIGNATURE - NRC INSPECTOR <i>Jay Foxworth</i>	DATE 5/25/94
----------------------	------	--	-----------------