



# CHARLTON MEMORIAL HOSPITAL

HIGHLAND AVENUE AT NEW BOSTON ROAD  
FALL RIVER, MASSACHUSETTS 02720  
(617) 679-3131

September 9, 1987

John R. White, Chief  
Nuclear Materials Safety Section C  
Division of Radiation Safety & Safeguards  
Region I  
631 Park Avenue  
King of Prussia, PA 19406

RE: Docket No. 030-01875  
License No. 20-05691-01

Dear Mr. White:

We are writing this letter in response to violations found during an inspection on May 27, 1987.

ITEM 1

Radiation Safety Committee Meetings will be held as required but not less than once per quarter. The last meeting was held on 9/2/87.

ITEM 2

Constancy checks are now performed on each day that the dose calibrator is used for measurement of clinically used doses.

ITEM 3

It is now being documented that the final source container for packages is being wipe tested and the package itself is being surveyed before disposal.

ITEM 4

The wipe test results are now being recorded in terms of dpm. It has been verified that the Geiger counter used for this purpose is sensitive enough to detect 100 dpm and a conversion factor is applied to the cpm reading.

As of the date of this letter of 9/9/87 we are in full compliance.

Sincerely,

8801120372 871228  
REG1 LIC30  
20-05691-01 PDCD