

APPENDIX

NOTICE OF VIOLATION

Sooner Pride Services, Inc.
Hominy, Oklahoma 74035

Docket: 030-17803
License: 35-19505-01

During an NRC inspection conducted on May 26, 1994, a violation of NRC requirements was identified. In accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, the violation is listed below:

10 CFR 39.35(c) requires, in part, that each sealed source be tested at intervals not to exceed 6 months.

Contrary to the above, from December 4, 1992 to August 10, 1993, an interval greater than 6 months, the licensee did not leak test its 2 curie cesium-137 sealed source.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, Sooner Pride Services, Inc., is hereby required to submit a written statement or explanation to the U.S. Nuclear Regulatory Commission, ATTN: Document Control Desk, Washington, D.C. 20555, with a copy to the Regional Administrator, Region IV, 611 Ryan Plaza Drive, Suite 400, Arlington, Texas 76011, within 30 days of the date of the letter transmitting this Notice of Violation (Notice). This reply should be clearly marked as a "Reply to a Notice of Violation" and should include for each violation: (1) the reason for the violation, or, if contested, the basis for disputing the violation, (2) the corrective steps that have been taken and the results achieved, (3) the corrective steps that will be taken to avoid further violations, and (4) the date when full compliance will be achieved. If an adequate reply is not received within the time specified in this Notice, an order or a Demand for Information may be issued to show cause why the license should not be modified, suspended, or revoked, or why such other action as may be proper should not be taken. Where good cause is shown, consideration will be given to extending the response time.

Dated at Arlington, Texas
this 17th day of June, 1994

9406220204 940617
PDR ADOCK 03017803
C PDR

bcc:
DMB - Original (IE-07)
LJCallan
SJCcollins
RAScarano
DWeiss, OC/LFDCB (T-9E10)
*WLFisher
*CLCain
*FAWenslawski
*MLMcLean
*NMIB
*MIS System
*RIV Files (2)

*W/IFS Form

RIV:NMIB	AC:NMIB <i>CLC</i>	DD:DRSS	D:DRSS	
MLMcLean <i>W</i>	CLCain	RAScarano	SJCcollins	
06/15/94	06/16/94	06/ /94	06/ /94	

IFS Data Entry Form

Reviewed By: UAC

Date: 6/16/94

Reactor/Vessel Inspection (IFS Option 1) Docket Related/P21 Items (IFS Option 4)
 Items Opened (Y/N): _____ Enter Log Number: _____
 Materials Inspection (IFS Option 2) LER Items (IFS Option 5)
 SR (Y/N): N Enter LER Number: _____
 LER (Y/N): Y
 Clear (Y/N): N Non-Docket Related Items (IFS Option 6)

Site Name: SOONER POWER SERVICES

Report Transmitted Date: 06/17/94

Lead Inspector: MJM Responsible Org. Code: 43104 Report End Date: 5/26/94 Region: 4

	Report NBR	Docket NBR
A	<u>94001</u>	<u>030-17803</u>
B	_____	_____
C	_____	_____

Materials Only
License NBR
<u>35-19505-01</u>

Docket Name

Update? (Y/N): _____ Opened IR/LEP/P21 LOG/IFS Number: _____

***Sequence NBR: 01 Item Type: VIO **Severity: 4 **Supplement: 6

Status	*UPD VR	*Proj. Closeout	*Actual Closeout	Materials Only	License Cond.	Tie Down
A	<u>0</u>	<u>1/1</u>	<u>1/1</u>	10 CFR	_____	_____
B	_____	<u>1/1</u>	<u>1/1</u>	<u>39.35</u>	_____	_____
C	_____	<u>1/1</u>	<u>1/1</u>	_____	_____	_____

Title: FAILURE TO PERFORM LEAK TEST AT REQUIRED FREQ (55 character width)

*Closeout Org: _____ *Closeout EMP: _____ *Contact EMP: _____ *Procedure: _____ *Funct Area: _____
 *Cause CD: _____ **EA Number: _____ **NOV/NNC Issue Date: 06/17/94

Text: _____

See attached violation highlighted

Update? (Y/N): _____ Opened IR/LEP/P21 LOG/IFS Number: _____

***Sequence NBR: 02 Item Type: VIO **Severity: 4 **Supplement: _____

Status	*UPD VR	*Proj. Closeout	*Actual Closeout	Materials Only	License Cond.	Tie Down
A	<u>0</u>	<u>1/1</u>	<u>1/1</u>	10 CFR	_____	_____
B	_____	<u>1/1</u>	<u>1/1</u>	_____	_____	_____
C	_____	<u>1/1</u>	<u>1/1</u>	_____	_____	_____

Title: _____ (55 character width)

*Closeout Org: _____ *Closeout EMP: _____ *Contact EMP: _____ *Procedure: _____ *Funct Area: _____
 *Cause CD: _____ **EA Number: _____ **NOV/NNC Issue Date: 1/1

Text: _____

* Optional Fields.

** Severity, Supplement, and NOV/NCC only applicable for Violations; EA Number only applicable for Apparent Violations.

*** Sequence NBR is not applicable for docket related/P21, LER, or non-docket related items.

ITEMS CONTINUED? (Y/N): _____