

FLEXEL, INC.

Highway 136 West
Covington, Indiana 47932
317/793-2211

Sumner

FLEXEL

March 30, 1994

VIA CERTIFIED MAIL

U.S. Nuclear Regulatory Commission
Region III
401 Warrenville Road
Lisle, IL 60532-4351

Dear NRC Representative:

Re: Reply to a Notice of Violation
License No. 13-24621-01

This letter is in response to the Notice of Violation dated March 19, 1994 regarding an inspection conducted at Flexel Indiana, Inc. in Covington, Indiana on January 20, 1994. There were two violations requiring some action.

1. Two spare krypton-85 gauging devices were located in an electrical workshop, which you feel is not secure against unauthorized removal or access. Even though we feel otherwise, the two spares have been placed in a lead lined crate and are now stored in the electrical supervisor's office, which is kept locked during the off shifts. The box has been labeled as containing radioactive material.
2. We could not provide evidence that the survey meter had been calibrated since 1984. We state in our license application that the meter will be calibrated periodically, which you are interpreting to mean annually. The meter was calibrated on February 14, 1994, and will be calibrated on an annual basis in the future.

We feel that full compliance has been achieved. If you have any further questions, please contact me at 317-793-2211, extension 202.

Sincerely,

Denise K. Vollmer

Denise K. Vollmer
Sr. Environmental Engineer
Radiation Safety Officer

9406220310 940609
PDR ADOCK 03018687
C PDR

"Excellence in Flexible Packaging"

APR 5 1994