

NOTICE OF VIOLATION

St. Luke's Hospital
Maumee, Ohio

Docket No. 030-12120
License No. 34-17056-01

As a result of the inspection conducted on November 26, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1990) (Enforcement Policy) the following violations were identified:

1. License Condition No. 14 states that the license is based on statements and representations contained in certain referenced applications and letters.

The referenced application dated July 11, 1988, states that the safety rules described in Appendix I of Regulatory Guide 10.8 will be followed. Appendix I states that after each procedure or before leaving the area, hands will be monitored for contamination.

Contrary to the above, hands were not always monitored*for contamination since the license was issued on September 20, 1988.

This is a Severity Level IV violation (Supplement VI)

2. License Condition No. 14 states that the license is based on statements and representations contained in certain referenced applications and letters.

The referenced application dated July 11, 1988, states that the safety rules described in Appendix I of Regulatory Guide 10.8 will be followed. Appendix I states that doses that differ from the prescribed dose by more than 10% will not be used.

Contrary to the above, doses that differed from the prescribed dose by more than 10% were used. For example, technetium-99m MDP doses that were from 25 to 33% greater than the prescribed dose were used on October 17 and 18, 1990.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the

corrective steps that will be taken to avoid further violations; and
(3) the date when full compliance will be achieved. Consideration may
be given to extending your response time for good cause shown.

JAN 3 1991

Dated _____

William H. Schultz

William H. Schultz, Chief
Nuclear Materials Safety
Section 1