

NOTICE OF VIOLATION

Portage View Hospital
Hancock, Michigan

License No. 21-18525-01
Docket No. 030-13827

As a result of the inspection conducted on December 17, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1990) (Enforcement Policy) the following violation was identified:

License Condition No. 16.A states that the license is based upon the licensee's statements and representations in certain referenced documents.

The referenced application dated June 6, 1988, states in Item 8.1 that Appendix A of Regulatory Guide 10.8, Revision 2 will be followed for training of ancillary personnel (nursing) whose duties may require them to work in the vicinity of radioactive material.

Appendix A states that ancillary personnel (nursing) will be instructed about radiation hazards and appropriate precautions during annual refresher training.


Contrary to the above, annual refresher training for ancillary personnel (nursing) was not performed since approximately October 1989.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

JAN 7 1991

Dated _____



Roy J. Caniano, Chief
Nuclear Materials Safety
Section 2

9101110151 910107
REG3 LIC30
21-18525-01 PDR