



A Non-Profit Osteopathic Facility † Directed by a Volunteer Board of Directors

oak hill hospital

(417) 623-4640

THIRTY-FOURTH and INDIANA AVENUE

JOPLIN, MISSOURI 64801

KEITH ADAMS Administrator

November 15, 1982

Mr. D. G. Wiedeman, Chief
Material Radiation Protection
U. S. Nuclear Regulatory Commission
799 Roosevelt Road
Glen Ellyn, IL 60137

RE: #24-16612-01

Dear Mr. Wiedeman:

Pursuant to your correspondence of October 19, 1982 concerning the inspection of our facility, we offer the following:

1. The ¹³⁷Cesium source #NSH 1318 has been re-tested and the report should be available to us within 30 days. This source was tested on a timely basis but there is no record of the result; in the future we will audit at time of testing. Full compliance was achieved October 29, 1982.
2. The quarterly linearity test has been implemented, In the future, this will be on the agenda of our quarterly Radiation Safety Committee meeting to assure compliance. Full compliance was achieved October 29, 1982.

Your kind attention is sincerely appreciated.

Very truly yours,

Allan Paris, Jr., D.O.
Radiation Safety Officer

lcc

cc: K. Adams, Administrator

8212080473 821202
NMS LIC30
24-16612-01 PDR

NOV 18 1982