

NOTICE OF VIOLATION

Bio-Metric Systems, Inc.
Eden Prairie, Minnesota

License No. 22-20307-01
Docket No. 030-18228

As a result of the inspection conducted on December 12, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1990) (Enforcement Policy) the following violations were identified:

1. License Condition No. 15 requires the licensee to conduct its program in accordance with the statements, representations, and procedures contained in certain referenced documents.

The referenced application dated September 30, 1987, states that loose contamination surveys will be done daily where millicurie amounts of radioactive materials are used.

Contrary to the above, no daily contamination surveys were done on January 29, October 11, November 8, 1988; and May 10, July 23, and September 13, 1990; days when millicurie amounts of radioactive materials were used.

This is a Severity Level IV violation (Supplement VI).

2. 10 CFR 20.401(b) requires in part, that each licensee maintain records of disposals made under 10 CFR Part 20.303.


Contrary to the above, as of December 12, 1990, the licensee did not maintain records of the disposals of licensed materials via the sanitary sewerage system.

This is a Severity Level V violation (Supplement IV).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

JAN 3 1991

Dated _____



Roy J. Caniano, Chief
Nuclear Materials Safety
Section 2

9101110161 910103
REG3 LIC30
22-20307-01 PDR