

LICENSEE EVENT REPORT

UPDATE REPORT - PREVIOUS

REPORT DATE 3/24/80

CONTROL BLOCK: ①

(PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

⑦ ⑧ ⑨ G A E I H 2 ⑭ ⑮ 2 0 0 - 0 0 0 0 0 0 - 0 0 0 ⑳ ㉑ ㉒ ㉓ ㉔ ㉕ ㉖ ㉗ ㉘ ㉙ ㉚ ㉛ ㉜ ㉝ ㉞ ㉟ ①

LICENSEE CODE LICENSE NUMBER LICENSE TYPE CAT 58

CON'T

⑦ ⑧ ⑨ 0 1 ⑩ L ⑪ ⑫ 6 0 5 0 0 0 3 6 6 ⑬ ⑭ 7 0 2 2 7 8 0 ⑮ ⑯ 8 1 1 2 3 8 2 ⑰ ⑱ ⑲

REPORT SOURCE DOCKET NUMBER EVENT DATE REPORT DATE

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES ⑩

⑦ ⑧ ⑨ 0 2 With Unit Two operating steady-state at 99 percent power, an instrument
0 3 check disclosed an erroneous reading on the "A" Drywell H2-O2 Analyzer
0 4 recorder. The "A" analyzer was declared inoperable. This event is
0 5 contrary to Tech. Specs. Section 3.3.6.4.a. The redundant "B" analyzer
0 6 was operable. The unit was placed in a 30-day LCO. The health and safety
0 7 of the public were not affected by this event. This event is repetitive
0 8 as last reported on LER No. 50-366/1980-017. ①

⑦ ⑧ ⑨ 0 9 ⑩ S ⑪ E ⑫ B ⑬ P U M P X X ⑭ C ⑮ Z ⑯ ⑰ ⑱ ⑲

SYSTEM CODE CAUSE CODE CAUSE SUBCODE COMPONENT CODE COMP. SUBCODE VALVE SUBCODE

⑦ ⑧ ⑨ 0 1 ⑩ 8 0 ⑪ 0 1 9 ⑫ 0 3 ⑬ X ⑭ 1 ⑮ ⑯ ⑰ ⑱ ⑲

LER/RO REPORT NUMBER EVENT YEAR SEQUENTIAL REPORT NO. OCCURRENCE CODE REPORT TYPE REVISION NO.

⑦ ⑧ ⑨ 0 1 ⑩ C ⑪ Z ⑫ Z ⑬ 0 0 0 0 ⑭ N ⑮ N ⑯ A ⑰ D 0 9 6 ⑱ ⑲

ACTION TAKEN FUTURE ACTION EFFECT ON PLANT SHUTDOWN METHOD HOURS ATTACHMENT SUBMITTED NPRD-4 FORM SUB. PRIME COMP. SUPPLIER COMPONENT MANUFACTURER

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS ⑲

⑦ ⑧ ⑨ 1 0 The cause of this event was determined to be diaphragm failure in the
1 1 "A" analyzer vacuum pump. The diaphragm was replaced. The analyzer was
1 2 calibrated and functionally tested satisfactorily. The analyzer was
1 3 returned to service within the Tech. Specs. time requirement. A new
1 4 analyzer and its associated vacuum pump will be installed upon receipt. ①

⑦ ⑧ ⑨ 1 5 ⑩ E ⑪ 0 9 9 ⑫ NA ⑬ A ⑭ Operator Observation ⑮ ⑯ ⑰ ⑱ ⑲

FACILITY STATUS % POWER OTHER STATUS METHOD OF DISCOVERY DISCOVERY DESCRIPTION

⑦ ⑧ ⑨ 1 6 ⑩ Z ⑪ Z ⑫ NA ⑬ NA ⑭ NA ⑮ ⑯ ⑰ ⑱ ⑲

ACTIVITY CONTENT RELEASED OF RELEASE AMOUNT OF ACTIVITY LOCATION OF RELEASE

⑦ ⑧ ⑨ 1 7 ⑩ 0 0 0 ⑪ 7 ⑫ NA ⑬ ⑭ ⑮ ⑯ ⑰ ⑱ ⑲

PERSONNEL EXPOSURES NUMBER TYPE DESCRIPTION

⑦ ⑧ ⑨ 1 8 ⑩ 0 0 0 ⑪ 40 ⑫ NA ⑬ ⑭ ⑮ ⑯ ⑰ ⑱ ⑲

PERSONNEL INJURIES NUMBER DESCRIPTION

⑦ ⑧ ⑨ 1 9 ⑩ Z ⑪ NA ⑫ ⑬ ⑭ ⑮ ⑯ ⑰ ⑱ ⑲

LOSS OF OR DAMAGE TO FACILITY TYPE DESCRIPTION

⑦ ⑧ ⑨ 2 0 ⑩ N ⑪ NA ⑫ ⑬ ⑭ ⑮ ⑯ ⑰ ⑱ ⑲

PUBLICITY ISSUED DESCRIPTION

8212090391 821123
PDR ADOCK 05000366
S PDR

NRC USE ONLY

NAME OF PREPARER S. B. TippsPHONE: (912) 367-7851