

APPENDIX A
NOTICE OF VIOLATION

Tennessee Valley Authority
Sequoyah Nuclear Plant

Docket Nos. 50-327 and 50-328
License Nos. DPR-77 and DPR-79

As a result of the inspection conducted on October 12-15, 1982 and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violation was identified.

Technical Specification 6.8.4.b requires that written procedures shall be established, implemented and maintained by the Radiological Hygiene Branch covering the Quality Assurance Program and environmental monitoring, using the guidance contained in Regulatory Guide 4.15, December 1977.

Procedure No. NR-OPS-FO-NRE-62.3 (Calibration of Radiation Monitors) which implements Technical Specification 6.8.4.b requires calibration of environmental radiation monitoring equipment at six (6) month intervals, and posting of calibration stickers on or near the monitor.

Contrary to the above, Procedure No. NR-OPS-FO-NRE-62.3 was not implemented in that:

1. Background radiation monitors and air monitor flow meters at all assigned air particulate/radioiodine monitoring stations were not calibrated within the required six (6) month interval in the case of stations LM-1, LM-2, and PM-1 through PM-8.
2. Calibration stickers documenting the most recent calibrations were not posted on or near the monitors as required.

This is a Severity Level V Violation (Supplement IV.E).

Pursuant to the provisions of 10 CFR 2.201, you are hereby required to submit to this office within thirty days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Date: _____