

DSS

MAR 21 1991

Docket No. 030-03111

License No. 37-09016-01

Citizens General Hospital
Nuclear Medicine Department
ATTN: Ronald J. Clearfield, M.D.
Radiation Safety Officer
651 Fourth Avenue
New Kensington, Pennsylvania 15068-6591

Gentlemen:

Subject: Inspection No. 030-03111/90-002

This refers to your letter dated November 9, 1990, in response to our 90-002 inspection conducted November 6, 1990.

Thank you for informing us of the corrective and preventive actions documented in your letter. These actions will be examined during a future inspection of your licensed program.

With respect to performing the linearity test to include the 100 to 200 millicuries dose range prior to the administration of therapeutic doses of iodine-131, your revised procedure to test linearity in those ranges does not relieve you from the requirement to perform linearity tests on a quarterly frequency. The inclusion of the testing of linearity in the therapeutic ranges is an additional test unless the therapeutic administration occurs concurrently with the quarterly linearity test.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By:
Mohamed M. Shanbaky

9104010031 910321
REG1 LIC30
37-09016-01 PDR

Mohamed M. Shanbaky, Chief
Nuclear Materials Safety Section A
Division of Radiation Safety
and Safeguards

cc:
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
Commonwealth of Pennsylvania

OFFICIAL RECORD COPY

RL CITIZENS - 0001.0.0
02/14/91

RETURN ORIGINAL TO / IE:07
REGION I

290066

bcc:
Region I Docket Room (w/concurrences)

RI:DRSS
Darden/bj
A.H. Darden
02/21/91

RI:DRSS
Shanbaky
MS
02/15/91

OFFICIAL RECORD COPY

RL CITIZENS - 0002.0.0
02/14/91