

Appendix

NOTICE OF VIOLATION

St. Elizabeth Hospital Medical Center

License No. 34-01131-01

As a result of the inspection conducted on February 10, 1983, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violations were identified:

1. License Condition No. 18 states that the licensee shall possess and use licensed material in accordance with statements, representations, and procedures contained in application dated September 4, 1979; and letter dated December 4, 1979.

Application dated September 4, 1979, states in attachment 2 that the Medical Isotope Committee shall meet as often as necessary to conduct its business but not less than once in each calendar quarter, and that written records of the meetings be maintained.

Contrary to the above, the Medical Isotope Committee had only met twice in 1981, and once in 1982. Specifically, the Medical Isotope Committee met on April 6, 1981, November 10, 1981, and October 12, 1982. The frequencies between these meetings exceed the required frequency of once in each calendar quarter.

This is a Severity Level IV violation (Supplement VI).

2. License Condition No. 18 states that the licensee shall possess and use licensed material in accordance with statements, representations, and procedures contained in application dated September 4, 1979; and letter dated December 4, 1979.

Application dated September 4, 1979, states in attachment 5 that calibration of survey meters shall be performed with radionuclide sources at least on an annual basis and after servicing.


Contrary to the above, your Thyac III survey meter, which is used on a daily basis, was last calibrated on September 30, 1981, a period of more than one year.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated

2/23/83


D. J. Sreniawski, Chief
Materials Radiation Protection
Section 2