



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION II  
101 MARIETTA STREET, N.W., SUITE 2900  
ATLANTA, GEORGIA 30323-0199

*Official Record Copy*

JUN 01 1994

RADIOLOGY ASSOCIATES, P.C.  
5205 LYNBATE COURT  
BURKE, VA 22015

ATTN: ROBERT A. FERRIS, M.D.

RE: Docket Number: 3019932  
License Number: 45-21209-01  
Plan File Date: 16-FEB-93  
Region Number: 2

Dear Dr. Ferris:

This refers to the review of your written Quality Management Program (QMP) submitted in accordance with 10 CFR 35.32. A review of the QMP was performed to determine whether policies and procedures have been developed to meet the objectives of the rule. Based on this submission, there appear to be significant weaknesses and potential substantial failure of your QMP to meet the objectives in 10 CFR 35.32 in that:

Regarding I-125 and /or I-131 > 30 Microcuries

The written directive must be an order for a specific patient, dated and signed by an authorized user or physician under the supervision of an authorized user, and, for any administration of quantities greater than 30 microcuries of either I-125 or I-131, the dosage. Your QMP is missing procedures to require that the written directive for I-125 and/or I-131 > 30 microcuries:

- is dated and signed by the authorized user

A footnote to 10 CFR 35.32(a)(1) provides that an oral revision to a written directive is acceptable if, because of the patient's condition, a delay in order to provide a written revision to an existing written directive would jeopardize the patient's health. Oral revisions must be documented immediately in the patient's record and a revised written directive must be signed and dated by an authorized user or physician under the supervision of an authorized user within 48 hours of the oral revision. Please include such a policy in your QMP.

9407290123 940601  
PDR ADOCK 03019932  
C PDR

IE09

JUN 01 1994

If, because of the emergent nature of the patient's condition, a delay in order to provide a written directive would jeopardize the patient's health, an oral directive will be acceptable provided that the information provided in the oral directive is documented immediately in the patient's record and a written directive is prepared within 24 hours of the oral directive. Please include such a policy in your QMP.

Revisions to written directives may be made for any diagnostic or therapeutic procedure provided that the revision is dated and signed by an authorized user prior to the administration of the radiopharmaceutical dosage. Your QMP must include a policy/procedure that requires that revisions to written directives will be made prior to administration.

The dosage should be confirmed by the person administering the radiopharmaceutical to verify agreement with the written directive; that is, the dosage should be measured in the dose calibrator and the results compared with the prescribed dosage in the written directive. Please provide such (or similar) procedures in your QMP.

A commitment to retain each written directive and a record of each administered radiopharmaceutical dosage for three years after the date of administration is required in 10 CFR 35.32(d). Describe the procedure for an authorized user or a qualified individual under the supervision of an authorized user (e.g., a nuclear medicine physician, physicist or technologist), after administering a radiopharmaceutical, to make, date, sign or initial a written record that documents the administered dosage in an auditable form.

Your QMP must include policies/procedures to institute corrective actions to be taken after an unintended deviation has been identified.

As required in 10 CFR 35.32(c), the licensee shall evaluate and respond, within 30 days after discovery of the recordable event, to each recordable event by: (a) assembling the relevant facts including the cause, (b) identifying what, if any, corrective action is required to prevent recurrence, and (c) retaining a record, in an auditable form, for three years, of the relevant facts and what corrective action was taken. Please include such a provision in your QMP.

Your QMP review procedure does not provide an evaluation of: (a) an adequate representative sample of patient administrations, (b) all recordable events, and (c) all misadministrations since the last

JUN 01 1994

review as required in 10 CFR 35.32(b)(1). The number of patient cases to be sampled should be based on the principles of statistical acceptance sampling and should represent each modality performed in the institution (e.g., radiopharmaceutical, teletherapy, brachytherapy, and gamma stereotactic radiosurgery). You may develop a sampling procedure of your own; use the chart provided in 10 CFR 32.110 (assuming an error rate of 2 percent); or a representative sample may be selected including (at a minimum): 20% if the number of cases performed is greater than 100, 20 cases if the number of cases is between 20 and 100, and all, if the number of cases is less than 20.) Provide a copy of your revised QMP to include this provision.

Your QMP should include a procedure to expand the number of cases reviewed when a misadministration or recordable event is uncovered during the periodic review of your QMP. Please include such a provision in your QMP.

Describe your procedures to evaluate the effectiveness of the QMP, and, if necessary, to make modifications to meet the objectives of the program as required by 10 CFR 35.32 (b)(2).

Please provide assurance that records of each QMP review and evaluation will be maintained for three years as required in 10 CFR 35.32 (b)(3).

To meet the requirements in 10 CFR 35.32, you may choose to utilize the procedures described in Regulatory Guide 8.33 (enclosed), or submit procedures that are equivalent. If you choose to use Regulatory Guide 8.33, be certain that the procedures you select are adjusted to meet the specific needs of your program as necessary. Additionally, you are reminded that training and/or instruction of supervised individuals in your QMP is required by 10 CFR 35.25.

Due to the apparent failure of your written QMP to meet the objectives in 10 CFR 35.32, you must immediately modify your written QMP to address the items listed above, and provide those modifications to your NRC regional office within 30 days of the date of this letter. NRC will review these matters during your next routine NRC inspection to determine whether violations of NRC requirements have occurred. Enforcement action may be taken at that time for failure to meet the requirements of 10 CFR 35.32.

Please be advised that this QMP will not be incorporated into your license by condition. This allows you the flexibility to make changes to your quality management program without obtaining prior NRC

JUN 01 1994

approval. When modifications are made to your program, You should submit any changes to your QMP to this Office within 30 days as required by 10 CFR 35.32(e).

Thank you for your cooperation in this matter. If you have any questions, please call Mr. John M. Pelchat at 404/331-5083.

Sincerely,

Original Signed By  
D. M. Collins  
Douglas M. Collins, Chief  
Nuclear Materials Safety and  
Safeguards Branch  
Division of Radiation Safety  
and Safeguards

Enclosure:  
Reg. Guide 8.33, "Quality  
Management Program"

bcc: Document Control Desk

|      |      |          |          |           |
|------|------|----------|----------|-----------|
| SEND | OFC  | RII:DRSS | RII:DRSS | RII:DRSS  |
| TO   | NAME | JPelchat | JPPotter | DMCollins |
| PDR? | DATE | 05/26/94 | 05/26/94 | 05/26/94  |
| Yes  | No   | COPY?    | Yes      | No        |
|      |      |          |          | Yes       |

OFFICIAL RECORD COPY

DOCUMENT NAME: g:\drss\qmp\002484.ra

13-2484.ra