

Appendix

NOTICE OF VIOLATION

Rockford School of Medicine

License No. 12-18786-01

As a result of the inspection conducted on July 21, 1982, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violations were identified:

1. License Condition No. 15 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced applications and letters.

The referenced application dated July 31, 1979, states in Item No. 11 that survey meters will be calibrated on a quarterly basis.

Contrary to this requirement, your survey meter has been calibrated on an annual basis instead of the required quarterly basis.

This is a Severity Level V violation (Supplement VI).

2. License Condition No. 15 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced applications and letters.

The referenced letter dated October 17, 1979, states in Item No. 1 that the exposure rate of packages containing radioactive materials will be measured at a distance of one meter, at the surface of the package, and that the empty container be monitored before discarding.

Contrary to this requirement, only a visual inspection of packages containing radioactive material has been made since December 1979.

This is a Severity Level V violation (Supplement VI).

3. License Condition No. 15 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced applications and letters.

The referenced letter dated October 17, 1979, states in Item No. 7 that all preparation, animal injection and waste storage areas will be surveyed daily with a low range thin-window G-M survey meter and that all other laboratory areas be surveyed on a weekly and monthly basis.

The weekly and monthly survey will consist of a measurement of radiation levels with a survey meter and a series of wipe tests to measure contamination levels.

Contrary to this requirement, only monthly wipe test of the laboratory areas have been performed since December 1979 and during 1981 there was a span of five months when these wipe tests were not performed.

This is a Severity Level V violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

7-27-82
Dated

Orig signed D. G. Wiedeman

D. G. Wiedeman, Chief
Materials Radiation Protection
Section 1