

DEC 15 1982

APPENDIX A

NOTICE OF VIOLATION

Radiology Professional Association
Jersey City, New Jersey 07307

Docket No.: 030-15018
License No.: 29-18407-01

As a result of the inspection conducted on November 22, 1982, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violation was identified:

Condition 13 of License No. 29-18407-01 requires that licensed materials be possessed and used in accordance with the statements, representations and procedures contained in application dated January 20, 1979.

Item 3 of the "Instructions for Personnel Opening and Monitoring Packages Containing Radioactive Material", on page 3 of item 13 of this application, requires that packages be monitored with a GM meter at 3 feet from the package, and at the surface, before opening. Item 11 of the application also requires that packages be wipe tested for contamination.

Contrary to the above, as of November 22, 1982, packages of radiopharmaceuticals were not routinely monitored with a GM meter or by wipe testing.

This is a Severity Level IV violation. (Supplement 10).

Pursuant to the provisions of 10 CFR 2.201, Radiology Professional Association is hereby required to submit to this office within thirty days of the date of this Notice, a written statement or explanation in reply, including; (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

OFFICIAL RECORD COPY

8212290365 821215
NMS LIC30
29-18407-01 PDR