

MAR 03 1983

MEMORANDUM FOR: R. C. Lewis, Director, Division of Project and Resident Programs, Region II

FROM: U. Potapovs, Chief, Vendor Program Branch, Division of Vendor and Technical Programs, Region IV

SUBJECTS: (1) AITS NO. F02700220, AMBIGUOUS OUTPUTS FROM BARTON PRESSURE TRANSMITTERS (VITS NO. 82-130)

(2) AITS NO. F02700259, POSTACCIDENT READINGS FROM CONTAINMENT SUMP LEVEL TRANSMITTERS (VITS NO. 82-238)

The aforementioned items were evaluated during a Vendor Program Branch (VPB) inspection at Barton Instruments. Findings are: regarding Item 1 - It appears that the current generation of pressure transmitters has reached the peak for accuracy; and regarding Item 2 - There were no indications of improper performance on the part of the supplier.

Details of the inspection are contained in the attached VPB Inspection Report No. 99900113/82-04, with followup on the deficient areas to be accomplished during the routine inspection effort. These AITS items have now been closed by the VPB. Please contact us if you have any questions.

"Original Signed by
U. POTAPOVS"

Uldis Potapovs, Chief
Vendor Program Branch

Attachment: As Stated

cc/att: R. L. Baer, IE
R. F. Heishman, IE
J. T. Collins, RIV
R. L. Bangart, RIV
J. E. Gagliardo, RIV

bcc: W. E. Foster
7010 VPB File

RIV
R&CPS
WEFoster/rc
2/4/83
FB

SC:R&CPS
IBarnes
2/4/83
FB

BC:VPB UP
UPotapovs
2/4/83

D:V&TP
RLBangart
2/7/83

8303070074 830303
PDR GA999 EMV*****
99900113 PDR

15:09