

APR 29 1983

Evanston and Glenbrook Hospitals
ATTN: Mr. Mark R. Neaman
Vice President
2650 Ridge Avenue
Evanston, Illinois 60201

License No. 12-00437-01

30-1373

Gentlemen:

Thank you for your letter dated April 13, 1983, informing us of the steps you have taken to correct the noncompliance which we brought to your attention in our letter dated March 18, 1983.

This also refers to the telephone conversations on April 26 and April 28, 1983, between S. Guru Prasad, Ph.D. of your staff and Mr. W. P. Reichhold of this office concerning your corrective actions.

It is our understanding that the following corrections to your letter dated April 13, 1983, were made:

1. Item 4a which references Appendix II was changed to Appendix III.
2. Item 10 which references Appendix IV was changed to Appendix II.

Also, it is our understanding that your corrective actions will include the following:

1. Technologists will be instructed to block and brace packages containing radioactive materials when packages are transported in their personal automobiles.
2. Daily surveys will be performed at Glenbrook and Evanston Hospitals, and the survey results will be recorded.
3. In July 1983, you will request that your license be changed to allow the disposal of radioactive urine into the sanitary sewer.

EHQ FILE COPY

8305120360 B30429
NMS LIC30
12-00437-01 PDR

OFFICE							
SURNAME							
DATE							

DMB:DE07

APR 29 1983

4. A reminder system has been established to ensure that the reference and calibration sources will be wipe tested every 6 months.
5. Survey meters will be calibrated at the recommended frequency. Currently, the recommended frequency is annually.


During the telephone conversation with S. Guru Prasad, Ph.D., it was stated that Item 9 will be corrected when your license is renewed in July 1983. Be advised that no activity is in accord with NRC regulatory requirements unless and until it is specifically authorized in an NRC license.

If your understanding differs from the above, please contact this office.

We will examine these matters during a future inspection.

Your cooperation with us is appreciated.

Sincerely,



J. A. Hind, Director
 Division of Radiological and
 Materials Safety Programs

cc: S. Guru Prasad, Ph.D.
 Chief, Radiological
 Physics

cc w/ltr dtd 4/13/83:
 DMB/Document Control Desk (RIDS)

OFFICE	RIII	RIII	RIII	RIII	RIII	RIII
SURNAME	Mullauer:rr	REichhold	Wiedeman	Miller	WWS Schultz	Hind
DATE	4/29/83	4/29/83	4/29/83	4/29/83	S-2-83	5/2/83