

APPENDIX A

NOTICE OF VIOLATION

Department of the Army
Harry Diamond Laboratories
Adelphi, Maryland 20783
License No. 19-17250-01

Based on the results of an NRC inspection conducted on May 25-26, 1977, it appears that one of your activities was not conducted in full compliance with NRC regulations as indicated below:

- A. 10 CFR 20.203(c)(2), requires that you control access to high radiation areas that have been established for periods of 30 days or more by one of the following methods: (1) Equip each entrance or access point with a control device which shall cause the level of radiation to be reduced below that at which an individual might receive a dose of 100 millirems in one hour upon entry into the area; or (2) Equip each entrance or access point with a control device which shall energize a conspicuous visible or audible alarm signal in such a manner that the individual entering the high radiation area and the licensee or supervisor of the activity are made aware of the entry; or (3) Maintain the area locked except during periods when access to the area is required, with a positive control over each individual entry.

Contrary to this requirement, on the day of the inspection, May 25, 1977, you failed to control access to all of the entrances to your high radiation area inside your calibration facility by the employment of any of the described methods.

This is an infraction.

8307220073 830307
PDR FOIA
SIMPSUNB2-608 PDR

OFFICE▶							
JURNAME▶							
DATE▶							