

AUG 10 1994

TAYLOR HOSPITAL
175 E. CHESTER PIKE
RIDLEY PARK, PA 19078

ATTN: PATRICIA SALUK M.D.

RE: Docket Number: 030-17757
License Number: 37-16507-02
Plan File Date: 19-MAY-93
Region Number: 1

Dear Dr. Saluk:

We have received your statement that, although you are licensed for I-125 and /or I-131 > 30 microcuries , you do not intend to use it. Please be aware that before using I-125 and /or I-131 > 30 microcuries for human use, you must provide an applicable quality management program that meets the requirements in 10 CFR 35.32 to the NRC regional office. Use of this material without the submission of an applicable quality management program could result in enforcement action.

No reply is required in response to this letter. Use of this modality will be reviewed at the next regular inspection of your facility.

Your QMP was reviewed by an NRC contractor following a standard review plan and related checklist provided by the NRC staff. This letter outlining the findings of that review was prepared by the NRC contractor utilizing standard paragraphs previously reviewed and approved by NRC headquarters and regional management.

Your QMP was reviewed by an NRC contractor following a standard review plan and related checklist provided by the NRC staff. This letter outlining the findings of that review was prepared by the contractor utilizing standard paragraphs previously reviewed and approved by NRC headquarters and regional management. If you have any questions about this review, you may call me at (610)337-5309. Thank you for your cooperation in this matter.

Sincerely,

Original Signed By:
James P. Dwyer

James P. Dwyer
Quality Management Program Coordinator
Region I

9411150227 940810
PDR ADOCK 03017757
C PDR

140057

RETURN ORIGINAL TO
REGION I

OFFICIAL RECORD COPY - D:\QMP4\003197_1 - 08/08/94

ML 10

IE-07