

APPENDIX A

NOTICE OF VIOLATION

Clinica Oncologica

License No. 52-11832-01

As a result of the inspection conducted on July 28, 1983, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violations were identified.

- A. 10 CFR 35.21(a) requires that full calibration measurements be performed annually on each teletherapy unit.

Contrary to the above, as of the date of inspection, full calibration measurements on both teletherapy units had not been performed since March 1982, an interval of more than one year.

This is a Severity Level IV Violation (Supplement VI).

- B. 10 CFR 35.23(a) requires that calibration, measurements on teletherapy units required by 10 CFR 35.21(a) be performed using a dosimetry system that has been calibrated by the National Bureau of Standards or by a Regional Calibration Laboratory accredited by the American Association of Physicists in Medicine. The dosimetry system shall have been calibrated within the previous two years and after any servicing that may have affected system calibration.

Contrary to the above, as of the date inspection, the dosimetry system used for the last full calibration had not been calibration as required since January 1980.

This is a Severity Level IV Violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are hereby required to submit to this office within thirty days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Date: AUG 17 1983