

FCML:LT
(12-17503-01)

OCT 11 1979

Stan A. Huber Consultants, Inc.
ATTN: Mr. Stan A. Huber
235 Essex Lane
New Lenox, IL 60451

Gentlemen:

Enclosed is Amendment No. 6 to License No. 12-17503-01.

Please note that Items 6.G., 7.G., and 8.G. only authorize use for up to 100 millicuries of cesium-137 in the J. L. Shepherd 28-6A calibrator. This action was taken because your letter dated September 12, 1979, included shielding calculations using a 100 millicurie cesium-137 source. In addition, your letter stated that the maximum activity you would be using at any one time would be 100 millicuries. If you wish to use a source of greater activity, you should submit calculations to show the radiation levels outside your facility using the higher activity source.

Sincerely,

James A. Jones
Material Licensing Branch
Division of Fuel Cycle and
Material Safety

Enclosure: Amendment No. 6 to
License No. 12-17503-01

XA Copy Has Been Sent to PDR

791730269

OFFICE	FCML <i>alt</i>	FCML <i>alt</i>			
BURNARD	LTremeper:bjb	JAJones			
DATE	9/18/79	9/11/79			

A/73