

JAN 25 1984

Docket No. 30-11353

License No. 06-16624-01

Johnson Memorial Hospital
ATTN: Mr. George Koller
Assistant Administrator
Stafford Springs, Connecticut 06076

Gentlemen:

Subject: Inspection No. 30-11353/83-01

This refers to your letter dated November 7, 1983, in response to our letter dated October 21, 1983.

Thank you for informing us of the corrective and preventive actions documented in your letter. With regard to item A, it is our understanding from a telephone conversation on November 16, 1983 between Mr. Stanley Scroggins of your staff and Mr. C. Stearns of this office, that a log has been set up to ensure that linearity testing of your dose calibrator will be done quarterly. With regard to item B, it is our understanding that the actual readings obtained during surveys of waste will be recorded and that the waste will not be disposed to normal trash unless these readings are not significantly greater than background. We will review these actions during the next inspection of your licensed program.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By:

John D. Kinneman

Thomas T. Martin, Director,
Division of Engineering and
Technical Programs

cc:
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
State of Connecticut
Stanley E. Scroggins, Director of Radiology

8402020004 840125
NMS LIC30
06-16624-01 PDR

OFFICIAL RECORD COPY

214STEARNS12/8/83 - 0001.0.0
01/23/84

Johnson Memorial Hospital

2

JAN 25 1984

bcc:
Region I Docket Room (w/concurrences)

for Kinneman
RI:DETP
Stearns/djh
12/8/83

Kinneman
RI:DETP
Kinneman
~~12/ /83~~
1/24/84

OFFICIAL RECORD COPY

2.4-STEARN12/8/83 - 0002.0.0
01/23/84