

REGISTRY OF RADIOACTIVE SEALED SOURCES AND DEVICES
SAFETY EVALUATION OF DEVICE

NO: NR-367-D-101-S

DATE: OCT 6 1982

PAGE 1 OF 1

DEVICE TYPE: Aerosol Analyzer

MODEL:

MANUFACTURER/DISTRIBUTOR:

IITRI Research Institute
Chicago, IL

MANUFACTURER/DISTRIBUTOR:

SEALED SOURCE MODEL DESIGNATION:

Amersham Model AM1Q-2367

ISOTOPE: Americium-241

MAXIMUM ACTIVITY: 125 microcuries

LEAK TEST FREQUENCY: 6 months

PRINCIPAL USE: (S) Foil Source

CUSTOM DEVICE: X YES NO

CUSTOM USER:

U.S. Department of the Interior
Bureau of Mines
Twin Cities Research Center
5629 Minnehaha Avenue, South
Minneapolis, MN 55417

8403020259 840209
PDR FOIA
HAMMITT84-74 PDR

REFERENCES: NRC registry document No. 81-88 dated October 17, 1981, attached.

ISSUING AGENCY: U.S. Nuclear Regulatory Commission