

ML:IB:JBB

SEP 14 1966

F & M Scientific
Division of Hewlett-Packard
Route 41 and Starr Road
Avondale, Pennsylvania 19311

Attention: Mr. Jim Peters

Gentlemen:

We have evaluated the Model 2-6195 detector cell incorporating two millicuries of Nickel 63 and the Model P-6761 Leak Test Kit for use with the detector cell.

Both items are acceptable for licensing purposes and we are prepared to accept license applications from your customers.

Very truly yours,

William O. Miller
Isotopes Branch
Division of Materials
Licensing

bcc: Isotopes Branch Standard Distribution
✓ Sealed Source Files

9504200168 950410
PDR RC *
SSD PDR

ML:IB
JMBell:bjs

ML:IB
WOMiller

9-12-66

9- -66

9504200168

Talked to Peters June 13, 1966, re American Cyanamid Company application 78010 dated June 7, 1966, for Ni-63 in an F & M cell.

Peters says unit is being advertised, but not yet in production. Says he will contact Cyanamid and point out that F & M must furnish information to AEC before license can be issued.

Peters says he is thinking about visiting us to discuss Ni-63 cell. Told him I didn't think a visit was necessary. He says maybe he will send package in prior to further consideration of a visit. I briefly outlined the information we would need for a Ni-63 cell.

Jack M. Bell
6-13-66