

APPENDIX A
NOTICE OF VIOLATION

Philadelphia Electric Company
Peach Bottom Atomic Power Station, Units 2 and 3

Docket Nos. 50-277
50-278
License Nos. DPR-44
DPR-56

As a result of the inspection conducted on May 8, 1984, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), published in the Federal Register on March 8, 1984 (49 FR 8583), the following violation was identified:

Appendix "A" to License Nos. DPR-44 and DPR-56, Technical Specifications Section 6.11, requires that procedures for personnel radiation protection shall be approved, maintained and adhered to for all operations.

Contrary to the above, written procedures were not approved and maintained for operation of the whole-body counting systems. The moving bed system and associated software was placed in service on February 10, 1984 and the stand-up system and associated software was placed in service on May 7, 1984.

This is a Severity Level V violation (Supplement I).

Pursuant to the provisions of 10 CFR 2.201, Philadelphia Electric Company is hereby required to submit to this office within thirty days of the date of the letter which transmitted this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

OFFICIAL RECORD COPY

DL50-277/84-16 - 0003.0.0
06/04/84

8407130344 840615
PDR ADOCK 05000277
Q PDR