


DUKE POWER COMPANY

OCONEE NUCLEAR STATION

EMERGENCY PLAN IMPLEMENTING PROCEDURES



APPROVED:

M. S. Tuckman

M. S. Tuckman, Station Manager

3/23/84

Date Approved

March 21, 1984

Effective Date

Revision 84-1

TABLE OF CONTENTS

	Emergency Telephone Numbers - (Rev. 84-2, 3/1/84)
AP/O/B/1000/01	Event Index - (10/29/82)
AP/O/B/1000/02	Unusual Event - (08/10/83)
AP/O/B/1000/03	Alert - (08/10/83)
AP/O/B/1000/04	Site Area Emergency - (08/10/83)
AP/O/B/1000/05	General Emergency - (08/10/83)
AP/O/B/1000/06	Procedure for Initiating Protective Action Guides for the General Public in the Emergency Planning Zone - (08/10/83)
AP/O/B/1000/07	Procedure for Offsite Dose Calculations by Control Room Personnel or Emergency Coordinator during a Radiological Accident - (06/09/83)
AP/O/B/1000/08	Procedure for Response Actions for Accidents/Emergencies - (09/03/82)
RP/O/B/1000/10	Procedure for Emergency Evacuation of Station Personnel - (11/16/83)
Station Directive 2.9.1	Station Assembly and Evacuation Procedure - (09/01/83)
Station Directive 2.9.2	Emergency Response Organization - (06/23/82)
PT/O/B/2000/04	Procedure for Establishment and Inspection of the Technical Support Center - (11/16/83)
PT/O/B/2000/05	Oconee Nuclear Data Transmission from TSC (09/15/82)
CP/1/A/2002/04A	Post Accident Liquid Sampling of the Reactor Coolant System (12/01/81)
CP/2/A/2002/04A	Post Accident Liquid Sampling of the Reactor Coolant System (12/01/81)
CP/3/A/2002/04A	Post Accident Liquid Sampling of the Reactor Coolant System (12/01/81)
CP/1/A/2002/04B	Post Accident Liquid Sampling of the Low Pressure Injection System (12/01/81)

Revision 84-1
March, 1984

Table of Contents (Implementing Procedures - Continued)

HP/O/B/1009/12	Distribution of Potassium Iodide Tablets in the Event of a Radioiodine Release - (04/13/83)
HP/O/B/1009/13	Procedure for Implementation and Verification for the Availability of a Back-Up Source of Meteorological Data - (2/12/83)
HP/O/B/1009/14	Project Offsite Dose from Releases other than Through a Vent - (12/20/82)
HP/O/E/1009/15	Procedure for Sampling and Quantifying High Level Gaseous Radioiodine and Particulate Radioactivity - (1-25-83)
HP/O/B/1009/16	Procedure for Emergency Decontamination of Personnel and Vehicles on-site and from Off-site Remote Assembly Area - (9/16/82)
IP/O/B/1601/03	Environmental Equipment Checks - (09/28/83)
CP/3/A/2002/04C	Operation Procedure for Post-Accident Liquid Sampling (PALS) System (4/21/83)
HP/1/A/1009/17	Operating Procedure for Post-Accident Containment Air Sampling System (12/21/83)
HP/2/A/1009/17	Operating Procedure for Post-Accident Containment Air Sampling System (12/21/83)
HP/3/A/1009/17	Operating Procedure for Post-Accident Containment Air Sampling System (12/21/83)

Confidential
Not For Publication

EMERGENCY TELEPHONE NUMBERS

This enclosure provides a listing of telephone numbers for various personnel and agencies that may have a part in dealing with an emergency situation or providing other assistance as needed at Oconee Nuclear Station.

EMERGENCY TELEPHONE NUMBERS

This directory provides a listing of telephone numbers for various personnel and agencies that may have a part in dealing with an emergency situation or providing other assistance as needed at Oconee Nuclear Station.

TABLE OF CONTENTS

NOTIFICATION OF AGENCIES, CRISIS MANAGEMENT TEAM, KEY COMPANIES	1 - 5
EMERGENCY FACILITY LOCATIONS.	6
OCONEE TELEPHONE ACCESS CODES	7
TECHNICAL SUPPORT CENTER.	8 - 10
OPERATIONAL SUPPORT CENTER.	9
HEALTH PHYSICS CENTER	8
COMMUNICATIONS COORDINATION	10
EMERGENCY COUNT-ROOM.	10
CRISIS MANAGEMENT CENTER	11 - 13
GENERAL OFFICE SUPPORT CENTER	12
BACKUP CRISIS MANAGEMENT CENTER - (Liberty)	13
CRISIS NEWS CENTER.	14
NUCLEAR REGULATORY COMMISSION	5
NRC HEALTH PHYSICS NETWORK TELEPHONES	15
DUKE EMERGENCY RADIO	16
EMERGENCY NUMBERS - Pickens County	17
EMERGENCY NUMBERS - Oconee County	18

DUKE POWER COMPANY
OCONEE NUCLEAR STATION

NUMBER CODE FOR IDENTIFYING PERSONNEL/ACTIVITIES TO BE NOTIFIED

CODE

1. NUCLEAR REGULATORY COMMISSION by Red Phone within one hour.
2. UNIT COORDINATOR/OPERATIONS DUTY ENGINEER who will notify:
 - A. Superintendent of Operations
 - B. Station Manager/Emergency Coordinator (or alternate as listed in number 11.)
 - C. Nuclear Production Duty Engineer who will notify:
 1. Corporate Communications
 2. Crisis Management Organization
3. STATION MANAGER
 - M. S. Tuckman, Office
 - Home
4. BABCOCK AND WILCOX RESIDENT ENGINEER
 - Bill Street, Office
 - Home

(If Bill Street cannot be reached, call)

 - L. H. Williams, Office
 - Home
5. STATION HEALTH PHYSICIST/DUTY HEALTH PHYSICIST
 - C. T. Yongue, Office
 - Home

6. SOUTH CAROLINA DEPARTMENT OF HEALTH AND ENVIRONMENTAL CONTROL,
(Warning Point State of South Carolina)

Bureau of Radiological Health (0800-1700)
Answering Service after hours, weekends, holidays.

*State Emergency Operations Center, Columbia, S.C.

*Forward Emergency Operations Center, Clemson, S. C. Duke Ringdown

Alternate Number

*NOTE: These numbers are to be used once the State
has established their Emergency Operations.

7. COUNTY EMERGENCY PREPAREDNESS AGENCIES

Oconee County Emergency Preparedness Duke
Alternate Number - 24 hour, page
Alternate Number - 24 hour, page

Pickens County Emergency Preparedness. Duke
Alternate Number - (0830-1700)
Alternate Number - 24 hour, page

8. COUNTY SHERIFF'S DEPARTMENTS

Oconee County (24 hours) Duke
Alternate Number

Pickens County (24 hours) Duke
Alternate Number
Alternate Number
Alternate Number

9. MEDICAL ASSISTANCE

Oconee Memorial Hospital Ambulance Service
Oconee Memorial Hospital Switchboard/Supervisor or Nursing

Additional Medical assistance may be provided through the
following institutions:

Pickens County Ambulance Service
Cannon Memorial Hospital/Supervisor of Nursing
Easley Baptist Hospital/Supervisor of Nursing

10. FIRE ASSISTANCE

Oconee County Rural Fire Protection Association
Woods or Forest Fire (Oconee County, Oakway Tower)
Woods or Forest Fire (Pickens County, Woodall Mt. Tower)

17. EMERGENCY COORDINATOR AND ALTERNATES (TSC Activation)

(If the first person cannot be reached, go to the next person down the list until one person is contacted)

Assistant Station Manager

Superintendent of Technical Services

T. S. Barr, Office
Home

Superintendent of Maintenance

J. M. Davis, Office
Home

Superintendent of Operations

J. N. Pope, Office
Home

Operations Duty Engineer



12. WATER DEPARTMENTS

Should releases of radioactive effluent into Lake Keowee or Lake Hartwell potentially effect municipal water intakes or exceed technical specifications. Contact the appropriate authorities as indicated below:

Lake Keowee

Seneca, H. J. Balding, Office
Home

Lake Hartwell

City of Clemson

Mayor of Clemson, Office
Home

(If the mayor cannot be reached, call one of the following)

Clemson Administrator's Office
Home

Clemson Filter Plant (0700-1700)

Clemson University

President's Office
Home

Security - Police (24 hours)
(If the President cannot be reached, call)
Clemson University Physical Plant (0800-1630)

Anderson Water Works (24 Hr. Number)

AGENCIES THAT MAY RESPOND TO AN EMERGENCY AT THE OCONEE NUCLEAR STATION

LAW ENFORCEMENT (24-hour numbers)

S. C. Highway Patrol (Greenville, S.C.)

S. C. Enforcement Division (Columbia, S.C.)

FBI (Columbia, S.C.)

BOMB DISPOSAL

Explosives Ordnance Disposal Control (24-hour)
(Fort Jackson, Columbia, S.C.)

RADIATION AND CONTAMINATION

- REACTS, Department of Energy (Oak Ridge, Tennessee) . . .
(24 hr. number - after 1700 ask for Beeper number) . . .
- DOE Emergency Radiological Monitoring Team (Aiken, S.C.) .
- N. C. Division of Emergency Management
(Warning Point - State of North Carolina)
- Georgia Department of Natural Resources
Environmental Radiation Program
(Warning Point - State of Georgia)

NUCLEAR REGULATORY COMMISSION

- NRC Operations Center (via Bethesda Central Office) . . .
- NRC Operations Center (via Silver Spring Central Office) .
- Health Physics Network to NRC Operations Center
- Health Physics Network to NRC, Region II
- NRC Operator (Via Bethesda Central Office)
- US NRC, Region II (Operations Center).
- US NRC, Oconee Resident Inspectors

Jack Bryant	Home
Dolan Falconer	Home

BUS TRANSPORTATION

- Anderson Retail Office (24 hour number)
(Contact John Holland, Jerry Whitefield)

NATIONAL WEATHER SERVICE - METEOROLOGICAL BACK-UP SOURCE

- Greenville-Spartanburg Weather Service (24 hour) .

FEDERAL AERONAUTICS AGENCY

Private Aircraft

- Flight Standards District Office (0800-1700). . . .
- Flight Service Station (After hours, weekends, holidays) .

Military Aircraft

- Air Station Mgr. (Shaw AF Base)

OCONEE NUCLEAR STATION
CRISIS COMMUNICATIONS DIRECTORY

The crisis directory is intended for use should the Oconee Emergency Plan require implementation. Both station and corporate level telephone numbers are provided. The station's emergency organization will operate from the Technical Support Center near the Units 1 and 2 Control Room. The corporate emergency organization will operate from the Crisis Management Center located in the Visitors Center and Oconee Training Center.

EMERGENCY FACILITY LOCATIONS

Technical Support Center - Control Rooms 1 and 2

Operational Support Center - Control Room 3

Crisis Management Center - Oconee Training Center

Alternate Location: Liberty Retail Office

Crisis News Center - Keowee-Toxaway Visitors Center

Alternate Location: Liberty Retail Office

OCONEE NUCLEAR STATION
TELEPHONE DIRECTORY

Seneca Lines



Easley Lines



Anderson Line



Six Mile Line



Dial Code
(Micro-Wave)



(Charlotte General Office)

(Catawba)

(McGuire)

Attendant (To access
Bell Line)

Seneca

Easley

Anderson

Six Mile

OCONEE NUCLEAR STATION
CRISIS PHONE DIRECTORY
TECHNICAL SUPPORT CENTER

<u>POSITION/NAME</u>	<u>Telephone Number</u>	
	<u>Outside Line</u>	<u>Station Number</u>
Emergency Coordinator		
Supt. of Operations		
Supt. of Technical Services		
Supt. of Maintenance		
Supt. of Administration		
NRC Resident Engineer FTS		
B&W Resident Engineer		
Station Health Physicist		
<u>HEALTH PHYSICS CENTER</u>		
Field Monitoring Coordinator		
Data Report Coordinator (Off-Site Dose Projection)		
Dose Coordination to CMC		
FTS Line to NRC		

	<u>Telephone Number</u>	
	<u>Outside</u>	<u>Station</u>
	<u>Line</u>	<u>Number</u>

TECHNICAL SERVICES GROUP (Located in Computer Room CR 1&2)

- Performance
- Compliance
- Chemistry

OPERATIONAL SUPPORT CENTER

(Support group consists of Health Physics, Chemistry, Maintenance, Safety Operations group)

- Operational Support Center Coordinator
- Mechanical Maintenance Engineer
 - Mechanical Maintenance Supervisor
- I & E Engineer
 - I & E Supervisor
- Health Physics Support
 - Dose Control
 - S & C Coordinator
 - Support Function Coordinator.
- Chemistry Support
- Medical Support
- Clerical
- Operations Group
 - Unit 3 Operations Offices

- Nuclear Equipment Operators (Unit 1 & 2 Emergencies)
- Nuclear Equipment Operators (Unit 3 Emergencies)



CONTROL ROOM


Unit 1
Unit 2
Unit 3
Shift Supervisor (Unit 1 & 2)
 Unit 3

COMMUNICATIONS COORDINATION

Data Transmission Coordinator
Data Release (Vax Computer Program)

Telecopier (Technical Support Center)
Offsite Communicator
TSC Clerical Support
Emergency Response

EMERGENCY COUNT ROOM (Located in Visitor's Center)



OCONEE NUCLEAR STATION
CRISIS PHONE DIRECTORY
CRISIS MANAGEMENT CENTER

POSITION/NAME

PRIVATE
LINE

ONS
SWITCHBOARD

RECOVERY MANAGER

State of S.C. (FEOC Line)
(Duke Line)

SCHEDULING/PLANNING

RADIOLOGICAL SUPPORT

Bureau of Radiological Health (Duke Line)
(FEOC Line)

OFFSITE RADIOLOGICAL COORDINATOR

TECHNICAL SUPPORT

DESIGN AND CONSTRUCTION SUPPORT

ADMINISTRATION AND LOGISTICS

DATA COORDINATION

TELECOPIER

ADVISORY SUPPORT

NUCLEAR REGULATORY COMMISSION


BABCOCK & WILCOX (NSSS SUPPLIER)

CORPORATE HEADQUARTERS

(Contact with the Governor)

A. C. Thies

W. H. Owen



- 12 -
OCONEE NUCLEAR STATION
CRISIS PHONE DIRECTORY
GENERAL OFFICE SUPPORT CENTER

WACHOVIA CENTER

RECOVERY MANAGER (Room 1010) (Speaker Phone)
(Dedicated line to State Director)

NRC

SCHEDULING/PLANNING (Room 1010)

RADIOLOGICAL SUPPORT (Room 2390)

OFFSITE RADIOLOGICAL COORDINATOR (Room 1222)

NRC FTS LINE

TECHNICAL SUPPORT (Room 1704)

ADMINISTRATION AND LOGISTICS (Room 0925)

NUCLEAR REGULATORY COMMISSION (Room 1488)

ELECTRIC CENTER

DESIGN AND CONSTRUCTION SUPPORT (Room 32, 3rd Floor)

POWER BUILDING


CRISIS NEWS GROUP - DUKE (Rooms 5010, 5012, 5014)

S.C. PUBLIC INFORMATION OFFICERS (Rooms 5020, 5022)

NRC NEWS STAFF (Room 5024)

*Dedicated line to State Center

OCONEE NUCLEAR STATION
CRISIS PHONE DIRECTORY
BACKUP CRISIS MANAGEMENT CENTER
LIBERTY RETAIL OFFICE, LIBERTY, S.C.

AREA CODE - 
Telephone Number

RECOVERY MANAGER

SCHEDULING/PLANNING

PUBLIC INFORMATION OFFICERS*

State of South Carolina
Oconee County
Pickens County

DESIGN AND CONSTRUCTION

TECHNICAL SUPPORT

OFFSITE RADIOLOGICAL COORDINATOR

ADMINISTRATION AND LOGISTICS

HEALTH PHYSICS/RADWASTE

GOVERNMENT AGENCIES*

NRC
State of South Carolina
Oconee County
Pickens County

*NOTE: Call any one of the numbers listed to reach the desired representative.

OCONEE NUCLEAR STATION
CRISIS PHONE DIRECTORY
CRISIS NEWS CENTER
KEOWEE-TOXAWAY VISITORS' CENTER

<u>Position/Name</u>	<u>Private Line</u>	<u>Telephone Number</u> <u>ONS Switchboard</u>
<u>CRISIS NEWS DIRECTOR</u> Mary Cartwright	[REDACTED]	[REDACTED]
<u>COMMERCIAL NEWS MEDIA</u> (Active Numbers) For drill purposes only	[REDACTED]	
<u>COMMERCIAL NEWS MEDIA</u> (Inactive Numbers) Activated only during an actual emergency	[REDACTED]	
<u>NRC/STATE/COUNTY PUBLIC INFORMATION OFFICERS (PIO'S)</u>	[REDACTED]	
NRC Oconee County Pickens County	[REDACTED]	
State of S.C. (FECO Line)	[REDACTED]	
(Duke Line)	[REDACTED]	

*Note: NRC, Oconee County or Pickens County may be reached on any one of these phones.


NRC HEALTH PHYSICS NETWORK TELEPHONES

The NRC's Health Physics Network (HPN or Black Phone) connects all Nuclear Power Plants and Fuel Facilities to NRC Regional Offices and to NRC Headquarters Operations Center. The phone is intended to support Health Physics Operations in an emergency but can be used for daily voice traffic and facsimile transmittal.

The Station has jacks for the HPN phones in the Performance Office (Control Room 1 & 2) and in the Oconee Training Center.

The phone is used normally with the exception; NO DIAL TONE OR RINGING IS HEARD. In addition, ringing only lasts 30 seconds, so after 30 seconds if the party has not answered, you must hang up and redial.

For convenience, the codes most often used are listed below:

- | <u>HPN Phone</u> | <u>Code</u> |
|---|--|
| 1. NRC region 2 (Atlanta) office |  |
| 2. NRC headquarters (24 hours) | |
| 3. B&W Research Center | |
| 4. Oconee NRC Resident Inspector | |
| 5. Oconee Nuclear Station | |
| 6. <u>All</u> NRC region 2 Resident Inspectors | |
| 7. <u>All</u> region 2 Operating Nuclear Plants | |

In addition, the calling party may "conference" any phones during conversation by simply dialing the appropriate code(s). Any number of stations may be added in this manner.

OCONEE NUCLEAR STATION EMERGENCY RADIO

The call letters identify the Emergency Radio frequency. The following is a listing of radio locations, unit call letters, and identifiers. Use identifiers to begin a transmission and the call letters to close out the radio transmission. (For example: Oconee Nuclear Station Control Room to Pickens County Law Enforcement Center. Close out with [redacted] off.)

ONS Base Station Remotes

<u>Location</u>	<u>Unit Call Letters</u>	<u>Identifier</u>
1. Unit 1&2 Control Room	[redacted]	Oconee Control Room
2. Crisis Management Center	[redacted]	Oconee CMC
3. Technical Support Center	[redacted]	Oconee TSC

Coded Squelch Radios

<u>Location</u>	<u>Encode</u>	<u>Unit Call Letters</u>	<u>Identifier</u>
4. Pickens LEC	[redacted]	[redacted]	[redacted]
5. Pickens LOC	[redacted]	[redacted]	[redacted]
6. Pickens EPD	[redacted]	[redacted]	[redacted]
5. Oconee LEC	[redacted]	[redacted]	[redacted]
6. State FEOC - (Clemson)	[redacted]	[redacted]	[redacted]

ALL ABOVE RADIOS MAY BE ACTIVATED BY ENCODING NO

Field Monitoring Teams

<u>Location</u>	<u>Unit Call Letters</u>	<u>Identifier</u>
8. Field Monitor Coordinator	[redacted]	[redacted]
9. Field Monitor Team	[redacted]	[redacted]
10. Field Monitor Team	[redacted]	[redacted]
11. Field Monitor Team	[redacted]	[redacted]
12. Field Monitor Team	[redacted]	[redacted]
13. Field Monitor Team	[redacted]	[redacted]
14. Field Monitor Team	[redacted]	[redacted]

TO COMMUNICATE BETWEEN BASE STATION REMOTES (1, 2, 3), THE INTERCOM MUST BE USED! The following procedure must be used:

1. Push INTERCOM button and hold
2. Push MIKE button and hold
3. Send message (example, CMC to TSC)
4. Release both buttons to receive a response.

EMERGENCY OPERATION CENTER

Pickens County

Primary Number

EXECUTIVE GROUP*

Emergency Preparedness
County Administrator
County Council
Legal Officer

OPERATIONS GROUP*

Law Enforcement
Rescue Squad
EMS

Fire Service
Medical Service
Health Service
Dept. of Public Works

ASSESSMENT*

Transportation
Emergency Welfare Service
Shelter Service
Red Cross

Public Information
RADEF

Mental Health
Damage Assessment
Supply and Procurement

ALTERNATE NUMBER (to any group)

PUBLIC INFORMATION OFFICER

CRISIS NEWS CENTER-ONS*

State of South Carolina
Oconee County
Pickens County
NRC

CRISIS NEWS CENTER LIBERTY RETAIL OFFICE*

State of South Carolina
Oconee County
Pickens County
NRC

*Call any one of the listed numbers to reach group desired.

EMERGENCY OPERATION CENTER

Oconee County

Primary Number (24-hour)

OPERATIONS*

Fire Protection

Police

Public Roads

Emergency Medical Services

Rescue Squads

ASSESSMENT*

Emergency Welfare Services

Radiological Defense

Damage Assessment

EXECUTIVE GROUP*

Supervisor/Chairman County Council

EOC Director

Financial Officer

FNF Representative

PUBLIC INFORMATION OFFICER

CRISIS NEWS CENTER-ONS

State of South Carolina
Oconee County
Pickens County
NRC

CRISIS NEWS CENTER LIBERTY RETAIL OFFICE

State of South Carolina
Oconee County
Pickens County
NRC

*Call any one of the listed numbers to reach group desired.



INFORMATION ONLY

CONTROL COPY

Form SPD-1002-1

DUKE POWER COMPANY
PROCEDURE PREPARATION
PROCESS RECORD

(1) ID No: HP/1/A/1009/17
Change(s) 3 to
NA Incorporated

(2) STATION: Oconee

(3) PROCEDURE TITLE: Operating Procedure for Post-Accident Containment
Air Sampling System

(4) PREPARED BY: Sarah Cox DATE: 12-20-83

(5) REVIEWED BY: [Signature] DATE: 12-21-83

Cross-Disciplinary Review By: _____ N/R: g

(6) TEMPORARY APPROVAL (IF NECESSARY):

By: _____ (SRO) Date: _____

By: _____ Date: _____

(7) APPROVED BY: Jerry D. Bunn Date: 12/21/83

(8) MISCELLANEOUS:

Reviewed/Approved By: _____ Date: _____

Reviewed/Approved By: _____ Date: _____

INFORMATION ONLY

DUKE POWER COMPANY

OCONEE NUCLEAR STATION

OPERATING PROCEDURE FOR POST-ACCIDENT CONTAINMENT

AIR SAMPLING SYSTEM

1.0 Purpose

This procedure describes the operation of the Post-Accident Containment Air Sampling System which is used to obtain a prompt containment air sample under accident conditions while keeping radiation exposure ALARA. This procedure is also used to perform the semi-annual functional test of the system.

2.0 References

- 2.1 Duke Power Company Nuclear Station Post-Accident Containment Air Sampling System Manual
- 2.2 HP/O/B/1006/07, Procedure for Preparation of Gas Calibration Sources
- 2.3 CP/1/A/2002/0-C, Operating Procedure for the Post Accident Liquid Sampling (PALS) System
- 2.4 HP/O/B/1009/15, Procedure for Sampling and Quantifying High Level Gaseous, Radiiodine and Particulate Radioactivity
- 2.5 EP/O/A/1800/04, Loss of Coolant
- 2.6 Station Directive 4.2.5, Independent Verification Requirements
- 2.7 Station Directive 3.1.37, Interim Control of Independent Verification Requirements

3.0 Limits and Precautions

- 3.1 The sampling cycle will require two (2) qualified technicians approximately one (1) hour per sample, of which about ten (10) minutes will be spent in the sample panel area. One qualified technician will operate the control panel while the other will perform transit duties to and from the panel.
- 3.2 Personnel communications can be achieved by phone.
Unit 1 & 2 - Ext. [REDACTED] (by column [REDACTED])
Unit 3 - Ext. [REDACTED] (by door to [REDACTED])
- 3.3 The following items will never be used on the panel.
 - a. Trap Area Evacuation
 - b. Fast Sample Dilution

- 3.4 The Recirc Pump must never be used at any pressure other than 0 inches of Mercury.
- 3.5 Moving the Selector Switch from one mode to another stops all current system operations. Depressing the Activate button starts operation of the newly selected mode.
- 3.6 The radiation monitor on the control panel will provide levels of radiation at the sample panel. If the radiation monitor is not working properly, then a portable survey instrument will be used to determine radiation levels.
- 3.7 If problems with the pressure and/or temperature gauge are evident, such as going off scale or erratic response, the Selector Switch must be turned to the OFF mode and sampling discontinued until the problem is corrected.
- 3.8 If the sampling system cannot be operated, then HP/O/B/1009/15 (Ref. 2.4) will be used as an alternate method for obtaining a containment air sample.
- 3.9 Enclosure 5.5 will be used to check off the steps as the procedure is completed.
- 3.10 Operations must complete Enclosure 5.3 or 5.4 to bypass the Hydrogen Analyzer to bring containment air to the sampling system and to return the Hydrogen Analyzer to service after sampling is complete by Enclosure 5.6 or 5.7. These enclosures shall be independently verified to ensure that containment integrity is maintained. The Operations Unit Supervisor shall designate one "doer" and one "verifier" as required by Reference 2.6 and 2.7.
- 3.11 The front side of the sample panel is the side which contains the door. The left and right side of the sample panel will be determined by using this fact.
- 3.12 If radiation levels exceed 16 R/hr and cannot be reduced by purging the system, secure operation of the panel, move to a low background area, and contact the Station Health Physicist or his designee for further instructions.
- 3.13 Before sampling operations begin, the decision must be made based on radiological conditions in the reactor building and the sampling area whether to use a 100 ml gas bomb or a calibrated syringe for the gas sample. During emergency conditions, this decision will be made by the Station Health Physicist or his designee.
- 3.14 Enclosure 5.8, Valve Checklist for Sample Panel, may be used to provide assistance in determining flow inside the sample panel. It is not intended to provide a verification for valve operation.
- 3.15 During accident conditions, the keys needed for sampling will be located in the Shiftman's key cabinet.

- 3.16 The sampling system must not be used if reactor building pressure is greater than 40 psig.

4.0 Procedure

- 4.1 Locate the Shift Supervisor for Operations and request that Operations complete Enclosure 5.3 or 5.4 to bypass the Hydrogen Analyzer so the Post Accident Containment Gas Sampling System may be operated.

- 4.2 Obtain equipment necessary to perform sampling, including the thio-sulfate solution. Also obtain keys to the control panel and the sixth floor Ventilation Equipment Room.

NOTE: Necessary equipment for sampling is listed on Enclosure 5.1.

- 4.3 Open the valve on the nitrogen bottle next to the sampling panel to 40 psig.

- 4.4 Open the

- (a) DI Water Inlet
- (b) Instrument Air Inlet
- (c) N₂ Inlet

located on the left side of the sample panel.

NOTE: Open inlets by rotating the back switches counterclockwise one-quarter turn to the upward position.

- 4.4.1 Ensure the test tees on the sample inlet and outlet lines are closed.

- 4.4.2 Ensure inlet valve on gas sampler is open (black switch parallel with lines).

- 4.4.3 Ensure DI water supply line is open to the panel.

- 4.5 Position the thiosulfate funnel directly over the fill port located on top of the sample panel. Attach the hose on the funnel to the fill port and pour the 500 ml of thiosulfate solution into the funnel.

- 4.6 Set the switches listed below as follows:

- (a) Sample Volume Select - set on SMALL
- (b) Dilution Volume Select - set on LARGE
- (c) Selector Switch - set on OFF
- (d) System Purge - set on NORMAL
- (e) Refill Switch - set on OFF (down)
- (f) TC Switch - set on POSITION 1 (thermocouple measures sample line temperature)

- (g) Sample Line Select Switch - turn to Unit and Hydrogen Analyzer (Train A or B) being used for this operation of the sampling system
- 4.7 Turn the Key Lock Switch to POWER ON and ensure the power on light has come on.
- 4.8 Turn the Radiation Monitor toggle switch ON (up).
- 4.8.1 Turn the selector on the Radiation Monitor to BATT and ensure the needle is in the "red test region." Turn the selector to the MR/HR or R/HR scale.
- NOTE: If the Radiation Monitor is not functioning properly, note that it is not working on Enclosure 3.5, Step 9 and use a portable survey instrument to determine radiation levels during sampling.
- 4.9 Purge the Sample Panel.
- 4.9.1 Turn Select Switch to SYSTEM PURGE
- 4.9.2 Move Normal - Sample Purge to SAMPLE PURGE
- 4.9.3 Depress ACTIVATE button.
- 4.9.4 Depress EVAC button (Evac light on) and watch pressure gauge slowly drop to ~ - 19" of Hg. Depress STOP.
- 4.9.5 Press down and release the GAS PURGE toggle switch and watch the pressure gauge swiftly rise to + 10" of Hg. Depress STOP button.
- 4.9.6 Depress the EVAC button and watch the pressure gauge drop to 0" of Hg. Depress STOP button.
- 4.9.7 Depress the PUMP button and wait for 30 seconds. Depress STOP button.
- 4.9.8 Repeat Step 4.9.4 through 4.9.7 twice to purge the sample panel two more times.
- 4.9.9 Move Normal - Sample Purge to NORMAL.
- 4.9.10 Turn Selector Switch to SOLUTION CHANGE OUT.
- 4.10 Preparation for Sampling
- 4.10.1 Set the 500 ml sample bottle in a clear poly bag. Place the portable shielded container on the floor under the Thio-sulfate sampler (left side of panel), and place the sample bottle in the shielded container.

- 4.10.2 Detach the left side of the flexible tubing on the thiosulfate sampler located on the left side of the sample panel near the floor.
- 4.10.3 Insert the free end of the tubing into the 500 ml sample bottle.
- 4.10.4. Complete Steps a) and b) below if a 100 ml gas bomb will be used for the gas sample. If the gas sample will be drawn by syringe, go to Step 4.10.5.
- a) Detach the side of the flexible tubing on the gas sampler between the inlet valve and the hard piping.
 - b) Attach a 100 ml gas bomb between the free end of the flexible tubing and the hard piping on the gas sampler. Ensure valves on gas bomb are open.
- 4.10.5 Record sample line temperature reading for sample volume calculations on Enclosure 5.2.
- 4.11 Flush Thiosulfate Sampler and fill with Thiosulfate.
- 4.11.1 Depress ACTIVATE button.
 - 4.11.2 Depress FLUSH button and hold for 30 seconds.
 - 4.11.3 Depress PURGE button and hold for 30 seconds.
 - 4.11.4 Depress EMPTY button and hold for 45 seconds.
 - 4.11.5 Open the TS (thiosulfate) valve located inside the sample panel directly below the fillport. (Open valve in same manner as valves in Step 4.4).
 - 4.11.6 Move Refill toggle switch to ON (up) and wait 2 minutes. Move Refill to OFF (down).
 - 4.11.7 Turn Selector Switch to DILUTION VOLUME EVACUATION.
- 4.12 Evacuate the Dilution Volume.
- 4.12.1 Depress ACTIVATE button and watch pressure gauge drop to ~ - 19" of Hg. Turn Selector Switch to SAMPLE RECIRC.
- 4.13 Recirc Containment Air and Trap a Sample.
- 4.13.1 Depress ACTIVATE button and wait 10 minutes.
 - 4.13.2 Return to sample panel and note pressure gauge reading on sample inlet line. Record pressure on Enclosure 5.2.
 - 4.13.3 Depress SAMPLE button and wait 1 minute.

- 4.13.4 Depress TRAP button and wait 10 seconds.
- 4.13.5 Turn Selector Switch to SAMPLE DILUTION.
- 4.14 Dilute Sample with N₂ and Recirc.
 - 4.14.1 Depress ACTIVATE button.
 - 4.14.2 Depress SLOW button and watch pressure gauge slowly rise to 0" of Hg. Depress STOP button.
 - 4.14.3 Depress RECIRC button and wait 5 minutes.
 - 4.14.4 Complete step a) if a syringe will be used for the gas sample. If a 100 ml gas bomb is being used for the gas sample, continue on to Step 4.14.5.
 - a) Insert the calibrated gas syringe into the septum on the gas sampler. Withdraw a 5 cc sample of gas and place the syringe into the portable shielded container.
 - 4.14.5 Depress the STOP button on the control panel.
 - 4.14.6 Turn the Selector Switch to SOLUTION CHANGEOUT.
- 4.15 Collect Particulate and Iodine Sample.
 - 4.15.1 Depress ACTIVATE button.
 - 4.15.2 Depress TS SAMPLE button.
 - 4.15.3 Depress and hold EMPTY button until thiosulfate solution has drained into 500 ml sample bottle.
 - 4.15.4 Depress TS SAMPLE GRAB button.
 - 4.15.5 Depress PURGE button and hold for 1 minute.
 - 4.15.6 Turn the Selector Switch to SYSTEM PURGE.
- 4.16 Purge the Sample Panel.
 - 4.16.1 Depress ACTIVATE button.
 - 4.16.2 Depress EVAC button and watch pressure gauge slowly drop to ~ - 19" of Hg. Depress STOP button.
 - 4.16.3 Press down and release GAS PURGE toggle switch and watch pressure swiftly rise to + 10" of Hg. Depress STOP button.
 - 4.16.4 Depress EVAC button and watch the pressure gauge drop to 0" of Hg. Depress STOP button.

- 4.16.5 Depress the PUMP button and wait 30 seconds. Depress STOP button.
- 4.16.6 Repeat Steps 4.16.2 through 4.16.5 to purge the sample panel one additional time.
- 4.17 Remove Samples from Sample Panel.
- 4.17.1 Return to the sample panel and close both valves on the gas bomb (if used) and close the inlet valve on the gas sampler.
- 4.17.2 Disconnect the gas bomb (if used) from the sample panel. Place gas bomb in portable shielded container.
- 4.17.3 Reconnect the gas sampler line and open the inlet valve. Reconnect the thiosulfate sampler line.
- 4.17.4 Tightly cap the 500 ml sample bottle.
- 4.18 Switching the Sample System Off.
- 4.18.1 Turn the Selector Switch to OFF.
- 4.18.2 Turn the Radiation Monitor to OFF.
- 4.18.3 Turn the Keylock Switch to OFF.
- 4.18.4 Close the following valves:
- a) Nitrogen bottle - next to sample panel
 - b) TS Valve - inside sample panel
 - c) DI Water Inlet, Instrument Air Inlet, N₂ Inlet
(On left side of sample panel)
- 4.19 Transport the samples to the Count Room for analysis.
- 4.20 Calculate the sample volume using the data from Enclosure 5.2. Record this volume on sample data sticker.
- NOTE: If sample cannot be counted because of high activity, further dilute the gas samples as per procedure HP/0/B/1006/07.
- 4.21 Transmit sample analysis results to the Station Health Physicist or his designee.
- 4.22 Request Operations to return the Hydrogen Analyzer to service per Enclosure 5.6 or 5.7.
- 4.23 Clean the area around the sample panel and pump out the sump.
- NOTE: This step may be N/A if additional samples will be pulled or radiological conditions do not allow clean up.

5.0 Enclosures

- 5.1 Sampling Equipment
- 5.2 Sample Data Sheet
- 5.3 Operations Checklist for Bypassing H₂ Analysis Panel currently in Standby Mode
- 5.4 Operations Checklist for Bypassing H₂ Analysis Panel currently in Analyze Mode
- 5.5 Checklist for Operation of Sample Panel
- 5.6 Operations Checklist for Returning H₂ Analysis Panel Back to Service in Standby Mode
- 5.7 Operations Checklist for Returning H₂ Analysis Panel Back to Service in Analyze Mode
- 5.8 Valve Checklist for Sample Panel
- 5.9 Control Panel Diagram
- 5.10 Flow Diagram

ENCLOSURE 5.1

HP/1/A/1009/17

SAMPLING PANEL EQUIPMENT

- 1 Nalgene 500 ml Thiosulfate sample bottle.
- 2 Stainless Steel Gas Bombs
- 1 9/16" Combination Wrench
- 1 Stainless Steel Portable Shielded Container
- 1 Stopwatch
- 1 bottle Thiosulfate Solution (500 ml)
- 2 10" x 12" Clear Poly Bags
- 1 Calibrated Gas Syringe
- 1 Bucket

ENCLOSURE 5.2

HP/1/A/1009/17

SAMPLE DATA SHEET

1) NAME _____
 DATE _____
 UNIT, _____

2) Sample Line Temperature _____

3) Sample Inlet Line Pressure _____

4) Gas Sample Volume = SV

$$SV = \frac{4307.1 (STV)}{(275.224 + .555 [^{\circ}F]) (14.7 + P)} = \text{_____ ml}$$

where:

$^{\circ}F$ = Sample Line Temperature

P = Sample Inlet Line Pressure

STV = Sample Trap Volume

Unit 1 = 1.3 ml

5) Diluted Volume = $\frac{SV}{1E4} = \text{_____ ml}$

6) Record Diluted Volume as Gas Sample Volume on Sample Label.

7) Record Iodine and Particulate Sample Volume as 1.3 ml of sample in 500 ml of thiosulfate solution on sample label.

Checked Control Copy _____

Date _____

ENCLOSURE 5.3

HP/1/A/1009/17

OPERATIONS CHECKLIST FOR BYPASSING

H₂ ANALYSIS PANEL CURRENTLY IN STANDBY MODE

	<u>DATE</u> <u>INIT./TIME</u>	<u>VERIFICATION</u> <u>DATE</u> <u>INIT./TIME</u>
1.0 <u>Initial Conditions</u>		
1.1 Containment Integrity is required.	_____	_____
1.2 Designate a Licensed Operator assigned to immediately close containment isolation valves from the Control Room if an ES actuation occurs. This person may have other responsibilities, but they shall not prevent him from performing this evolution.	_____	
License Operator _____		
Unit Supervisor _____		
1.3 Record that the containment isolation valves will be opened on Enclosures 5.1 and 5.6 of OP/O/A/1102/20 (Shift Turnover). (1PR-81 and 1PR-84 or 1PR-90 and 1PR-87).	_____	
1.4 H ₂ Analysis Panel is in Standby Mode.	_____	
1.5 Reactor building pressure is less than 40 psig.	_____	
2.0 <u>Procedure</u>		
2.1 Place Post Accident Sampler in service as follows:		
2.1.1 Select which train to be used. Circle one: Trn. "A" or Trn. "B".	_____	
2.1.2 Ensure train is in standby mode by observing red light in gray cabinet.	_____	

NOTE: Use other train if not in standby.

ENCLOSURE 5..

HP/1/A/1009/17

		<u>DATE</u> <u>INIT./TIME</u>	<u>VERIFICATION</u> <u>DATE</u> <u>INIT./TIME</u>
2.1.3	At the selected train "Remote" Panel (blue cabinet), depress <u>both</u> black <u>ON</u> buttons for ' <u>BYP TO POST AC</u> '. Opens (1PR-83, 1PR-86) or (1PR-89, (1PR-92).	_____	_____
2.1.4	Turn sample valve selector switch to ' <u>Top Cont</u> '. Opens 1PR-71 or 1PR-76. (Red light will come on).	_____	_____
2.1.5	From the Control Room, open 1PR-81 and 1PR-84 (Containment Isolation Valves) if train "A" was selected. OR Open 1PR-87 and 1PR-90 (Containment Isolation Valves) if train "B" was selected.	_____	_____
CAUTION: If ES actuation occurs, immediately close isolation valves for containment isolation.			
2.1.6	Notify Unit Supervisor which train is selected. Unit Supervisor _____	_____	_____
2.1.7	Return completed enclosure to Health Physics Personnel operating Sample Panel.	_____	_____

Checked Control Copy _____

Date _____

ENCLOSURE 5.4

HP/1/A/1009/17

OPERATIONS CHECKLIST FOR BYPASSING

H₂ ANALYSIS PANEL CURRENTLY IN ANALYZE MODE

	<u>DATE INIT./TIME</u>	<u>VERIFICATION DATE INIT./TIME</u>
1.0 <u>Initial Conditions</u>		
1.1 H ₂ Analyzer is in Analyze Mode.	_____	
1.2 Reactor building pressure is less than 40 psig.	_____	
2.0 <u>Procedure</u>		
2.1 Place Post Accident Sampler in service as follows:		
2.1.1 Select which train is to be used. Circle one: Trn. "A" or Trn. "B".	_____	
2.1.2 At the "Remote" Panel (blue cabinet), position the "Off Standby, Analyze" selector to "Standby" and observe red light in grey cabinet.	_____	
2.1.3 At selected train "Remote" Panel (blue cabinet), depress both black ON buttons for 'BYP TO POST AC'. Opens (IPR-83, IPR-86) or (IPR-89, IPR-92).	_____	
2.1.4 Notify Unit Supervisor which train is selected.	_____	
Unit Supervisor _____		
2.1.5 Return completed enclosure to Health Physics Personnel operating Sample Panel.	_____	

Checked Control Copy _____

Date _____

ENCLOSURE 5.5
HP/1/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

NOTE: Complete steps in order listed. Initial steps as completed.

A) Switching System On

- ___ 1) Operations Bypass H₂ Analyzer by Enclosure 5.3 or 5.4.
- ___ 2) Obtain Sampling Equipment and Keys.
- ___ 3) Open Nitrogen bottle to 40 psig.
- ___ 4) Open:
 - a) DI Water Inlet
 - b) Instrument Air Inlet
 - c) N₂ Inlet
- ___ 5)
 - a) Ensure test tees on sample inlet and outlet lines are closed.
 - b) Ensure inlet valve on gas sampler is open
 - c) Ensure DI water supply line is open to the panel
- ___ 6)
 - a) Position thiosulfate funnel
 - b) Attach hose to fill port
 - c) Pour 500 ml of thiosulfate into funnel.
- ___ 7) Set switches on control panel:
 - a) Sample Volume Select - set on SMALL.
 - b) Dilution Volume Select - set on LARGE.
 - c) Selector Switch - set on OFF.
 - d) System Purge - set on NORMAL.
 - e) Refill Switch - set on OFF (down).
 - f) TC Switch - set on POSITION 1 (measures sample line temperature).
 - g) Sample Line Select Switch - Unit and Hydrogen Analyzer Train A
or B
- ___ 8) Key Lock Switch - POWER ON

ENCLOSURE 5.5
HP/1/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

- ___ 9) a) Radiation Monitor - ON (up)
 b) Radiation Monitor Selector - BATT (red test region)
 c) Radiation Monitor Selector - MR/HR or R/HR
- B) Purge the Sample Panel
- ___ 10) Selector Switch - SYSTEM PURGE
- ___ 11) Normal - Sample Purge - SAMPLE PURGE
- 12) a) ACTIVATE
 b) EVAC
 c) Pressure slowly drops to ~ - 19" of Hg.
 d) STOP
- 13) a) GAS PURGE - press down and release.
 b) Pressure swiftly rises to + 10" of Hg.
 c) STOP
- 14) a) EVAC
 b) Pressure drops to 0" of Hg.
 c) STOP
- 15) a) PUMP - wait 30 seconds
 b) STOP
- 16) a) Purge sample panel two (2) more times by completing Steps 12 through 15 two (2) more times.
- ___ 17) Normal - Sample Purge - NORMAL

ENCLOSURE 5.5
HP/1/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

___ 18) Selector Switch - SOLUTION CHANGEOUT

C) Preparation for Sampling

___ 19) Attach 500 ml sample bottle to TS Sampler.

___ 20) Attach gas bomb to gas sampler and ensure valves on gas bomb are open
(N/A step if syringe will be used instead of gas bomb).

___ 21) Record sample line temperature on Enclosure 5.2.

D) Flush Thiosulfate Sampler and fill with Thiosulfate

___ 22) a) ACTIVATE

b) FLUSH - hold 30 seconds

c) PURGE - hold 30 seconds

d) EMPTY - hold 45 seconds

e) Open TS (thiosulfate) valve

f) Refill - ON - wait 2 minutes

g) Refill - OFF

___ 23) Selector Switch - DILUTION VOLUME EVACUATION

ENCLOSURE 5.5
HP/1/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

E) Evacuate the Dilution Volume

- ___ 24) a) Activate
- b) Pressure slowly drops to ~ - 19" of Hg.
- c) Selector Switch - SAMPLE RECIRC

F) Recirc Containment Air and Trap a Sample

- ___ 25) ACTIVATE - wait 10 minutes
- ___ 26) Return to sample panel - note and record sample inlet line pressure on Enclosure 5.2.
- ___ 27) a) SAMPLE - wait 1 minute
- b) TRAP - wait 20 seconds
- c) Selector Switch - SAMPLE DILUTION

G) Dilute Sample with N₂ and Recirc.

- ___ 28) a) ACTIVATE
- b) SLOW
- c) Pressure slowly rises to 0" of Hg.
- d) STOP
- ___ 29) RECIRC - wait 5 minutes
- ___ 30) Complete a) if syringe will be used for gas sample. If gas bomb is being used, N/A this step and continue on to Step 31.
- a) Withdraw a 5 cc gas sample from the septum of the gas sampler using calibrated syringe. Place syringe in portable shielded container.

ENCLOSURE 5.5
HP/1/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

- 31) a) STOP
 b) Selector Switch - SOLUTION CHANGEOUT

H) Collect Particulate and Iodine Sample

- 32) a) ACTIVATE
 b) TS SAMPLE
 c) EMPTY - hold button until thiosulfate solution has drained into sample bottle.
 d) TS SAMPLE GRAB
 e) PURGE - hold button 1 minute
- 33) Selector Switch - SYSTEM PURGE

I) Purge the Sample Panel.

- 34) a) ACTIVATE
 b) EVAC
 c) Pressure slowly drops to ~ - 19" of Hg.
 d) STOP
- 35) a) GAS PURGE - press down and release
 b) Pressure swiftly rise to + 10" of Hg.
 c) STOP

ENCLOSURE 5.5
HP/1/A/1C 4/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

- 36) a) EVAC
 - b) Pressure drop to 0" of Hg.
 - c) STOP
- 37) a) PUMP - wait 30 seconds
 - b) STOP
- 38) Repeat Steps 34 through 37 one additional time.

J) Remove Samples from Sample Panel

- 39) Return to sample panel and close both valves on the gas bomb and the inlet valve on the gas sampler. (N/A step if gas bomb not used).
- 40) Disconnect gas bomb from sample panel. Place gas bomb in portable shielded container. (N/A step if gas bomb not used).
- 41) a) Reconnect gas sampler line and open inlet valve.
 - b) Reconnect thiosulfate sampler line.
 - c) Tightly cap sample bottle.

K) Switching System Off

- 42) Selector Switch - OFF
- 43) Turn the Radiation Monitor - OFF
- 44) Key Lock Switch - OFF

ENCLOSURE 5.5
HP/1/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

- ___ 45) Close:
 - a) Nitrogen bottle
 - b) TS Valve - inside sample panel
 - c) DI Water Inlet
 - d) Instrument Air Inlet
 - e) N₂ Inlet
- ___ 46) Transport samples to Count Room for analysis.
- ___ 47) Calculate sample volume using data from Enclosure 5.2. (Dilute gas samples per HP/O/B/1006/07 if needed).
- ___ 48) Transmit sample analysis results to Station Health Physicist or his designee.
- ___ 49) Request Operations to return the Hydrogen Analyzer to service per Enclosures 5.6 or 5.7.
- ___ 50) Clean area around sample panel and pump out sump. (This step may be N/A if additional samples will be pulled or radiological conditions do not allow cleanup.)

Checked Control Copy _____

Date _____

ENCLOSURE 5.6

HP/1/A/1009/17

OPERATIONS CHECKLIST FOR RETURNING H₂ ANALYSIS PANEL
BACK TO SERVICE IN STANDBY MODE

	DATE		VERIFICATION	
	INIT./TIME		DATE INIT./TIME	
1.0 Procedure				
1.1 Return the H ₂ Analysis train back to service as follows:				
1.1.1 Turn Sample Valve Selector switch to OFF. (Red light will go off). Closes IPR-71 or IPR-76.				
1.1.2 Depress the OFF buttons on both 'BYP TO POST AC' switches. Closes (IPR-83, IPR-86) or (IPR-89, IPR-92).				
1.1.3 From the Control Room, Close IPR-81 and IPR-84 if train "A" is selected. OR Close IPR-87 and IPR-90 if train "B" was selected.				
NOTE: This will regain containment integrity. Remove the containment isolation valves from Enclosure 5.1 and 5.6 of OP/0/A/1102/20 (Shift Turnover).				
1.1.4 Notify the Unit Supervisor the H ₂ Analysis Train is back in service.				
1.1.5 Return completed enclosure to personnel operating Post Accident Sample Panel.				

Checked Control Copy _____

Date _____

ENCLOSURE 5.7

HP/1/A/1009/17

OPERATIONS CHECKLIST FOR RETURNING UNIT 1 H₂ ANALYSIS

PANEL BACK TO SERVICE IN ANALYZE MODE

	DATE <u>INIT./TIME</u>	VERIFICATION DATE <u>INIT./Time</u>
1.0 <u>Initial Conditions</u>		
1.1 H ₂ Analysis Panel has been switched to Standby Mode for Post Accident sampling and is to be returned to Analyze Mode.	_____	
2.0 <u>Procedure</u>		
2.1 Return the H ₂ Analysis train back to service as follows:		
2.1.1 Depress the OFF buttons on both " <u>BYP to Post AC</u> " switches. Closes (1PR-83, 1PR-86) or (1PR-89, 1PR-92).	_____	
2.1.2 Position the " <u>Off, Standby, Analyze</u> " Selector to <u>Analyze</u> .	_____	
NOTE: When Analyze is selected, the indication will go up scale resulting in a possible High Hydrogen Alarm on both panels and in the Control Room. Then return down scale to the correct reading in approximately 3 minutes.		
2.1.3 Push the <u>Remote Selector</u> button to ensure control is from Remote Panel.	_____	
2.1.4 Reset the Common Alarm after the meter reading stabilizes.	_____	
2.1.5 Notify the Unit Supervisor the H ₂ Analysis Train is back in service.	_____	
2.1.6 Return completed enclosure to Health Physics Personnel Operating Post Accident Sample Panel.	_____	








ENCLOSURE 5.8


HP/1/A/1009/17

VALVE CHECKLIST FOR SAMPLE PANEL

NOTE: This checklist may be used to provide assistance in determining flow inside the sample panel. It is not intended to provide a verification for valve operation.

<u>ACTION</u>	<u>RESPONSE</u>
<u>Dilution Volume Evacuation</u>	
- Activate	Energize 1, 2, 9, 12, 19, 17
<u>Sample Recirculate</u>	
- Activate	Energize 1, 2, 5, 6, 27
- Sample	De-energize 27
- Trap	De-energize 2, 5
<u>Sample Dilution</u>	
- Activate	Energize 12, 17
- Slow	Energize 3, 20
- Stop	De-energize 3, 20
- Recirc	Energize Recirc Pump 16, 18
NOTE: Valve #17 will de-energize when selector switch is moved to another position.	
<u>System Purge</u>	
- Activate	Energize 9, 12, 19, 27
- Evac	Energize 12, 22
- Stop	Energize 1, 2
- Gas Purge (down)	De-energize 1, 2
- Stop	Energize 15, De-energize 1, 2
- Normal - Sample Purge (Sample Purge)	De-energize 15
- Pump	Energize 17
- Stop	Energize Pump 16, 18
	De-energize 15, or 1 and 2
	De-energize Pump 16, 18
<u>Solution Change Out</u>	
- Empty	Energize 10, 11, 13
- Flush	Energize 14, 10
- Purge	Energize 15, 10
- Refill	Energize 11, 13
- TS Sample	Energize 21
- TS Sample Grab	De-energize 21





INFORMATION ONLY

CONTROL COPY

Form SPD-1002-1

DUKE POWER COMPANY
PROCEDURE PREPARATION
PROCESS RECORD

(1) ID No: HP/2/A/1009/17
Change(s) 1 to
N/A Incorporated

(2) STATION: Oconee

(3) PROCEDURE TITLE: Operating Procedure For Post-Accident Containment Air
Sampling System

(4) PREPARED BY: Sarah Coy DATE: 12-13-83

(5) REVIEWED BY: Chris Young DATE: 12-19-83

Cross-Disciplinary Review By: _____ N/R: g

(6) TEMPORARY APPROVAL (IF NECESSARY):
By: _____ (SRO) Date: _____

By: _____ Date: _____

(7) APPROVED BY: Jerry S. Ban Date: 12/21/83

(8) MISCELLANEOUS:
Reviewed/Approved By: _____ Date: _____

Reviewed/Approved By: _____ Date: _____

INFORMATION ONLY

DUKE POWER COMPANY
OCONEE NUCLEAR STATION
OPERATING PROCEDURE FOR POST-ACCIDENT CONTAINMENT
AIR SAMPLING SYSTEM

1.0 Purpose

This procedure describes the operation of the Post-Accident Containment Air Sampling System which is used to obtain a prompt containment air sample under accident conditions while keeping radiation exposure ALARA. This procedure is also used to perform the semi-annual functional test of the system.

2.0 References

- 2.1 Duke Power Company Nuclear Station Post-Accident Containment Air Sampling System Manual
- 2.2 HP/0/B/1006/07, Procedure for Preparation of Gas Calibration Sources
- 2.3 CP/1/A/2002/04C, Operating Procedure for the Post Accident Liquid Sampling (PALS) System
- 2.4 HP/0/B/1009/15, Procedure for Sampling and Quantifying High Level Gaseous, Radioiodine and Particulate Radioactivity
- 2.5 EP/0/A/1800/04, Loss of Coolant
- 2.6 Station Directive 4.2.5, Independent Verification Requirements
- 2.7 Station Directive 3.1.37, Interim Control of Independent Verification Requirements

3.0 Limits and Precautions

3.1 The sampling cycle will require two (2) qualified technicians approximately one (1) hour per sample, of which about ten (10) minutes will be spent in the sample panel area. One qualified technician will operate the control panel while the other will perform transit duties to and from the panel.

3.2 Personnel communications can be achieved by phone.

Unit 1 & 2 - Ext. 1268 (by column AX-38)
Unit 3 - Ext. 1396 (by door to RCA)

3.3 The following items will never be used on the panel.

- a. Trap Area Evacuation
- b. Fast Sample Dilution

- 3.4 The Recirc Pump must never be used at any pressure other than 0 inches of Mercury.
- 3.5 Moving the Selector Switch from one mode to another stops all current system operations. Depressing the Activate button starts operation of the newly selected mode.
- 3.6 The radiation monitor on the control panel will provide levels of radiation at the sample panel. If the radiation monitor is not working properly, then a portable survey instrument will be used to determine radiation levels.
- 3.7 If problems with the pressure and/or temperature gauge are evident, such as going off scale or erratic response, the Selector Switch must be turned to the OFF mode and sampling discontinued until the problem is corrected.
- 3.8 If the sampling system cannot be operated, then HP/O/B/1009/15 (Ref. 2.4) will be used as an alternate method for obtaining a containment air sample.
- 3.9 Enclosure 5.5 will be used to check off the steps as the procedure is completed.
- 3.10 Operations must complete Enclosure 5.3 or 5.4 to bypass the Hydrogen Analyzer to bring containment air to the sampling system and to return the Hydrogen Analyzer to service after sampling is complete by Enclosure 5.6 or 5.7. These enclosures shall be independently verified to ensure that containment integrity is maintained. The Operations Unit Supervisor shall designate one "doer" and one "verifier" as required by References 2.6 and 2.7.
- 3.11 The front side of the sample panel is the side which contains the door. The left and right side of the sample panel will be determined by using this fact.
- 3.12 If radiation levels exceed 16 R/hr and cannot be reduced by purging the system, secure operation of the panel, move to a low background area, and contact the Station Health Physicist or his designee for further instructions.
- 3.13 Before sampling operations begin, the decision must be made based on radiological conditions in the reactor building and the sampling area whether to use a 100 ml gas bomb or a calibrated syringe for the gas sample. During emergency conditions, this decision will be made by the Station Health Physicist or his designee.
- 3.14 Enclosure 5.8, Valve Checklist for Sample Panel, may be used to provide assistance in determining flow inside the sample panel. It is not intended to provide a verification for valve operation.
- 3.15 During accident conditions, the keys needed for sampling will be located in the Shiftman's key cabinet.

3.16 The sampling system must not be used if reactor building pressure is greater than 40 psig.

4.0 Procedure

4.1 Locate the Shift Supervisor for Operations and request that Operations complete Enclosure 5.3 or 5.4 to bypass the Hydrogen Analyzer so the Post Accident Containment Gas Sampling System may be operated.

4.2 Obtain equipment necessary to perform sampling, including the thio-sulfate solution. Also obtain keys to the control panel and the sixth floor Ventilation Equipment Room.

NOTE: Necessary equipment for sampling is listed on Enclosure 5.1.

4.3 * Open the valve on the nitrogen bottle next to the sampling panel to 40 psig.

4.4 Open the

- (a) DI Water Inlet
- (b) Instrument Air Inlet
- (c) N₂ Inlet

located on the left side of the sample panel.

NOTE: Open inlets by rotating the back switches counterclockwise one-quarter turn to the upward position.

4.4.1 Ensure the test tees on the sample inlet and outlet lines are closed.

4.4.2 Ensure inlet valve on gas sampler is open (black switch parallel with line).

4.4.3 Ensure DI water supply line is open to the panel.

4.5 Position the thiosulfate funnel directly over the fill port located on top of the sample panel. Attach the hose on the funnel to the fill port and pour the 500 ml of thiosulfate solution into the funnel.

4.6 Set the switches listed below as follows:

- (a) Sample Volume Select - set on SMALL
- (b) Dilution Volume Select - set on LARGE
- (c) Selector Switch - set on OFF
- (d) System Purge - set on NORMAL
- (e) Refill Switch - set on OFF (down)

- (f) TC Switch - set on POSITION 1 (thermocouple measures sample line temperature)
 - (g) Sample Line Select Switch - turn to Unit and Hydrogen Analyzer (Train A or B) being used for this operation of the sampling system
- 4.7 Turn the Key Lock Switch to POWER ON and ensure the power on light has come on.
- 4.8 Turn the Radiation Monitor toggle switch ON (up).
- 4.8.1 Turn the selector on the Radiation Monitor to BATT and ensure the needle is in the "red test region." Turn the selector to the MR/HR or R/HR scale.
- NOTE: If the Radiation Monitor is not functioning properly, note that it is not working on Enclosure 5.5, Step 9 and use a portable survey instrument to determine radiation levels during sampling.
- 4.9 Purge the Sample Panel.
- 4.9.1 Turn Select Switch to SYSTEM PURGE
 - 4.9.2 Move Normal - Sample Purge to SAMPLE PURGE
 - 4.9.3 Depress ACTIVATE button.
 - 4.9.4 Depress EVAC button (Evac light on) and watch pressure gauge slowly drop to ~ - 19" of Hg. Depress STOP.
 - 4.9.5 Press down and release the GAS PURGE toggle switch and watch the pressure gauge swiftly rise to + 10" of Hg. Depress STOP button.
 - 4.9.6 Depress the EVAC button and watch the pressure gauge drop to 0" of Hg. Depress STOP button.
 - 4.9.7 Depress the PUMP button and wait for 30 seconds. Depress STOP button.
 - 4.9.8 Repeat Step 4.9.4 through 4.9.7 twice to purge the sample panel two more times.
 - 4.9.9 Move Normal - Sample Purge to NORMAL.
 - 4.9.10 Turn Selector Switch to SOLUTION CHANGE OUT.
- 4.10 Preparation for Sampling
- 4.10.1 Set the 500 ml sample bottle in a clear poly bag. Place the portable shielded container on the floor under the Thio-sulfate sampler (left side of panel), and place the sample bottle in the shielded container.

- 4.10.2 Detach the left side of the flexible tubing on the thiosulfate sampler located on the left side of the sample panel near the floor.
 - 4.10.3 Insert the free end of the tubing into the 500 ml sample bottle.
 - 4.10.4. Complete Steps a) and b) below if a 100 ml gas bomb will be used for the gas sample. If the gas sample will be drawn by syringe, go to Step 4.10.5.
 - a) Detach the side of the flexible tubing on the gas sampler between the inlet valve and the hard piping.
 - b) Attach a 100 ml gas bomb between the free end of the flexible tubing and the hard piping on the gas sampler. Ensure valves on gas bomb are open.
 - 4.10.5 Record sample line temperature reading for sample volume calculations on Enclosure 5.2.
- 4.11 Flush Thiosulfate Sampler and fill with Thiosulfate.
- 4.11.1 Depress ACTIVATE button.
 - 4.11.2 Depress FLUSH button and hold for 30 seconds.
 - 4.11.3 Depress PURGE button and hold for 30 seconds.
 - 4.11.4 Depress EMPTY button and hold for 45 seconds.
 - 4.11.5 Open the TS (thiosulfate) valve located inside the sample panel directly below the fillport. (Open valve in same manner as valves in Step 4.4).
 - 4.11.6 Move Refill toggle switch to ON (up) and wait 2 minutes. Move Refill to OFF (down).
 - 4.11.7 Turn Selector Switch to DILUTION VOLUME EVACUATION.
- 4.12 Evacuate the Dilution Volume.
- 4.12.1 Depress ACTIVATE button and watch pressure gauge drop to ~ - 19" of Hg. Turn Selector Switch to SAMPLE RECIRC.
- 4.13 Recirc Containment Air and Trap a Sample.
- 4.13.1 Depress ACTIVATE button and wait 10 minutes.
 - 4.13.2 Return to sample panel, and note pressure gauge reading on sample inlet line. Record pressure on Enclosure 5.2.
 - 4.13.3 Depress SAMPLE button and wait 1 minute.

- 4.13.4 Depress TRAP button and wait 10 seconds.
- 4.13.5 Turn Selector Switch to SAMPLE DILUTION.
- 4.14 Dilute Sample with N₂ and Recirc.
 - 4.14.1 Depress ACTIVATE button.
 - 4.14.2 Depress SLOW button and watch pressure gauge slowly rise to 0" of Hg. Depress STOP button.
 - 4.14.3 Depress RECIRC button and wait 5 minutes.
 - 4.14.4 Complete step a) if a syringe will be used for the gas sample. If a 100 ml gas bomb is being used for the gas sample, continue on to Step 4.14.5.
 - a) Insert the calibrated gas syringe into the septum on the gas sampler. Withdraw a 5 cc sample of gas and place the syringe into the portable shielded container.
 - 4.14.5 Depress the STOP button on the control panel.
 - 4.14.6 Turn the Selector Switch to SOLUTION CHANGEOUT.
- 4.15 Collect Particulate and Iodine Sample.
 - 4.15.1 Depress ACTIVATE button.
 - 4.15.2 Depress TS SAMPLE button.
 - 4.15.3 Depress and hold EMPTY button until thiosulfate solution has drained into 500 ml sample bottle.
 - 4.15.4 Depress TS SAMPLE GRAB button.
 - 4.15.5 Depress PURGE button and hold for 1 minute.
 - 4.15.6 Turn the Selector Switch to SYSTEM PURGE.
- 4.16 Purge the Sample Panel.
 - 4.16.1 Depress ACTIVATE button.
 - 4.16.2 Depress EVAC button and watch pressure gauge slowly drop to ~ - 19" of Hg. Depress STOP button.
 - 4.16.3 Press down and release GAS PURGE toggle switch and watch pressure swiftly rise to + 10" of Hg. Depress STOP button.
 - 4.16.4 Depress EVAC button and watch the pressure gauge drop to 0" of Hg. Depress STOP button.

- 4.16.5 Depress the PUMP button and wait 30 seconds. Depress STOP button.
- 4.16.6 Repeat Steps 4.16.2 through 4.16.5 to purge the sample panel one additional time.
- 4.17 Remove Samples from Sample Panel.
- 4.17.1 Return to the sample panel and close both valves on the gas bomb (if used) and close the inlet valve on the gas sampler.
- 4.17.2 Disconnect the gas bomb (if used) from the sample panel. Place gas bomb in portable shielded container.
- 4.17.3 Reconnect the gas sampler line and open the inlet valve. Reconnect the thiosulfate sampler line.
- 4.17.4 Tightly cap the 500 ml sample bottle.
- 4.18 Switching the Sample System Off.
- 4.18.1 Turn the Selector Switch to OFF.
- 4.18.2 Turn the Radiation Monitor to OFF.
- 4.18.3 Turn the Keylock Switch to OFF.
- 4.18.4 Close the following valves:
- a) Nitrogen bottle - next to sample panel
 - b) TS-Valve - inside sample panel
 - c) DI Water Inlet, Instrument Air Inlet, N₂ Inlet
(On left side of sample panel)
- 4.19 Transport the samples to the Count Room for analysis.
- 4.20 Calculate the sample volume using the data from Enclosure 5.2. Record this volume on sample data sticker.
- NOTE: If sample cannot be counted because of high activity, further dilute the gas samples as per procedure HP/0/B/1006/07.
- 4.21 Transmit sample analysis results to the Station Health Physicist or his designee.
- 4.22 Request Operations to return the Hydrogen Analyzer to service per Enclosure 5.6 or 5.7.
- 4.23 Clean the area around the sample panel and pump out the sump.
- NOTE: This step may be N/A if additional samples will be pulled or radiological conditions do not allow clean up.

5.0 Enclosures

- 5.1 Sampling Equipment
- 5.2 Sample Data Sheet
- 5.3 Operations Checklist for Bypassing H₂ Analysis Panel Currently in Standby Mode
- 5.4 Operations Checklist for Bypassing H₂ Analysis Panel Currently in Analyze Mode
- 5.5 Checklist for Operation of Sample Panel
- 5.6 Operations Checklist for Returning H₂ Analysis Panel Back to Service in Standby Mode
- 5.7 Operations Checklist for Returning H₂ Analysis Panel Back to Service in Analyze Mode
- 5.8 Valve Checklist for Sample Panel
- 5.9 Control Panel Diagram
- 5.10 Flow Diagram

ENCLOSURE 5.1

HP/2/A/1009/17

SAMPLING PANEL EQUIPMENT

- 1 Nalgene 500 ml Thiosulfate sample bottle.
- 2 Stainless Steel Gas Bombs
- 1 9/16" Combination Wrench
- 1 Stainless Steel Portable Shielded Container
- 1 Stopwatch
- 1 bottle Thiosulfate Solution (500 ml)
- 2 10" x 12" Clear Poly Bags
- 1 Calibrated Gas Syringe
- 1 Bucket

ENCLOSURE 5.2

HP/2/A/1009/17

SAMPLE DATA SHEET

- 1) NAME _____
 DATE _____
 UNIT _____

2) Sample Line Temperature _____

3) Sample Inlet Line Pressure _____

4) Gas Sample Volume = SV

$$SV = \frac{4307.1 (STV)}{(275.224 + .555 [^{\circ}F]) (14.7 + P)} = \text{_____ ml}$$

where:

$^{\circ}F$ = Sample Line Temperature

P = Sample Inlet Line Pressure

STV = Sample Trap Volume

Unit 2 = 1.3 ml

5) Diluted Volume = $\frac{SV}{1E4} = \text{_____ ml}$

6) Record Diluted Volume as Gas Sample Volume on Sample Label.

7) Record Iodine and Particulate Sample Volume as 1.3 ml of sample in 500 ml of thiosulfate solution on sample label.

Checked Control Copy _____

Date _____

ENCLOSURE 5.3

HP/2/A/1009/17

OPERATIONS CHECKLIST FOR BYPASSING

H₂ ANALYSIS PANEL CURRENTLY IN STANDBY MODE

	<u>DATE</u> <u>INIT./TIME</u>	<u>VERIFICATION</u> <u>DATE</u> <u>INIT./TIME</u>
1.0 <u>Initial Conditions</u>		
1.1 Containment Integrity is required.	_____	_____
1.2 Designate a Licensed Operator assigned to immediately close containment isolation valves from the Control Room if an ES actuation occurs. This person may have other responsibilities, but they shall not prevent him from performing this evolution.	_____	
License Operator _____		
Unit Supervisor _____		
1.3 Record that the containment isolation valves will be opened on Enclosures 5.1 and 5.6 of OP/0/A/1102/20 (Shift Turnover). (2PR-81 and 2PR-84 or 2PR-90 and 2PR-87)	_____	
1.4 H ₂ Analysis Panel is in Standby Mode.	_____	
1.5 Reactor building pressure is less than 40 psig.	_____	
2.0 <u>Procedure</u>		
2.1 Place Post Accident Sampler in service as follows:		
2.1.1 Select which train to be used. Circle one: Trn. "A" or Trn. "B".	_____	
2.1.2 Ensure train is in standby mode by observing red light in gray cabinet.	_____	

NOTE: Use other train if not in standby.

ENCLOSURE 5.3

HP/2/A/1009/17

		<u>DATE</u> <u>INIT./TIME</u>	<u>VERIFICATION</u> <u>DATE</u> <u>INIT./TIME</u>
2.1.3	At the selected train "Remote" Panel (blue cabinet), depress both black ON buttons for 'BYP TO POST AC'. Opens (2PR-83, 2PR-86) or (2PR-89, 2PR-92).	_____	_____
2.1.4	Turn sample valve selector switch to 'Top Cont'. Opens 2PR-71 or 2PR-76. (Red light will come on).	_____	_____
2.1.5	From the Control Room, open 2PR-81 and 2PR-84 (Containment Isolation Valves) if train "A" was selected. OR Open 2PR-87 and 2PR-90 (Containment Isolation Valves) if train "B" was selected.	_____	_____
CAUTION: If ES actuation occurs, immediately close isolation valves for containment isolation.			
2.1.6	Notify Unit Supervisor which train is selected. Unit Supervisor _____	_____	_____
2.1.7	Return completed enclosure to Health Physics Personnel operating Sample Panel.	_____	_____

Checked Control Copy _____

Date _____

ENCLOSURE 5.4

HP/2/A/1009/17

OPERATIONS CHECKLIST FOR BYPASSING

H₂ ANALYSIS PANEL CURRENTLY IN ANALYZE MODE

	<u>DATE</u> <u>INIT./TIME</u>	<u>VERIFICATION</u> <u>DATE</u> <u>INIT./TIME</u>
1.0 <u>Initial Conditions</u>		
1.1 H ₂ Analyzer is in Analyze Mode.	_____	
1.2 Reactor building pressure is less than 40 psig.	_____	
2.0 <u>Procedure</u>		
2.1 Place Post Accident Sampler in service as follows:		
2.1.1 Select which train is to be used. Circle one: Trn. "A" or Trn. "B".	_____	
2.1.2 At the "Remote" Panel (blue cabinet), position the "Off Standby, Analyze" selector to "Standby" and observe red light in grey cabinet.	_____	
2.1.3 At selected train "Remote" Panel (blue cabinet), depress <u>both</u> black <u>ON</u> buttons for ' <u>BYP TO POST AC</u> '. Opens (2PR-83, 2PR-86) or (2PR-89, 2PR-92).	_____	
2.1.4 Notify Unit Supervisor which train is selected. Unit Supervisor _____	_____	
2.1.5 Return completed enclosure to Health Physics Personnel operating Sample Panel.	_____	

Checked Control Copy _____

Date _____

ENCLOSURE 5.5
HP/2/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

NOTE: Complete steps in order listed. Initial steps as completed.

A) Switching System On

- ___ 1) Operations Bypass H₂ Analyzer by Enclosure 5.3 or 5.4.
- ___ 2) Obtain Sampling Equipment and Keys.
- ___ 3) Open Nitrogen bottle to 40 psig.
- ___ 4) Open:
 - a) DI Water Inlet
 - b) Instrument Air Inlet
 - c) N₂ Inlet
- ___ 5)
 - a) Ensure test tees on sample inlet and outlet lines are closed.
 - b) Ensure inlet valve on gas sampler is open.
 - c) Ensure DI water supply line is open to the panel.
- ___ 6)
 - a) Position thiosulfate funnel
 - b) Attach hose to fill port
 - c) Pour 500 ml of thiosulfate into funnel.
- ___ 7) Set switches on control panel:
 - a) Sample Volume Select - set on SMALL.
 - b) Dilution Volume Select - set on LARGE.
 - c) Selector Switch - set on OFF.
 - d) System Purge - set on NORMAL.
 - e) Refill Switch - set on OFF (down).
 - f) TC Switch - set on POSITION 1 (measures sample line temperature).
 - g) Sample Line Select Switch - Unit and Hydrogen Analyzer Train A
or B
- ___ 8) Key Lock Switch - POWER ON

ENCLOSURE 5.5
HP/2/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

- ___ 9) a) Radiation Monitor - ON (up)
- b) Radiation Monitor Selector - BATT (red test region)
- c) Radiation Monitor Selector - MR/HR or R/HR

B) Purge the Sample Panel

- ___ 10) Selector Switch - SYSTEM PURGE
- ___ 11) Normal - Sample Purge - SAMPLE PURGE

- 12) a) ACTIVATE
- b) EVAC
- c) Pressure slowly drops to ~ - 19" of Hg.
- d) STOP
- 13) a) GAS PURGE - press down and release.
- b) Pressure swiftly rises to + 10" of Hg.
- c) STOP
- 14) a) EVAC
- b) Pressure drops to 0" of Hg.
- c) STOP
- 15) a) PUMP - wait 30 seconds
- b) STOP
- 16) a) Purge sample panel two (2) more times by completing Steps 12 through 15 two (2) more times.
- ___ 17) Normal - Sample Purge - NORMAL

ENCLOSURE 5.5
HP/2/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

___ 18) Selector Switch - SOLUTION CHANGEOUT

C) Preparation for Sampling

___ 19) Attach 500 ml sample bottle to TS Sampler.

___ 20) Attach gas bomb to gas sampler and ensure valves on gas bomb are open
(N/A step if syringe will be used instead of gas bomb).

___ 21) Record sample line temperature on Enclosure 5.2.

D) Flush Thiosulfate Sampler and fill with Thiosulfate

___ 22) a) . ACTIVATE

b) FLUSH - hold 30 seconds

c) PURGE - hold 30 seconds

d) EMPTY - hold 45 seconds

e) Open TS (thiosulfate) valve

f) Refill - ON - wait 2 minutes

g) Refill - OFF

___ 23) Selector Switch - DILUTION VOLUME EVACUATION

E) Evacuate the Dilution Volume

___ 24) a) Activate

b) Pressure slowly drops to ~ - 19" of Hg.

c) Selector Switch - SAMPLE RECIRC

ENCLOSURE 5.5
HP/2/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

F) Recirc Containment Air and Trap a Sample

- ___ 25) ACTIVATE - wait 10 minutes
- ___ 26) Return to sample panel - note and record sample inlet line pressure on Enclosure 5.2.
- ___ 27) a) SAMPLE - wait 1 minute
b) TRAP - wait 20 seconds
c) Selector Switch - SAMPLE DILUTION

G) Dilute Sample with N₂ and Recirc.

- ___ 28) a) ACTIVATE
b) SLOW
c) Pressure slowly rises to 0" of Hg.
d) STOP
- ___ 29) RECIRC - wait 5 minutes
- ___ 30) Complete a) if syringe will be used for gas sample. If gas bomb is being used, N/A this step and continue on to Step 31.
 - a) Withdraw a 5 cc gas sample from the septum of the gas sampler using calibrated syringe. Place syringe in portable shielded container.
- ___ 31) a) STOP
b) Selector Switch - SOLUTION CHANGEOUT

ENCLOSURE 5.5
HF/2/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

H) Collect Particulate and Iodine Sample

- 32) a) ACTIVATE
 b) TS SAMPLE
 c) EMPTY - hold button until thiosulfate solution has drained into
 sample bottle.
 d) TS SAMPLE GRAB
 e) PURGE - hold button 1 minute
- 33) Selector Switch - SYSTEM PURGE

I) Purge the Sample Panel.

- 34) a) ACTIVATE
 b) EVAC
 c) Pressure slowly drops to ~ - 19" of Hg.
 d) STOP
- 35) a) GAS PURGE - press down and release
 b) Pressure swiftly rise to + 10" of Hg.
 c) STOP
- 36) a) EVAC
 b) Pressure drop to 0" of Hg.
 c) STOP

ENCLOSURE 5.5
HP/2/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

- 37 a) PUMP - wait 30 seconds
- b) STOP
- 38) Repeat Steps 34 through 37 one additional time.

J) Remove Samples from Sample Panel

- 39) Return to sample panel and close both valves on the gas bomb and the inlet valve on the gas sampler (N/A step if gas bomb not used).
- 40) Disconnect gas bomb from sample panel. Place gas bomb in portable shielded container. (N/A step if gas bomb not used).
- 41) a) Reconnect gas sampler line and open inlet valve.
b) Reconnect thiosulfate sampler line.
c) Tightly cap sample bottle.

K) Switching System Off

- 42) Selector Switch - OFF
- 43) Turn the Radiation Monitor - OFF
- 44) Key Lock Switch - OFF
- 45) Close:
 - a) Nitrogen bottle
 - b) TS Valve - inside sample panel
 - c) DI Water Inlet
 - d) Instrument Air Inlet
 - e) N₂ Inlet

ENCLOSURE 5.5
HP/2/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

- 46) Transport samples to Count Room for analysis.
- 47) Calculate sample volume using data from Enclosure 5.2. (Dilute gas samples per HP/0/B/1006/07 if needed).
- 48) Transmit sample analysis results to Station Health Physicist or his designee.
- 49) Request Operations to return the Hydrogen Analyzer to service per Enclosures 5.6 or 5.7.
- 50) Clean area around sample panel and pump out sump. (This step may be N/A if additional samples will be pulled or radiological conditions do not allow clean up.)

Checked Control Copy _____

Date _____

ENCLOSURE 5.6

HP/2/A/1009/17

OPERATIONS CHECKLIST FOR RETURNING H₂ ANALYSIS PANEL

BACK TO SERVICE IN STANDBY MODE

		<u>DATE</u> <u>INIT./TIME</u>	<u>VERIFICATION</u> <u>DATE</u> <u>INIT./TIME</u>
1.0	<u>Procedure</u>		
1.1	Return the H ₂ Analysis train back to service as follows:		
1.1.1	Turn Sample Valve Selector switch to <u>OFF</u> . (Red light will go off). Closes 2PR-71 or 2PR-76.	_____	_____
1.1.2	Depress the <u>OFF</u> buttons on both ' <u>BYP TO POST AC</u> ' switches. Closes (2PR-83, 2PR-86) or (2PR-89, 2PR-92).	_____	_____
1.1.3	From the Control Room, Close 2PR-81 and 2PR-84 if train "A" is selected. OR Close 2PR-87 and 2PR-90 if train "B" was selected.	_____	_____
NOTE:	This will regain containment integrity. Remove the containment isolation valves from Enclosure 5.1 and 5.6 of OP/O/A/1102/20 (Shift Turnover).		
1.1.4	Notify the Unit Supervisor the H ₂ Analysis Train is back in service.	_____	
1.1.5	Return completed enclosure to personnel operating Post Accident Sample Panel.	_____	

Checked Control Copy _____

Date _____

ENCLOSURE 5.7

HP/2/A/1009/17

OPERATIONS CHECKLIST FOR RETURNING UNIT 2 H₂ ANALYSIS

PANEL BACK TO SERVICE IN ANALYZE MODE

	<u>DATE</u> <u>INIT./TIME</u>	<u>VERIFICATION</u> <u>DATE</u> <u>INIT./Time</u>
1.0 <u>Initial Conditions</u>		
1.1 H ₂ Analysis Panel has been switched to Standby Mode for Post Accident sampling and is to be returned to Analyze Mode.	_____	
2.0 <u>Procedure</u>		
2.1 Return the H ₂ Analysis train back to service as follows:		
2.1.1 Depress the <u>OFF</u> buttons on both "BYP to Post AC" switches. Closes (2PR-83, 2PR-86) or (2PR-89, 2PR-92).	_____	
2.1.2 Position the " <u>Off, Standby, Analyze</u> " Selector to <u>Analyze</u> .	_____	
NOTE: When Analyze is selected, the indication will go up scale resulting in a possible High Hydrogen Alarm on both panels and in the Control Room. Then return down scale to the correct reading in approximately 3 minutes.		
2.1.3 Push the <u>Remote Selector</u> button to ensure control is from Remote Panel.	_____	
2.1.4 Reset the Common Alarm after the meter reading stabilizes.	_____	
2.1.5 Notify the Unit Supervisor the H ₂ Analysis Train is back in service.	_____	
2.1.6 Return completed enclosure to Health Physics Personnel Operating Post Accident Sample Panel.	_____	


ENCLOSURE 5.3

HP/2/A/1009/17


VALVE CHECKLIST FOR SAMPLE PANEL

NOTE: This checklist may be used to provide assistance in determining flow inside the sample panel. It is not intended to provide a verification for valve operation.

<u>ACTION</u>	<u>RESPONSE</u>
<u>Dilution Volume Evacuation</u>	
- Activate	Energize 1, 2, 9, 12, 19, 17
<u>Sample Recirculate</u>	
- Activate	Energize 1, 2, 5, 6, 27
- Sample	De-energize 27
- Trap	De-energize 2, 5
<u>Sample Dilution</u>	
- Activate	Energize 12, 17
- Slow	Energize 3, 20
- Stop	De-energize 3, 20
- Recirc	Energize Recirc Pump 16, 18
NOTE: Valve #17 will de-energize when selector switch is moved to another position.	
<u>System Purge</u>	
- Activate	Energize 9, 12, 19, 27
- Evac	Energize 12, 22
- Stop	Energize 1, 2
- Gas Purge (down)	De-energize 1, 2
- Stop	Energize 15, De-energize 1, 2
- Normal - Sample Purge (Sample Purge)	De-energize 15
- Pump	Energize 17
	Energize Pump 16, 18
	De-energize 15, or 1 and 2
- Stop	De-energize Pump 16, 18
<u>Solution Change Out</u>	
- Empty	Energize 10, 11, 13
- Flush	Energize 14, 10
- Purge	Energize 15, 10
- Refill	Energize 11, 13
- TS Sample	Energize 21
- TS Sample Grab	De-energize 21



300	T
250	E
200	P
150	E
100	R
50	A
	T
	U
	R
	E
	-
-25	Hg
-20	
-15	
-10	
-5	
0	V
+5	A



CONTROL COPY

DUKE POWER COMPANY
PROCEDURE PREPARATION
PROCESS RECORD

(1) ID No: HP/3/A/1009/17
Change(s) 1 to
N/A Incorporated

- (2) STATION: Oconee
- (3) PROCEDURE TITLE: Operating Procedure For Post-Accident Containment Air Sampling System
- (4) PREPARED BY: Sarah Coy DATE: 12-13-83
- (5) REVIEWED BY: Charlie Young DATE: 12-19-83
Cross-Disciplinary Review By: _____ N/R: g
- (6) TEMPORARY APPROVAL (IF NECESSARY):
By: _____ (SRO) Date: _____
By: _____ Date: _____
- (7) APPROVED BY: Jimmy D. Barr Date: 12/21/83
- (8) MISCELLANEOUS:
Reviewed/Approved By: _____ Date: _____
Reviewed/Approved By: _____ Date: _____

INFORMATION ONLY

DUKE POWER COMPANY
OCONEE NUCLEAR STATION
OPERATING PROCEDURE FOR POST-ACCIDENT CONTAINMENT
AIR SAMPLING SYSTEM

1.0 Purpose

This procedure describes the operation of the Post-Accident Containment Air Sampling System which is used to obtain a prompt containment air sample under accident conditions while keeping radiation exposure ALARA. This procedure is also used to perform the semi-annual functional test of the system.

2.0 References

- 2.1 Duke Power Company Nuclear Station Post-Accident Containment Air Sampling System Manual
- 2.2 HP/0/B/1006/07, Procedure for Preparation of Gas Calibration Sources
- 2.3 CP/1/A/2002/04C, Operating Procedure for the Post Accident Liquid Sampling (PALS) System
- 2.4 HP/0/B/1009/15, Procedure for Sampling and Quantifying High Level Gaseous, Radioiodine and Particulate Radioactivity
- 2.5 EP/0/A/1800/04, Loss of Coolant
- 2.6 Station Directive 4.2.5, Independent Verification Requirements
- 2.7 Station Directive 3.1.37, Interim Control of Independent Verification Requirements

3.0 Limits and Precautions

- 3.1 The sampling cycle will require two (2) qualified technicians approximately one (1) hour per sample, of which about ten (10) minutes will be spent in the sample panel area. One qualified technician will operate the control panel while the other will perform transit duties to and from the panel.
- 3.2 Personnel communications can be achieved by phone.

Unit 1 & 2 - Ext. 1268 (by column AX-38)
Unit 3 - Ext. 1396 (by door to RCA)
- 3.3 The following items will never be used on the panel.
 - a. Trap Area Evacuation
 - b. Fast Sample Dilution

- 3.4 The Recirc Pump must never be used at any pressure other than 0 inches of Mercury.
- 3.5 Moving the Selector Switch from one mode to another stops all current system operations. Depressing the Activate button starts operation of the newly selected mode.
- 3.6 The radiation monitor on the control panel will provide levels of radiation at the sample panel. If the radiation monitor is not working properly, then a portable survey instrument will be used to determine radiation levels.
- 3.7 If problems with the pressure and/or temperature gauge are evident, such as going off scale or erratic response, the Selector Switch must be turned to the OFF mode and sampling discontinued until the problem is corrected.
- 3.8 If the sampling system cannot be operated, then HP/0/B/1009/15 (Ref. 2.4) will be used as an alternate method for obtaining a containment air sample.
- 3.9 Enclosure 5.5 will be used to check off the steps as the procedure is completed.
- 3.10 Operations must complete Enclosure 5.3 or 5.4 to bypass the Hydrogen Analyzer to bring containment air to the sampling system and to return the Hydrogen Analyzer to service after sampling is complete by Enclosure 5.6 or 5.7. These enclosures shall be independently verified to ensure that containment integrity is maintained. The Operations Unit Supervisor shall designate one "doer" and one "verifier" as required by References 2.6 and 2.7.
- 3.11 The front side of the sample panel is the side which contains the door. The left and right side of the sample panel will be determined by using this fact.
- 3.12 If radiation levels exceed 16 R/hr and cannot be reduced by purging the system, secure operation of the panel, move to a low background area, and contact the Station Health Physicist or his designee for further instructions.
- 3.13 Before sampling operations begin, the decision must be made based on radiological conditions in the reactor building and the sampling area whether to use a 100 ml gas bomb or a calibrated syringe for the gas sample. During emergency conditions, this decision will be made by the Station Health Physicist or his designee.
- 3.14 Enclosure 5.8, Valve Checklist for Sample Panel, may be used to provide assistance in determining flow inside the sample panel. It is not intended to provide a verification for valve operation.
- 3.15 During accident conditions, the keys needed for sampling will be located in the Shiftman's key cabinet.

3.16 The sampling system must not be used if reactor building pressure is greater than 40 psig.

4.0 Procedure

4.1 Locate the Shift Supervisor for Operations and request that Operations complete Enclosure 5.3 or 5.4 to bypass the Hydrogen Analyzer so the Post Accident Containment Gas Sampling System may be operated.

4.2 Obtain equipment necessary to perform sampling, including the thio-sulfate solution. Also obtain keys to the control panel and the sixth floor Ventilation Equipment Room.

NOTE: Necessary equipment for sampling is listed on Enclosure 5.1.

4.3 Open the valve on the nitrogen bottle next to the sampling panel to 40 psig.

4.4 Open the

- (a) DI Water Inlet
- (b) Instrument Air Inlet
- (c) N₂ Inlet

located on the left side of the sample panel.

NOTE: Open inlets by rotating the back switches counterclockwise one-quarter turn to the upward position.

4.4.1 Ensure the test tees on the sample inlet and outlet lines are closed.

4.4.2 Ensure inlet valve on gas sampler is open (black switch parallel with line).

4.4.3 Ensure DI water supply line is open to the panel.

4.5 Position the thiosulfate funnel directly over the fill port located on top of the sample panel. Attach the hose on the funnel to the fill port and pour the 500 ml of thiosulfate solution into the funnel.

4.6 Set the switches listed below as follows:

- (a) Sample Volume Select - set on SMALL
- (b) Dilution Volume Select - set on LARGE
- (c) Selector Switch - set on OFF
- (d) System Purge - set on NORMAL
- (e) Refill Switch - set on OFF (down)

- (f) TC Switch - set on POSITION 1 (thermocouple measures sample line temperature)
 - (g) Sample Line Select Switch - turn to Unit and Hydrogen Analyzer (Train A or B) being used for this operation of the sampling system
- 4.7 Turn the Key Lock Switch to POWER ON and ensure the power on light has come on.
- 4.8 Turn the Radiation Monitor toggle switch ON (up).
- 4.8.1 Turn the selector on the Radiation Monitor to BATT and ensure the needle is in the "red test region." Turn the selector to the MR/HR or R/HR scale.
- NOTE: If the Radiation Monitor is not functioning properly, note that it is not working on Enclosure 5.5, Step 9 and use a portable survey instrument to determine radiation levels during sampling.
- 4.9 Purge the Sample Panel.
- 4.9.1 Turn Select Switch to SYSTEM PURGE
 - 4.9.2 Move Normal - Sample Purge to SAMPLE PURGE
 - 4.9.3 Depress ACTIVATE button.
 - 4.9.4 Depress EVAC button (Evac light on) and watch pressure gauge slowly drop to ~ - 19" of Hg. Depress STOP.
 - 4.9.5 Press down and release the GAS PURGE toggle switch and watch the pressure gauge swiftly rise to + 10" of Hg. Depress STOP button.
 - 4.9.6 Depress the EVAC button and watch the pressure gauge drop to 0" of Hg. Depress STOP button.
 - 4.9.7 Depress the PUMP button and wait for 30 seconds. Depress STOP button.
 - 4.9.8 Repeat Step 4.9.4 through 4.9.7 twice to purge the sample panel two more times.
 - 4.9.9 Move Normal - Sample Purge to NORMAL.
 - 4.9.10 Turn Selector Switch to SOLUTION CHANGE OUT.
- 4.10 Preparation for Sampling
- 4.10.1 Set the 500 ml sample bottle in a clear poly bag. Place the portable shielded container on the floor under the Thio-sulfate sampler (left side of panel), and place the sample bottle in the shielded container.

- 4.10.2 Detach the left side of the flexible tubing on the thiosulfate sampler located on the left side of the sample panel near the floor.
 - 4.10.3 Insert the free end of the tubing into the 500 ml sample bottle.
 - 4.10.4. Complete Steps a) and b) below if a 100 ml gas bomb will be used for the gas sample. If the gas sample will be drawn by syringe, go to Step 4.10.5.
 - a) Detach the side of the flexible tubing on the gas sampler between the inlet valve and the hard piping.
 - b) Attach a 100 ml gas bomb between the free end of the flexible tubing and the hard piping on the gas sampler. Ensure valves on gas bomb are open.
 - 4.10.5 Record sample line temperature reading for sample volume calculations on Enclosure 5.2.
- 4.11 Flush Thiosulfate Sampler and fill with Thiosulfate.
- 4.11.1 Depress ACTIVATE button.
 - 4.11.2 Depress FLUSH button and hold for 30 seconds.
 - 4.11.3 Depress PURGE button and hold for 30 seconds.
 - 4.11.4 Depress EMPTY button and hold for 45 seconds.
 - 4.11.5 Open the TS (thiosulfate) valve located inside the sample panel directly below the fillport. (Open valve in same manner as valves in Step 4.4).
 - 4.11.6 Move Refill toggle switch to ON (up) and wait 2 minutes. Move Refill to OFF (down).
 - 4.11.7 Turn Selector Switch to DILUTION VOLUME EVACUATION.
- 4.12 Evacuate the Dilution Volume.
- 4.12.1 Depress ACTIVATE button and watch pressure gauge drop to ~ - 19" of Hg. Turn Selector Switch to SAMPLE RECIRC.
- 4.13 Recirc Containment Air and Trap a Sample.
- 4.13.1 Depress ACTIVATE button and wait 10 minutes.
 - 4.13.2 Return to sample panel and note pressure gauge reading on sample inlet line. Record pressure on Enclosure 5.2.
 - 4.13.3 Depress SAMPLE button and wait 1 minute.

- 4.13.4 Depress TRAP button and wait 10 seconds.
- 4.13.5 Turn Selector Switch to SAMPLE DILUTION.
- 4.14 Dilute Sample with N₂ and Recirc.
 - 4.14.1 Depress ACTIVATE button.
 - 4.14.2 Depress SLOW button and watch pressure gauge slowly rise to 0" of Hg. Depress STOP button.
 - 4.14.3 Depress RECIRC button and wait 5 minutes.
 - 4.14.4 Complete step a) if a syringe will be used for the gas sample. If a 100 ml gas bomb is being used for the gas sample, continue on to Step 4.14.5.
 - a) Insert the calibrated gas syringe into the septum on the gas sampler. Withdraw a 5 cc sample of gas and place the syringe into the portable shielded container.
 - 4.14.5 Depress the STOP button on the control panel.
 - 4.14.6 Turn the Selector Switch to SOLUTION CHANGEOUT.
- 4.15 Collect Particulate and Iodine Sample.
 - 4.15.1 Depress ACTIVATE button.
 - 4.15.2 Depress TS SAMPLE button.
 - 4.15.3 Depress and hold EMPTY button until thiosulfate solution has drained into 500 ml sample bottle.
 - 4.15.4 Depress TS SAMPLE GRAB button.
 - 4.15.5 Depress PURGE button and hold for 1 minute.
 - 4.15.6 Turn the Selector Switch to SYSTEM PURGE.
- 4.16 Purge the Sample Panel.
 - 4.16.1 Depress ACTIVATE button.
 - 4.16.2 Depress EVAC button and watch pressure gauge slowly drop to ~ - 19" of Hg. Depress STOP button.
 - 4.16.3 Press down and release GAS PURGE toggle switch and watch pressure swiftly rise to + 10" of Hg. Depress STOP button.
 - 4.16.4 Depress EVAC button and watch the pressure gauge drop to 0" of Hg. Depress STOP button.

- 4.16.5 Depress the PUMP button and wait 30 seconds. Depress STOP button.
- 4.16.6 Repeat Steps 4.16.2 through 4.16.5 to purge the sample panel one additional time.

- 4.17 Remove Samples from Sample Panel.
 - 4.17.1 Return to the sample panel and close both valves on the gas bomb (if used) and close the inlet valve on the gas sampler.
 - 4.17.2 Disconnect the gas bomb (if used) from the sample panel. Place gas bomb in portable shielded container.
 - 4.17.3 Reconnect the gas sampler line and open the inlet valve. Reconnect the thiosulfate sampler line.
 - 4.17.4 Tightly cap the 500 ml sample bottle.

- 4.18 Switching the Sample System Off.
 - 4.18.1 Turn the Selector Switch to OFF.
 - 4.18.2 Turn the Radiation Monitor to OFF.
 - 4.18.3 Turn the Keylock Switch to OFF.
 - 4.18.4 Close the following valves:
 - a) Nitrogen bottle - next to sample panel
 - b) TS Valve - inside sample panel
 - c) DI Water Inlet, Instrument Air Inlet, N₂ Inlet
(On left side of sample panel)

- 4.19 Transport the samples to the Count Room for analysis.

- 4.20 Calculate the sample volume using the data from Enclosure 5.2. Record this volume on sample data sticker.

NOTE: If sample cannot be counted because of high activity, further dilute the gas samples as per procedure HP/0/B/1006/07.

- 4.21 Transmit sample analysis results to the Station Health Physicist or his designee.

- 4.22 Request Operations to return the Hydrogen Analyzer to service per Enclosure 5.6 or 5.7.

- 4.23 Clean the area around the sample panel and pump out the sump.

NOTE: This step may be N/A if additional samples will be pulled or radiological conditions do not allow clean up.

5.0 Enclosures

- 5.1 Sampling Equipment
- 5.2 Sample Data Sheet
- 5.3 Operations Checklist for Bypassing H₂ Analysis Panel currently in Standby Mode
- 5.4 Operations Checklist for Bypassing H₂ Analysis Panel currently in Analyze Mode
- 5.5 Checklist for Operation of Sample Panel
- 5.6 Operations Checklist for Returning H₂ Analysis Panel Back to Service in Standby Mode
- 5.7 Operations Checklist for Returning H₂ Analysis Panel Back to Service in Analyze Mode
- 5.8 Valve Checklist for Sample Panel
- 5.9 Control Panel Diagram
- 5.10 Flow Diagram

ENCLOSURE 5.1

HP/3/A/1009/17

SAMPLING PANEL EQUIPMENT

- 1 Nalgene 500 ml Thiosulfate sample bottle.
- 2 Stainless Steel Gas Bombs
- 1 9/16" Combination Wrench
- 1 Stainless Steel Portable Shielded Container
- 1 Stopwatch
- 1 bottle Thiosulfate Solution (500 ml)
- 2 10" x 12" Clear Poly Bags
- 1 Calibrated Gas Syringe
- 1 Bucket

ENCLOSURE 5.2

HP/3/A/1009/17

SAMPLE DATA SHEET

1) NAME _____
 DATE _____
 UNIT _____

2) Sample Line Temperature _____

3) Sample Inlet Line Pressure _____

4) Gas Sample Volume = SV

$$SV = \frac{4307.1 (STV)}{(275.224 + .555 [^{\circ}F]) (14.7 + P)} = \text{_____ ml}$$

where:

$^{\circ}F$ = Sample Line Temperature

P = Sample Inlet Line Pressure

STV = Sample Trap Volume

Unit 3 = 1.2 ml

5) Diluted Volume = $\frac{SV}{1E4}$ = _____ ml

6) Record Diluted Volume as Gas Sample Volume on Sample Label.

7) Record Iodine and Particulate Sample Volume as 1.2 ml of sample in 500 ml of thiosulfate solution on sample label.

Checked Control Copy _____

Date _____

ENCLOSURE 5.3

HP/3/A/1009/17

OPERATIONS CHECKLIST FOR BYPASSING

H₂ ANALYSIS PANEL CURRENTLY IN STANDBY MODE

	<u>DATE</u> <u>INIT./TIME</u>	<u>VERIFICATION</u> <u>DATE</u> <u>INIT./TIME</u>
1.0 <u>Initial Conditions</u>		
1.1 Containment Integrity is required.	_____	_____
1.2 Designate a Licensed Operator assigned to immediately close containment isolation valves from the Control Room if an ES actuation occurs. This person may have other responsibilities, but they shall not prevent him from performing this evolution.	_____	
License Operator _____		
Unit Supervisor _____		
1.3 Record the containment isolation valves that will be opened on Enclosures 5.1 and 5.6 of OP/0/A/1102/20 (Shift Turnover). (3PR-81 and 3PR-84 or 3PR-90 and 3PR-87)	_____	
1.4 H ₂ Analysis Panel is in standby mode.		
1.5 Reactor Building pressure is less than 40 psig.	_____	
2.0 <u>Procedure</u>		
2.1 Place Post Accident Sampler in service as follows:		
2.1.1 Select which train to be used. Circle one: Trn. "A" or Trn. "B".	_____	

ENCLOSURE 5.3

HP/3/A/1009/17

		<u>DATE</u> <u>INIT./TIME</u>	<u>VERIFICATION</u> <u>DATE</u> <u>INIT./TIME</u>
2.1.2	Ensure train is in standby mode by observing red light in gray cabinet.	_____	_____
NOTE:	Use other train if not in standby.		
2.1.3	At the selected train "Remote" Panel (blue cabinet), depress <u>both</u> black <u>CN</u> buttons for 'BYP TO POST AC'. Opens (3PR-83, 3PR-86) or (3PR-89, 3PR-92).	_____	_____
2.1.4	Turn sample valve selector switch to ' <u>Top Cont</u> '. Opens 3PR-71 or 3PR-76. (Red light will come on).	_____	_____
2.1.5	From the Control Room, open 3PR-81 and 3PR-84 (Containment Isolation Valves) if train "A" was selected. OR Open 3PR-87 and 3PR-90 (Containment Isolation Valves) if train "B" was selected.	_____	_____
CAUTION:	If ES actuation occurs, immediately close isolation valves for containment isolation.		
2.1.6	Notify Unit Supervisor which train is selected. Unit Supervisor _____	_____	_____
2.1.7	Return completed enclosure to Health Physics Personnel operating Sample Panel.	_____	_____

Checked Control Copy _____

Date _____

ENCLOSURE 5.4

HP/3/A/1009/17

OPERATIONS CHECKLIST FOR BYPASSING

H₂ ANALYSIS PANEL CURRENTLY IN ANALYZE MODE

	<u>DATE</u> <u>INIT./TIME</u>	<u>VERIFICATION</u> <u>DATE</u> <u>INIT./TIME</u>
1.0 <u>Initial Conditions</u>		
1.1 H ₂ Analyzer is in Analyze Mode.	_____	
1.2 Reactor Building pressure is less than 40 psig.	_____	
2.0 <u>Procedure</u>		
2.1 Place Post Accident Sampler in service as follows:		
2.1.1 Select which train is to be used. Circle one: Trn. "A" or Trn. "B".	_____	
2.1.2 At the "Remote" Panel (blue cabinet) position the "Off, Standby, Analyze" selector to "Standby" and observe red light in grey cabinet.	_____	
2.1.3 At selected train "Remote" Panel (blue cabinet), depress <u>both</u> black <u>ON</u> buttons for ' <u>BYP TO POST AC</u> '. Opens (3PR-83, 3PR-86) or (3PR-89, 3PR-92).	_____	
2.1.4 Notify Unit Supervisor which train is selected.	_____	
Unit Supervisor _____		
2.1.5 Return completed enclosure to Health Physics Personnel operating Sample Panel.	_____	

Checked Control Copy _____

Date _____

ENCLOSURE 5.5
HP/3/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

NOTE: Complete steps in order listed. Initial steps as completed.

A) Switching System On

- ___ 1) Operations Bypass H₂ Analyzer by Enclosure 5.3 or 5.4.
- ___ 2) Obtain Sampling Equipment and Keys.
- ___ 3) Open Nitrogen bottle to 40 psig.
- ___ 4) Open:
 - a) DI Water Inlet
 - b) Instrument Air Inlet
 - c) N₂ Inlet
- ___ 5)
 - a) Ensure test tees on sample inlet and outlet lines are closed.
 - b) Ensure inlet valve on gas sampler is open.
 - c) Ensure DI water supply line is open to the panel.
- ___ 6)
 - a) Position thiosulfate funnel
 - b) Attach hose to fill port
 - c) Pour 500 ml of thiosulfate into funnel
- ___ 7) Set switches on control panel:
 - a) Sample Volume Select - set on SMALL.
 - b) Dilution Volume Select - set on LARGE.
 - c) Selector Switch - set on OFF.
 - d) System Purge - set on NORMAL.
 - e) Refill Switch - set on OFF (down).
 - f) TC Switch - set on POSITION 1 (measures sample line temperature).
 - g) Sample Line Select Switch - Unit and Hydrogen Analyzer Train A
or B
- ___ 8) Key Lock Switch - POWER ON
- ___ 9) a) Radiation Monitor - ON (up)

ENCLOSURE 5.5
HP/3/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

- b) Radiation Monitor Selector - BATT (red test region)
- c) Radiation Monitor Selector - MR/HR or R/HR
- B) Purge the Sample Panel
- ___ 10) Selector Switch - SYSTEM PURGE
- ___ 11) Normal - Sample Purge - SAMPLE PURGE
- 12) a) ACTIVATE
 b) EVAC
 c) Pressure slowly drops to ~ - 19" of Hg.
 d) STOP
- 13) a) GAS PURGE - press down and release.
 b) Pressure swiftly rises to + 10" of Hg.
 c) STOP
- 14) a) EVAC
 b) Pressure drops to 0" of Hg.
 c) STOP
- 15) a) PUMP - wait 30 seconds
 b) STOP
- 16) a) Purge sample panel two (2) more times by completing
 Steps 12 through 15 two (2) more times.
- ___ 17) Normal - Sample Purge - NORMAL
- ___ 18) Selector Switch - SOLUTION CHANGEOUT

ENCLOSURE 5.5
HP/3/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

C) Preparation for Sampling

- ___ 19) Attach 500 ml sample bottle to TS Sampler.
- ___ 20) Attach gas bomb to gas sampler and ensure valves on gas bomb are open (N/A step if syringe will be used instead of gas bomb).
- ___ 21) Record sample line temperature on Enclosure 5.2.

D) Flush Thiosulfate Sampler and fill with Thiosulfate

- ___ 22) a) ACTIVATE
 - b) FLUSH - hold 30 seconds
 - c) PURGE - hold 30 seconds
 - d) EMPTY - hold 45 seconds
 - e) Open TS (thiosulfate) valve
 - f) Refill - ON - wait 2 minutes
 - g) Refill - OFF
- ___ 23) Selector Switch - DILUTION VOLUME EVACUATION

E) Evacuate the Dilution Volume

- ___ 24) a) Activate
 - b) Pressure slowly drops to ~ - 19" of Hg.
 - c) Selector Switch - SAMPLE RECIRC

ENCLOSURE 5.5
HP/3/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

F) Recirc Containment Air and Trap a Sample

- ___ 25) ACTIVATE - wait 10 minutes
- ___ 26) Return to sample panel - note and record sample inlet line pressure on Enclosure 5.2.
- ___ 27) a) SAMPLE - wait 1 minute
- b) TRAP - wait 20 seconds
- c) Selector Switch - SAMPLE DILUTION

G) Dilute Sample with N₂ and Recirc.

- ___ 28) a) ACTIVATE
- b) SLOW
- c) Pressure slowly rises to 0" of Hg.
- d) STOP
- ___ 29) RECIRC - wait 5 minutes
- ___ 30) Complete a) if syringe will be used for gas sample. If gas bomb is being used, N/A this step and continue on to Step 31.
- a) Withdraw a 5 cc gas sample from the septum of the gas sampler using calibrated syringe. Place syringe in portable shielded container.
- ___ 31) a) STOP
- b) Selector Switch - SOLUTION CHANGEOUT

ENCLOSURE 5.5
HP/3/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

H) Collect Particulate and Iodine Sample

- 32) a) ACTIVATE
 b) TS SAMPLE
 c) EMPTY - hold button until thiosulfate solution has drained into
 sample bottle.
 d) TS SAMPLE GRAB
 e) PURGE - hold button 1 minute
- 33) Selector Switch - SYSTEM PURGE

I) Purge the Sample Panel.

- 34) a) ACTIVATE
 b) EVAC
 c) Pressure slowly drops to ~ - 19" of Hg.
 d) STOP
- 35) a) GAS PURGE - press down and release
 b) Pressure swiftly rise to + 10" of Hg.
 c) STOP
- 36) a) EVAC
 b) Pressure drop to 0" of Hg.
 c) STOP

ENCLOSURE 5.5
HP/3/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

- 37 a) PUMP - wait 30 seconds
- b) STOP
- 38) Repeat Steps 34 through 37 one additional time.

J) Remove Samples from Sample Panel

- ___ 39) Return to sample panel and close both valves on the gas bomb and the inlet valve on the gas sampler. (N/A step if gas bomb not used).
- ___ 40) Disconnect gas bomb from sample panel. Place gas bomb in portable shielded container. (N/A step if gas bomb not used).
- ___ 41) a) Reconnect gas sampler line and open inlet valve.
- b) Reconnect thiosulfate sampler line.
- c) Tightly cap sample bottle.

K) Switching System Off

- ___ 42) Selector Switch - OFF
- ___ 43) Turn the Radiation Monitor - OFF
- ___ 44) Key Lock Switch - OFF
- ___ 45) Close:
 - a) Nitrogen bottle
 - b) TS Valve - inside sample panel
 - c) DI Water Inlet
 - d) Instrument Air Inlet
 - e) N₂ Inlet

ENCLOSURE 5.5
HP/3/A/1009/17
CHECKLIST FOR OPERATION OF SAMPLE PANEL

- ___ 46) Transport samples to Count Room for analysis.
- ___ 47) Calculate sample volume using data from Enclosure 5.2. (Dilute gas samples per HP/0/B/1006/07 if needed).
- ___ 48) Transmit sample analysis results to Station Health Physicist or his designee.
- ___ 49) Request Operations to return the Hydrogen Analyzer to service per Enclosures 5.6 or 5.7.
- ___ 50) Clean area around sample panel and pump out sump. (This step may be N/A if additional samples will be pulled or radiological conditions do not allow clean up.)

Checked Control Copy _____

Date _____

ENCLOSURE 5.6

HP/3/A/1009/17

OPERATIONS CHECKLIST FOR RETURNING H₂ ANALYSIS PANEL

BACK TO SERVICE IN STANDBY MODE

		<u>DATE</u> <u>INIT./TIME</u>	<u>VERIFICATION</u> <u>DATE</u> <u>INIT./TIME</u>
1.0	<u>Procedure</u>		
1.1	Return the H ₂ Analysis train back to service as follows:		
1.1.1	Turn Sample Valve Selector switch to <u>OFF</u> . (Red light will go off). Closes 3PR-71 or 3PR-76.	_____	_____
1.1.2	Depress the <u>OFF</u> buttons on both ' <u>BYP TO POST AC</u> ' switches. Closes (3PR-83, 3PR-86) or (3PR-89, 3PR-92).	_____	_____
1.1.3	From the Control Room, Close 3PR-81 and 3PR-84 if train "A" is selected. OR Close 3PR-87 and 3PR-90 if train "B" was selected.	_____	_____
NOTE:	This will regain containment integrity. Remove the containment isolation valves from Enclosure 5.1 and 5.6 of OP/0/A/1102/20 (Shift Turnover).		
1.1.4	Notify the Unit Supervisor the H ₂ Analysis Train is back in service.	_____	
1.1.5	Return completed enclosure to personnel operating Post Accident Sample Panel.	_____	

Checked Control Copy _____

Date _____

ENCLOSURE 5.7

HP/3/A/1009/17

OPERATIONS CHECKLIST FOR RETURNING UNIT 3 H₂ ANALYSIS

PANEL BACK TO SERVICE IN ANALYZE MODE

	<u>DATE</u> <u>INIT./TIME</u>	<u>VERIFICATION</u> <u>DATE</u> <u>INIT./Time</u>
1.0 <u>Initial Conditions</u>		
1.1 H ₂ Analysis Panel has been switched to Standby Mode for Post Accident sampling and is to be returned to Analyze Mode.	_____	
2.0 <u>Procedure</u>		
2.1 Return the H ₂ Analysis train back to service as follows:		
2.1.1 Depress the <u>OFF</u> buttons on both " <u>BYP to Post AC</u> " switches. Closes (3PR-83, 3PR-86) or (3PR-89, 3PR-92).	_____	
2.1.2 Position the " <u>Off, Standby, Analyze</u> " Selector to <u>Analyze</u> .	_____	
NOTE: When Analyze is selected, the indication will go up scale resulting in a possible High Hydrogen Alarm on both panels and in the Control Room. Then return down scale to the correct reading in approximately 3 minutes.		
2.1.3 Push the <u>Remote Selector</u> button to ensure control is from Remote Panel.	_____	
2.1.4 Reset the Common Alarm after the meter reading stabilizes.	_____	
2.1.5 Notify the Unit Supervisor the H ₂ Analysis Train is back in service.	_____	
2.1.6 Return completed enclosure to Health Physics Personnel Operating Post Accident Sample Panel.	_____	



ENCLOSURE 5.8

HP/3/A/1009/17

VALVE CHECKLIST FOR SAMPLE PANEL

NOTE: This checklist may be used to provide assistance in determining flow inside the sample panel. It is not intended to provide a verification for valve operation.

<u>ACTION</u>	<u>RESPONSE</u>
<u>Dilution Volume Evacuation</u>	
- Activate	Energize 1, 2, 9, 12, 19, 17
<u>Sample Recirculate</u>	
- Activate	Energize 1, 2, 5, 6, 27
- Sample	De-energize 27
- Trap	De-energize 2, 5
<u>Sample Dilution</u>	
- Activate	Energize 12, 17
- Slow	Energize 3, 20
- Stop	De-energize 3, 20
- Recirc	Energize Recirc Pump 16, 18
NOTE: Valve #17 will de-energize when selector switch is moved to another position.	
<u>System Purge</u>	
- Activate	Energize 9, 12, 19, 27
- Evac	Energize 12, 22
- Stop	Energize 1, 2
- Gas Purge (down)	De-energize 1, 2
- Stop	Energize 15, De-energize 1, 2
- Normal - Sample Purge (Sample Purge)	De-energize 15
- Pump	Energize 17
	Energize Pump 16, 18
	De-energize 15, or 1 and 2
- Stop	De-energize Pump 16, 18
<u>Solution Change Out</u>	
- Empty	Energize 10, 11, 13
- Flush	Energize 14, 10
- Purge	Energize 15, 10
- Refill	Energize 11, 13
- TS Sample	Energize 21
- TS Sample Grab	De-energize 21



	+5	300
	0	250
	-5	200
	-10	150
	-15	100
	-20	50
	-25	
VACUUM		TEMPERATURE

Flow Diagram

